



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB89409

Temp.: 22.3 °C

BalanceID: WC SC-4

SLOPE.: 97.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	08/10/2017	16:00
2	CAL2	1	Water			7.01	08/10/2017	16:01
3	CAL3	1	Water			10.03	08/10/2017	16:02
4	ICV	1	Water			7.00	08/10/2017	16:03
5	CCV1	1	Water			2.01	08/10/2017	16:05
6	I4705-03	1	Solid	20.02	20	5.82	08/10/2017	16:06
7	I4705-06	1	Solid	20.01	20	5.60	08/10/2017	16:07
8	I4706-03	1	Solid	20.03	20	8.75	08/10/2017	16:09
9	I4717-03	1	Solid	20.01	20	7.67	08/10/2017	16:10
10	I4736-15	1	Solid	20.02	20	8.62	08/10/2017	16:11
11	I4736-16	1	Solid	20.03	20	8.52	08/10/2017	16:12
12	I4736-17	1	Solid	20.05	20	9.00	08/10/2017	16:14
13	I4736-18	1	Solid	20.04	20	8.74	08/10/2017	16:15
14	I4736-19	1	Solid	20.02	20	9.04	08/10/2017	16:17
15	I4736-20	1	Solid	20.03	20	8.00	08/10/2017	16:19
16	CCV2	1	Water			12.01	08/10/2017	16:20
17	I4736-20DUP	1	Solid	20.05	20	8.01	08/10/2017	16:22
18	CCV3	1	Water			12.01	08/10/2017	16:24

WORKLIST(Hardcopy Internal Chain)

1329409

WorkList Name : CORROSIONITY I4706

WorkList ID : 101741

Date : 8/10/2017 2:13:03 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/15/2017	Solid	I4705-03	Corrosivity	Cool 4 deg C	PSEG01	L42	SP-A-COMP	08/10/2017	9045C
08/15/2017	Solid	I4705-06	Corrosivity	Cool 4 deg C	PSEG01	L42	SP-B-COMP	08/10/2017	9045C
08/15/2017	Solid	I4706-03	Corrosivity	Cool 4 deg C	PSEG05	L42	OR-02-081017	08/10/2017	9045C
08/15/2017	Solid	I4717-03	Corrosivity	Cool 4 deg C	PSEG05	L42	OK-02-081017	08/10/2017	9045C
	Solid	I4736-15	Corrosivity	Cool 4 deg C	LIRO01	L11	SB-32-COMP	08/10/2017	9045C
	Solid	I4736-16	Corrosivity	Cool 4 deg C	LIRO01	L11	SB-34-COMP	08/10/2017	9045C
	Solid	I4736-17	Corrosivity	Cool 4 deg C	LIRO01	L11	SB-37-COMP	08/10/2017	9045C
	Solid	I4736-18	Corrosivity	Cool 4 deg C	LIRO01	L11	SB-10-COMP	08/08/2017	9045C
	Solid	I4736-19	Corrosivity	Cool 4 deg C	LIRO01	L11	SB-36-COMP	08/10/2017	9045C
	Solid	I4736-20	Corrosivity	Cool 4 deg C	LIRO01	L11	SB-35-COMP	08/10/2017	9045C

Date/Time 08/10/17 2:13 PM
 Received by: J. C. A.
 Relinquished by: J. C. A.

Date/Time 08/10/17 7:00 PM
 Received by: C. A.
 Relinquished by: J. C. A.