



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB89410

Temp.: 22.3 °C

SLOPE.: 98.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	08/10/2017	18:15
2	CAL2	1	Water			7.01	08/10/2017	18:16
3	CAL3	1	Water			10.04	08/10/2017	18:17
4	ICV	1	Water			6.99	08/10/2017	18:19
5	CCV1	1	Water			2.01	08/10/2017	18:21
6	I4736-13	1	Water			6.72	08/10/2017	18:22
7	I4736-14	1	Water			6.60	08/10/2017	18:24
8	I4736-14DUP	1	Water			6.61	08/10/2017	18:26
9	CCV2	1	Water			12.01	08/10/2017	18:27

LB89410

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH 14736

WorkList ID : 101743

Date : 8/10/2017 5:55:53 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	14736-13-G	pH	Cool 4 deg C	LIRO01	L11	TWP-37	08/10/2017	SM 4500-PH B
	Water	14736-14 -G	pH	Cool 4 deg C	LIRO01	L11	TWP-32	08/10/2017	SM 4500-PH B

Date/Time 08.10.17 6.00 PM
 Received by: JP
 Relinquished by: CL

Date/Time 08.10.17 7.00 PM
 Received by: CL
 Relinquished by: JP