


7470A

LB 89422
INSTRUMENT ID: CV1

Linear

 μ Abs.:

34611



A= 0.0000e+000

B= 2.8867e-004

C= 2.4677e-002

Rho= 0.9998664

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.013	-0.013	-131	0.000	-131				
0.2	0.200	0.200	0.000	609	0.0 %	609				
2.5	2.500	2.475	-0.025	8489	0.0 %	8489				
5.0	5.000	5.112	0.112	17624	0.0 %	17624				
7.5	7.500	7.409	-0.091	25582	0.0 %	25582				
10.0	10.000	10.016	0.016	34611	0.0 %	34611				

LB89422 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 11 Aug 2017 11:20:50

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-131	-	0.0000	7470A	PPB	11 Aug 2017 11:25:07	S	Std
0.2 - 1	50.2	609	-	0.2000	7470A	PPB	11 Aug 2017 11:27:16	S	Std
2.5 - 1	52.5	8489	-	2.5000	7470A	PPB	11 Aug 2017 11:29:24	S	Std
5.0 - 1	55	17624	-	5.0000	7470A	PPB	11 Aug 2017 11:31:32	S	Std
7.5 - 1	57.5	25582	-	7.5000	7470A	PPB	11 Aug 2017 11:33:41	S	Std
10.0 - 1	510	34611	-	10.0000	7470A	PPB	11 Aug 2017 11:35:49	S	Std
ICV05 - 1	ICV05	13623	3.9572	-7470A	PPB	PPB	11 Aug 2017 11:39:07	U	SMPL
ICB05 - 1	ICB05	-196	-0.0319	-7470A	PPB	PPB	11 Aug 2017 11:41:13	U	SMPL
CCV42 - 1	CCV42	17509	5.0790	-7470A	PPB	PPB	11 Aug 2017 11:43:20	U	SMPL
CCB42 - 1	CCB42	-204	-0.0342	-7470A	PPB	PPB	11 Aug 2017 11:45:26	U	SMPL
CRA - 1	CRA	615	0.2022	-7470A	PPB	PPB	11 Aug 2017 11:47:33	U	SMPL
HighStd - 1	HighStd	34595	10.0112	-7470A	PPB	PPB	11 Aug 2017 11:49:39	U	SMPL
ChkStd - 1	ChkStd	24010	6.9556	-7470A	PPB	PPB	11 Aug 2017 11:51:46	U	SMPL
PB101430BL - 1	PBW	-244	-0.0458	-7470A	PPB	PPB	11 Aug 2017 11:53:55	U	SMPL
PB101430BS - 1	LCSW	12801	3.7199	-7470A	PPB	PPB	11 Aug 2017 11:56:03	U	SMPL
PB101430TB - 1	PB101402TB	-199	-0.0328	-7470A	PPB	PPB	11 Aug 2017 11:58:11	U	SMPL
I4683-02 - 1	CAPTAINSCOVE-WC-001	-124	-0.0111	-7470A	PPB	PPB	11 Aug 2017 12:00:19	U	SMPL
I4695-03 - 1	TR-04-080917	-125	-0.0114	-7470A	PPB	PPB	11 Aug 2017 12:02:27	U	SMPL
I4699-03 - 1	SU-04-080917	-130	-0.0129	-7470A	PPB	PPB	11 Aug 2017 12:04:36	U	SMPL
I4700-03 - 1	NB-01-080917-A	-122	-0.0105	-7470A	PPB	PPB	11 Aug 2017 12:06:43	U	SMPL
CCV43 - 1	CCV43	17968	5.2115	-7470A	PPB	PPB	11 Aug 2017 12:08:49	U	SMPL
CCB43 - 1	CCB43	-192	-0.0307	-7470A	PPB	PPB	11 Aug 2017 12:10:56	U	SMPL
I4700-06 - 1	NB-01-080917-B	-135	-0.0143	-7470A	PPB	PPB	11 Aug 2017 12:13:03	U	SMPL
I4700-09 - 1	NB-01-080917-C	-117	-0.0091	-7470A	PPB	PPB	11 Aug 2017 12:15:10	U	SMPL
I4704-03 - 1	DRUM-2-3-4-OILY-DEBRIS	-125	-0.0114	-7470A	PPB	PPB	11 Aug 2017 12:17:17	U	SMPL
I4704-05 - 1	DRUM-5-OILY-GRAVEL	-116	-0.0088	-7470A	PPB	PPB	11 Aug 2017 12:19:24	U	SMPL
I4704-05DUP - 1	DRUM-5-OILY-GRAVELDUP	-121	-0.0103	-7470A	PPB	PPB	11 Aug 2017 12:21:32	U	SMPL
I4704-05MS - 1	DRUM-5-OILY-GRAVELMS	12797	3.7188	-7470A	PPB	PPB	11 Aug 2017 12:23:39	U	SMPL
I4704-05MSD - 1	DRUM-5-OILY-GRAVELMSD	12819	3.7251	-7470A	PPB	PPB	11 Aug 2017 12:25:48	U	SMPL
I4704-05LX5 - 1		-173	-0.0253	-7470A	PPB	PPB	11 Aug 2017 12:27:56	U	SMPL
I4704-05A - 1		14258	4.1405	-7470A	PPB	PPB	11 Aug 2017 12:30:05	U	SMPL
CCV44 - 1	CCV44	17420	5.0533	-7470A	PPB	PPB	11 Aug 2017 12:32:11	U	SMPL
CCB44 - 1	CCB44	-184	-0.0284	-7470A	PPB	PPB	11 Aug 2017 12:34:18	U	SMPL