


7471A

LB89424
INSTRUMENT ID: CX1

Linear

 μ Abs.:

41193



A= 0.0000e+000

B= 2.4480e-004

C= 2.0317e-002

Rho= 0.9997816

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.001	0.001	-78	0.000	-78				
0.2	0.200	0.214	0.014	792	0.0 %	792				
2.5	2.500	2.542	0.042	10302	0.0 %	10302				
5.0	5.000	4.989	-0.011	20297	0.0 %	20297				
7.5	7.500	7.349	-0.151	29939	0.0 %	29939	0			
10.0	10.000	10.104	0.104	41193	0.0 %	41193				

LB89424 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 11 Aug 2017 13:14:30

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	-78	-	0.0007471A	PPB	11 Aug 2017 14:08:40	S	Std
0.2 - 1	50.2	792	-	0.20007471A	PPB	11 Aug 2017 14:10:57	S	Std
2.5 - 1	52.5	10302	-	2.50007471A	PPB	11 Aug 2017 14:13:15	S	Std
5.0 - 1	55	20297	-	5.00007471A	PPB	11 Aug 2017 14:15:32	S	Std
7.5 - 1	57.5	29939	-	7.50007471A	PPB	11 Aug 2017 14:17:50	S	Std
10.0 - 1	510	41193	-	10.00007471A	PPB	11 Aug 2017 14:20:06	S	Std
ICV06 - 1	ICV06	15678	3.8582	-7471A	PPB	11 Aug 2017 14:23:00	U	SMPL
ICB06 - 1	ICB06	-197	-0.0279	-7471A	PPB	11 Aug 2017 14:25:16	U	SMPL
CCV45 - 1	CCV45	20483	5.0345	-7471A	PPB	11 Aug 2017 14:27:32	U	SMPL
CCB45 - 1	CCB45	-211	-0.0313	-7471A	PPB	11 Aug 2017 14:29:47	U	SMPL
CRA - 1	CRA	752	0.2044	-7471A	PPB	11 Aug 2017 14:32:03	U	SMPL
HighStd - 1	HighStd	41169	10.0983	-7471A	PPB	11 Aug 2017 14:34:19	U	SMPL
ChkStd - 1	ChkStd	28590	7.0190	-7471A	PPB	11 Aug 2017 14:36:34	U	SMPL
PB101431BL - 1	PBS	-260	-0.0433	-7471A	PPB	11 Aug 2017 14:38:53	U	SMPL
PB101431BS - 1	LCSS	14086	3.4685	-7471A	PPB	11 Aug 2017 14:41:10	U	SMPL
I4706-01 - 1	OR-02-081017	1004	0.2661	-7471A	PPB	11 Aug 2017 14:43:27	U	SMPL
I4717-01 - 1	OK-02-081017	1118	0.2940	-7471A	PPB	11 Aug 2017 14:45:45	U	SMPL
I4723-01 - 1	OUTSIDE-FENCE	15671	3.8565	-7471A	PPB	11 Aug 2017 14:48:02	U	SMPL
I4723-03 - 1	ELECTRICAL-UNIT-BUS-SUPPORT	13648	3.3613	-7471A	PPB	11 Aug 2017 14:50:20	U	SMPL
I4723-05 - 1	STAINED-BRICK	775	0.2100	-7471A	PPB	11 Aug 2017 14:52:36	U	SMPL
CCV46 - 1	CCV46	20152	4.9534	-7471A	PPB	11 Aug 2017 14:54:52	U	SMPL
CCB46 - 1	CCB46	-189	-0.0259	-7471A	PPB	11 Aug 2017 14:57:08	U	SMPL
I4723-07 - 1	UPPER-BRICK	30	0.0277	-7471A	PPB	11 Aug 2017 14:59:24	U	SMPL
I4723-09 - 1	LEANING-WALL	1143	0.3001	-7471A	PPB	11 Aug 2017 15:01:40	U	SMPL
I4727-01 - 1	FK-PS-01-012	167	0.0612	-7471A	PPB	11 Aug 2017 15:03:56	U	SMPL
I4730-01 - 1	FK-PS-01-013	499	0.1425	-7471A	PPB	11 Aug 2017 15:06:13	U	SMPL
I4730-01DUP - 1	FK-PS-01-013DUP	507	0.1444	-7471A	PPB	11 Aug 2017 15:08:29	U	SMPL
I4730-01MS - 1	FK-PS-01-013MS	15479	3.8095	-7471A	PPB	11 Aug 2017 15:10:46	U	SMPL
I4730-01MSD - 1	FK-PS-01-013MSD	15321	3.7708	-7471A	PPB	11 Aug 2017 15:13:04	U	SMPL
I4730-01LX5 - 1		-67	0.0039	-7471A	PPB	11 Aug 2017 15:15:21	U	SMPL
I4730-01A - 1		16562	4.0746	-7471A	PPB	11 Aug 2017 15:17:39	U	SMPL
CCV47 - 1	CCV47	20201	4.9654	-7471A	PPB	11 Aug 2017 15:19:55	U	SMPL
CCB47 - 1	CCB47	-205	-0.0299	-7471A	PPB	11 Aug 2017 15:22:12	U	SMPL