

LB 89434


245.1

INSTRUMENT ID: CVJ

Linear

 μ Abs.:

27916



A= 0.0000e+000

B= 3.5876e-004

C= 6.3018e-002

Rho= 0.9998938

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.035	0.035	-77	0.000	-77				
0.2	0.200	0.224	0.024	450	0.0 %	450				
2.5	2.500	2.458	-0.042	6677	0.0 %	6677				
5.0	5.000	4.991	-0.009	13735	0.0 %	13735				
7.5	7.500	7.413	-0.087	20487	0.0 %	20487				
10.0	10.000	10.078	0.078	27916	0.0 %	27916				

LB89434 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 11 Aug 2017 15:36:26

Sample ID	Extended ID	μ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
0.0 - 1	50	-77	-	0.0000	245.1	PPB	11 Aug 2017 15:47:14	S	Std
0.2 - 1	50.2	450	-	0.2000	245.1	PPB	11 Aug 2017 15:49:23	S	Std
2.5 - 1	32.5	6677	-	2.5000	245.1	PPB	11 Aug 2017 15:51:31	S	Std
5.0 - 1	55	13735	-	5.0000	245.1	PPB	11 Aug 2017 15:53:40	S	Std
7.5 - 1	57.5	20487	-	7.5000	245.1	PPB	11 Aug 2017 16:03:16	S	Std
10.0 - 1	510	27916	-	10.0000	245.1	PPB	11 Aug 2017 16:07:50	S	Std
ICV07 - 1	ICV07	11415	4.1583	-245.1	PPB	PPB	11 Aug 2017 16:12:41	U	SMPL
ICB07 - 1	ICB07	-191	-0.0055	-245.1	PPB	PPB	11 Aug 2017 16:14:47	U	SMPL
CCV48 - 1	CCV48	13677	4.9698	-245.1	PPB	PPB	11 Aug 2017 16:16:54	U	SMPL
CCB48 - 1	CCB48	-207	-0.0112	-245.1	PPB	PPB	11 Aug 2017 16:19:00	U	SMPL
CRA - 1	CRA	431	0.2176	-245.1	PPB	PPB	11 Aug 2017 16:21:07	U	SMPL
HighStd - 1	HighStd	27919	10.0792	-245.1	PPB	PPB	11 Aug 2017 16:23:14	U	SMPL
ChkStd - 1	ChkStd	19608	7.0976	-245.1	PPB	PPB	11 Aug 2017 16:33:31	U	SMPL
PB101461BL - 1	PBW	-190	-0.0051	-245.1	PPB	PPB	11 Aug 2017 16:42:22	U	SMPL
PB101461BS - 1	LCSW	10852	3.9563	-245.1	PPB	PPB	11 Aug 2017 16:44:29	U	SMPL
I4681-10 - 1	TWP-17	4947	1.8378	-245.1	PPB	PPB	11 Aug 2017 16:48:45	U	SMPL
I4681-10DUP - 1	TWP-17DUP	4794	1.7829	-245.1	PPB	PPB	11 Aug 2017 16:50:53	U	SMPL
I4681-10MS - 1	TWP-17MS	16875	6.1171	-245.1	PPB	PPB	11 Aug 2017 16:53:02	U	SMPL
I4681-10MSD - 1	TWP-17MSD	16039	5.8172	-245.1	PPB	PPB	11 Aug 2017 16:55:09	U	SMPL
CCV49 - 1	CCV49	13267	4.8227	-245.1	PPB	PPB	11 Aug 2017 16:57:16	U	SMPL
CCB49 - 1	CCB49	-200	-0.0087	-245.1	PPB	PPB	11 Aug 2017 16:59:22	U	SMPL
I4681-11 - 1	TWP-09	-177	-0.0005	-245.1	PPB	PPB	11 Aug 2017 17:01:28	U	SMPL
I4736-13 - 1	TWP-37	28750	10.3774	-245.1	PPB	PPB	11 Aug 2017 17:03:35	U	SMPL
I4736-14 - 1	TWP-32	153	0.1179	-245.1	PPB	PPB	11 Aug 2017 17:05:41	U	SMPL
I4681-10LX5 - 1		608	0.2811	-245.1	PPB	PPB	11 Aug 2017 17:12:05	U	SMPL
I4681-10A - 1		18550	6.7180	-245.1	PPB	PPB	11 Aug 2017 17:14:14	U	SMPL
I3736-13DLX2 - 1		13239	4.8126	-245.1	PPB	PPB	11 Aug 2017 17:16:22	U	SMPL
CCV50 - 1	CCV50	13416	4.8761	-245.1	PPB	PPB	11 Aug 2017 17:18:31	U	SMPL
CCB50 - 1	CCB50	-193	-0.0062	-245.1	PPB	PPB	11 Aug 2017 17:20:38	U	SMPL