


7470A

LB 89437
INSTRUMENT ID: CV#

Linear

 μ Abs.:

29639



A= 0.0000e+000

B= 3.3565e-004

C= 3.9886e-004

Rho= 0.9999365

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.034	-0.034	-102	0.000	-102				
0.2	0.200	0.181	-0.019	539	0.0 %	539				
2.5	2.500	2.558	0.058	7620	0.0 %	7620				
5.0	5.000	4.993	-0.007	14874	0.0 %	14874				
7.5	7.500	7.553	0.053	22502	0.0 %	22502				
10.0	10.000	9.949	-0.051	29639	0.0 %	29639				

LB89437 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 11 Aug 2017 17:31:30

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-102	-	0.0000	7470A	PPB	11 Aug 2017 17:32:40	S	Std
0.2 - 1	50.2	539	-	0.2000	7470A	PPB	11 Aug 2017 17:34:49	S	Std
2.5 - 1	52.5	7620	-	2.5000	7470A	PPB	11 Aug 2017 17:36:57	S	Std
5.0 - 1	55	14874	-	5.0000	7470A	PPB	11 Aug 2017 17:39:06	S	Std
7.5 - 1	57.5	22502	-	7.5000	7470A	PPB	11 Aug 2017 17:43:40	S	Std
10.0 - 1	50	29639	-	10.0000	7470A	PPB	11 Aug 2017 17:50:40	S	Std
ICV08 - 1	ICV08	11918	4.0006	-7470A	PPB	11 Aug 2017 17:55:34	U	SMPL	
ICB08 - 1	ICB08	-201	-0.0671	-7470A	PPB	11 Aug 2017 17:57:41	U	SMPL	
CCV51 - 1	CCV51	15036	5.0472	-7470A	PPB	11 Aug 2017 18:02:11	U	SMPL	
CCB51 - 1	CCB51	-201	-0.0671	-7470A	PPB	11 Aug 2017 18:04:17	U	SMPL	
CRA - 1	CRA	490	0.1649	-7470A	PPB	11 Aug 2017 18:06:24	U	SMPL	
HighStd - 1	HighStd	29800	10.0026	-7470A	PPB	11 Aug 2017 18:08:30	U	SMPL	
ChkStd - 1	ChkStd	21007	7.0513	-7470A	PPB	11 Aug 2017 18:10:37	U	SMPL	
PB101462BL - 1	PBW	-214	-0.0714	-7470A	PPB	11 Aug 2017 18:12:46	U	SMPL	
PB101462BS - 1	LCSW	10391	3.4881	-7470A	PPB	11 Aug 2017 18:14:54	U	SMPL	
PB101462TB - 1	PB101425TB	-132	-0.0439	-7470A	PPB	11 Aug 2017 18:17:02	U	SMPL	
I4706-03 - 1	OR-02-081017	-94	-0.0312	-7470A	PPB	11 Aug 2017 18:19:10	U	SMPL	
I4717-03 - 1	OK-02-081017	-81	-0.0268	-7470A	PPB	11 Aug 2017 18:21:18	U	SMPL	
I4723-02 - 1	OUTSIDE-FENCE	-66	-0.0218	-7470A	PPB	11 Aug 2017 18:23:27	U	SMPL	
I4723-04 - 1	ELECTRICAL-UNIT-BUS-SUPPORT	-60	-0.0197	-7470A	PPB	11 Aug 2017 18:25:33	U	SMPL	
CCV52 - 1	CCV52	15064	5.0566	-7470A	PPB	11 Aug 2017 18:27:40	U	SMPL	
CCB52 - 1	CCB52	-178	-0.0593	-7470A	PPB	11 Aug 2017 18:29:47	U	SMPL	
I4723-06 - 1	STAINED-BRICK	-78	-0.0258	-7470A	PPB	11 Aug 2017 18:31:54	U	SMPL	
I4723-08 - 1	UPPER-BRICK	-51	-0.0167	-7470A	PPB	11 Aug 2017 18:34:01	U	SMPL	
I4723-10 - 1	LEANING-WALL	-4	-0.0009	-7470A	PPB	11 Aug 2017 18:36:07	U	SMPL	
I4736-15 - 1	SB-32-COMP	-76	-0.0251	-7470A	PPB	11 Aug 2017 18:38:14	U	SMPL	
I4736-16 - 1	SB-34-COMP	-56	-0.0184	-7470A	PPB	11 Aug 2017 18:40:22	U	SMPL	
I4736-17 - 1	SB-37-COMP	-79	-0.0261	-7470A	PPB	11 Aug 2017 18:42:29	U	SMPL	
I4736-18 - 1	SB-10-COMP	-58	-0.0191	-7470A	PPB	11 Aug 2017 18:44:37	U	SMPL	
I4736-19 - 1	SB-36-COMP	-56	-0.0184	-7470A	PPB	11 Aug 2017 18:46:45	U	SMPL	
I4736-20 - 1	SB-35-COMP	-36	-0.0117	-7470A	PPB	11 Aug 2017 18:48:54	U	SMPL	
I4736-20DUP - 1	SB-35-COMPDUP	-46	-0.0150	-7470A	PPB	11 Aug 2017 18:51:01	U	SMPL	
CCV53 - 1	CCV53	15038	5.0478	-7470A	PPB	11 Aug 2017 18:56:50	U	SMPL	
CCB53 - 1	CCB53	-180	-0.0600	-7470A	PPB	11 Aug 2017 18:58:57	U	SMPL	
I4736-20MS - 1	SB-35-COMPMS	10483	3.5190	-7470A	PPB	11 Aug 2017 19:01:04	U	SMPL	
I4736-20MSD - 1	SB-35-COMPMSD	10271	3.4478	-7470A	PPB	11 Aug 2017 19:03:11	U	SMPL	
I4736-20LX5 - 1		-202	-0.0674	-7470A	PPB	11 Aug 2017 19:15:57	U	SMPL	
I4736-20A - 1		14306	4.8021	-7470A	PPB	11 Aug 2017 19:18:07	U	SMPL	
CCV54 - 1	CCV54	15068	5.0579	-7470A	PPB	11 Aug 2017 19:20:14	U	SMPL	
CCB54 - 1	CCB54	-152	-0.0506	-7470A	PPB	11 Aug 2017 19:22:21	U	SMPL	