


7471A

LB89448
INSTRUMENT ID: CY1

Linear

 μ Abs.:

39990



A= 0.0000e+000

B= 2.5276e-004

C= -5.2723e-002

Rho= 0.9999117

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.045	0.045	386	0.000	386				
0.2	0.200	0.237	0.037	1145	0.0 %	1145				
2.5	2.500	2.421	-0.079	9785	0.0 %	9785				
5.0	5.000	4.978	-0.022	19904	0.0 %	19904				
7.5	7.500	7.465	-0.035	29742	0.0 %	29742	0			
10.0	10.000	10.055	0.055	39990	0.0 %	39990				

LB89448 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 14 Aug 2017 09:36:06

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	386	-	0.00007471A	PPB		14 Aug 2017 11:13:47	S	Std
0.2 - 1	50.2	1145	-	0.20007471A	PPB		14 Aug 2017 11:16:05	S	Std
2.5 - 1	52.5	9785	-	2.50007471A	PPB		14 Aug 2017 11:18:23	S	Std
5.0 - 1	55	19904	-	5.00007471A	PPB		14 Aug 2017 11:20:40	S	Std
7.5 - 1	57.5	29742	-	7.50007471A	PPB		14 Aug 2017 11:22:58	S	Std
10.0 - 1	510	39990	-	10.00007471A	PPB		14 Aug 2017 11:25:16	S	Std
ICV09 - 1	ICV09	16051	4.0043	-7471A	PPB		14 Aug 2017 11:28:09	U	SMPL
ICB09 - 1	ICB09	205	-0.0009	-7471A	PPB		14 Aug 2017 11:30:25	U	SMPL
CCV55 - 1	CCV55	20574	5.1475	-7471A	PPB		14 Aug 2017 11:32:41	U	SMPL
CCB55 - 1	CCB55	168	-0.0103	-7471A	PPB		14 Aug 2017 11:34:56	U	SMPL
CRA - 1	CRA	864	0.1657	-7471A	PPB		14 Aug 2017 11:37:12	U	SMPL
HighStd - 1	HighStd	39948	10.0444	-7471A	PPB		14 Aug 2017 11:39:28	U	SMPL
ChkStd - 1	ChkStd	28217	7.0793	-7471A	PPB		14 Aug 2017 11:41:44	U	SMPL
PB101489BL - 1	PBS	49	-0.0403	-7471A	PPB		14 Aug 2017 11:46:37	U	SMPL
PB101489BS - 1	LCSS	15955	3.9800	-7471A	PPB		14 Aug 2017 11:51:25	U	SMPL
I4705-01 - 1	SP-A-COMP	1336	0.2850	-7471A	PPB		14 Aug 2017 11:53:41	U	SMPL
I4705-01DUP - 1	SP-A-COMPDUP	1376	0.2951	-7471A	PPB		14 Aug 2017 11:55:59	U	SMPL
I4705-01MS - 1	SP-A-COMPMS	15907	3.9679	-7471A	PPB		14 Aug 2017 12:00:49	U	SMPL
I4705-01MSD - 1	SP-A-COMPMSD	16697	4.1676	-7471A	PPB		14 Aug 2017 12:03:07	U	SMPL
I4705-04 - 1	SP-B-COMP	668	0.1161	-7471A	PPB		14 Aug 2017 12:05:23	U	SMPL
CCV56 - 1	CCV56	20052	5.0156	-7471A	PPB		14 Aug 2017 12:07:39	U	SMPL
CCB56 - 1	CCB56	-176	-0.0972	-7471A	PPB		14 Aug 2017 12:09:55	U	SMPL
I4739-01 - 1	S11-WC-1	80279	20.2384	-7471A	PPB		14 Aug 2017 12:12:11	U	SMPL
I4739-02 - 1	S11-WC-2	73062	18.4142	-7471A	PPB		14 Aug 2017 12:14:27	U	SMPL
I4739-03 - 1	S11-WC-3	58045	14.6186	-7471A	PPB		14 Aug 2017 12:16:46	U	SMPL
I4739-04 - 1	S11-WC-4	86807	21.8884	-7471A	PPB		14 Aug 2017 12:19:04	U	SMPL
I4739-05 - 1	S11-WC-5	147761	37.2949	-7471A	PPB		14 Aug 2017 12:21:23	U	SMPL
I4739-06 - 1	S11-WC-6	59487	14.9830	-7471A	PPB		14 Aug 2017 12:23:42	U	SMPL
I4739-07 - 1	S11-WC-7	28955	7.2659	-7471A	PPB		14 Aug 2017 12:26:03	U	SMPL
I4750-02 - 1	B-15(0-10)COMP	17	-0.0484	-7471A	PPB		14 Aug 2017 12:28:23	U	SMPL
PB101490BL - 1	PBS	-204	-0.1043	-7471A	PPB		14 Aug 2017 12:30:42	U	SMPL
PB101490BS - 1	LCSS	13428	3.3413	-7471A	PPB		14 Aug 2017 12:32:58	U	SMPL
CCV57 - 1	CCV57	19889	4.9744	-7471A	PPB		14 Aug 2017 12:35:14	U	SMPL
CCB57 - 1	CCB57	-39	-0.0626	-7471A	PPB		14 Aug 2017 12:37:31	U	SMPL
I4753-01 - 1	DRUMS-COMP	7646	1.8799	-7471A	PPB		14 Aug 2017 12:39:47	U	SMPL
I4754-01 - 1	TOPSOIL-MATERIAL-COMP	1008	0.2021	-7471A	PPB		14 Aug 2017 12:42:03	U	SMPL
I4754-03 - 1	30-FINES-COMP	739	0.1341	-7471A	PPB		14 Aug 2017 12:44:20	U	SMPL
I4755-01 - 1	SA-06-081117-A	3062	0.7212	-7471A	PPB		14 Aug 2017 12:46:36	U	SMPL
I4755-04 - 1	SA-06-081117-B	1158	0.2400	-7471A	PPB		14 Aug 2017 12:48:53	U	SMPL
I4758-01 - 1	SU-02-081117	4525	1.0910	-7471A	PPB		14 Aug 2017 12:51:09	U	SMPL
I4760-02 - 1	SP-1-COMP	922	0.1803	-7471A	PPB		14 Aug 2017 12:53:27	U	SMPL
I4761-01 - 1	RUSB-01	896	0.1737	-7471A	PPB		14 Aug 2017 12:55:44	U	SMPL
I4761-02 - 1	RUSB-02	9752	2.4122	-7471A	PPB		14 Aug 2017 12:58:02	U	SMPL
I4761-03 - 1	RUSB-03	1013	0.2033	-7471A	PPB		14 Aug 2017 13:00:18	U	SMPL
CCV58 - 1	CCV58	20185	5.0492	-7471A	PPB		14 Aug 2017 13:02:34	U	SMPL
CCB58 - 1	CCB58	-58	-0.0674	-7471A	PPB		14 Aug 2017 13:04:51	U	SMPL
I4761-04 - 1	RUSB-04	2734	0.6383	-7471A	PPB		14 Aug 2017 13:07:07	U	SMPL
I4761-05 - 1	RURRSB-05	15864	3.9570	-7471A	PPB		14 Aug 2017 13:09:24	U	SMPL
I4761-06 - 1	RUSB-06	10346	2.5623	-7471A	PPB		14 Aug 2017 13:11:40	U	SMPL
I4761-06DUP - 1	RUSB-06DUP	11561	2.8694	-7471A	PPB		14 Aug 2017 13:13:56	U	SMPL
I4761-06MS - 1	RUSB-06MS	26609	6.6729	-7471A	PPB		14 Aug 2017 13:16:13	U	SMPL
I4761-06MSD - 1	RUSB-06MSD	26288	6.5918	-7471A	PPB		14 Aug 2017 13:18:30	U	SMPL
I4705-01LXS - 1		167	-0.0105	-7471A	PPB		14 Aug 2017 13:20:47	U	SMPL
I4705-01A - 1		17523	4.3763	-7471A	PPB		14 Aug 2017 13:23:05	U	SMPL
I4761-06LX5 - 1		1914	0.4311	-7471A	PPB		14 Aug 2017 13:25:23	U	SMPL
I4761-06A - 1		26971	6.7644	-7471A	PPB		14 Aug 2017 13:27:39	U	SMPL
CCV59 - 1	CCV59	19858	4.9665	-7471A	PPB		14 Aug 2017 13:29:56	U	SMPL
CCB59 - 1	CCB59	-71	-0.0707	-7471A	PPB		14 Aug 2017 13:32:12	U	SMPL
I4739-01DLX5 - 1		15635	3.8991	-7471A	PPB		14 Aug 2017 13:34:28	U	SMPL
I4739-02DLX5 - 1		14018	3.4904	-7471A	PPB		14 Aug 2017 13:36:45	U	SMPL
I4739-03DLX5 - 1		11094	2.7514	-7471A	PPB		14 Aug 2017 13:39:01	U	SMPL
I4739-04DLX5 - 1		16734	4.1769	-7471A	PPB		14 Aug 2017 13:41:18	U	SMPL
I4739-05DLX5 - 1		29714	7.4577	-7471A	PPB		14 Aug 2017 13:43:34	U	SMPL
I4739-06DLX5 - 1		11604	2.8803	-7471A	PPB		14 Aug 2017 13:45:51	U	SMPL
CCV60 - 1	CCV60	19995	5.0012	-7471A	PPB		14 Aug 2017 13:48:09	U	SMPL
CCB60 - 1	CCB60	-120	-0.0831	-7471A	PPB		14 Aug 2017 13:50:27	U	SMPL