


245.1

LB 89457
INSTRUMENT ID: CV2

Linear

 μ Abs.:

30744



A= 0.0000e+000

B= 3.2720e-004

C= -2.6900e-002

Rho= 0.9999345

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.000	-0.000	82	0.000	82				
0.2	0.200	0.214	0.014	737	0.0 %	737				
2.5	2.500	2.542	0.042	7850	0.0 %	7850				
5.0	5.000	4.912	-0.088	15095	0.0 %	15095				
7.5	7.500	7.499	-0.001	23002	0.0 %	23002				
10.0	10.000	10.033	0.033	30744	0.0 %	30744				

LB89457 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 14 Aug 2017 15:33:33

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	82	-	0.0000	245.1	PPB	14 Aug 2017 15:37:21	S	Std
0.2 - 1	50.2	737	-	0.2000	245.1	PPB	14 Aug 2017 15:39:29	S	Std
2.5 - 1	52.5	7850	-	2.5000	245.1	PPB	14 Aug 2017 15:41:38	S	Std
5.0 - 1	55	15095	-	5.0000	245.1	PPB	14 Aug 2017 15:43:46	S	Std
7.5 - 1	57.5	23002	-	7.5000	245.1	PPB	14 Aug 2017 15:45:54	S	Std
10.0 - 1	610	30744	-	10.0000	245.1	PPB	14 Aug 2017 15:50:28	S	Std
ICV10 - 1	ICV10	12060	3.9192	-245.1	PPB	PPB	14 Aug 2017 15:52:59	U	SMPL
ICB10 - 1	ICB10	-80	-0.0531	-245.1	PPB	PPB	14 Aug 2017 15:55:05	U	SMPL
CCV61 - 1	CCV61	15067	4.9030	-245.1	PPB	PPB	14 Aug 2017 15:57:12	U	SMPL
CCB61 - 1	CCB61	-31	-0.0370	-245.1	PPB	PPB	14 Aug 2017 15:59:18	U	SMPL
CRA - 1	CRA	709	0.2051	-245.1	PPB	PPB	14 Aug 2017 16:01:26	U	SMPL
HighStd - 1	HighStd	30635	9.9969	-245.1	PPB	PPB	14 Aug 2017 16:03:32	U	SMPL
ChkStd - 1	ChkStd	21348	6.9582	-245.1	PPB	PPB	14 Aug 2017 16:05:39	U	SMPL
PB101531BL - 1	PBW	-39	-0.0397	-245.1	PPB	PPB	14 Aug 2017 16:07:48	U	SMPL
PB101531BS - 1	LCSW	13221	4.2990	-245.1	PPB	PPB	14 Aug 2017 16:09:56	U	SMPL
I4348-04 - 1	0511B	22671	7.3911	-245.1	PPB	PPB	14 Aug 2017 16:16:21	U	SMPL
CCV62 - 1	CCV62	14908	4.8510	-245.1	PPB	PPB	14 Aug 2017 16:22:45	U	SMPL
CCB62 - 1	CCB62	-89	-0.0560	-245.1	PPB	PPB	14 Aug 2017 16:24:53	U	SMPL