



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB89469

Temp.: 22.2 °C

BalanceID: WC SC-4

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.00	08/14/2017	16:30
2	CAL2	1	Water			7.01	08/14/2017	16:31
3	CAL3	1	Water			10.02	08/14/2017	16:32
4	ICV	1	Water			7.00	08/14/2017	16:33
5	CCV1	1	Water			2.02	08/14/2017	16:34
6	I4793-03	1	Solid	20.02	20	7.92	08/14/2017	16:36
7	I4793-06	1	Solid	20.04	20	8.20	08/14/2017	16:37
8	I4793-09	1	Solid	20.03	20	8.50	08/14/2017	16:38
9	I4795-03	1	Solid	20.01	20	6.51	08/14/2017	16:39
10	I4798-05	1	Solid	20.01	20	10.96	08/14/2017	16:40
11	I4798-05DUP	1	Solid	20.02	20	10.97	08/14/2017	16:41
12	CCV2	1	Water			12.02	08/14/2017	16:42

1829469

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I4793

WorkList ID : 101876

Date : 8/14/2017 3:32:12 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/17/2017	Solid	I4793-03	Corrosivity	Cool 4 deg C	PSEG05	L43	EO-02-081417-A	08/14/2017	9045C
08/17/2017	Solid	I4793-06	Corrosivity	Cool 4 deg C	PSEG05	L43	EO-02-081417-B	08/14/2017	9045C
08/17/2017	Solid	I4793-09	Corrosivity	Cool 4 deg C	PSEG05	L43	EO-02-081417-C	08/14/2017	9045C
08/17/2017	Solid	I4795-03	Corrosivity	Cool 4 deg C	PSEG01	L52	SP-1-COMP	08/14/2017	9045C
08/17/2017	Solid	I4798-05	Corrosivity	Cool 4 deg C	PSEG01	L52	DRUM-1-6-COMP	08/14/2017	9045C

Date/Time

08/14/17 3:32:12 PM

Received by:

JO

Relinquished by:

Date/Time

08/14/17 6:40 PM

Received by:

CR

Relinquished by:

JO