



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB89470

Temp.: 22.1 °C

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	08/14/2017	16:25
2	CAL2	1	Water			7.01	08/14/2017	16:26
3	CAL3	1	Water			10.01	08/14/2017	16:27
4	ICV	1	Water			7.00	08/14/2017	16:29
5	CCV1	1	Water			2.01	08/14/2017	16:30
6	I4792-11	1	Water			8.18	08/14/2017	16:31
7	I4792-11DUP	1	Water			8.19	08/14/2017	16:32
8	CCV2	1	Water			12.01	08/14/2017	16:33

WORKLIST(Hardcopy Internal Chain)

2B 89470

WorkList Name : PH 14792

WorkList ID : 101879

Date : 8/14/2017 4:20:15 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	14792-11-12	pH	Cool 4 deg C	LIRO01	L63	TWP-01	08/14/2017	SM 4500-PH B

Date/Time 08-14-17 4:20 PM
 Received by: JP
 Relinquished by: JP

Date/Time 08-14-17 6:00 PM
 Received by: CD
 Relinquished by: JP