


7471A

LB89477
INSTRUMENT ID: LV2

Linear

 μ Abs.:

3851E



A= 0.0000e+000

B= 2.5917e-004

C= -2.2419e-002

Rho= 0.9999412

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.055	-0.055	-127	0.000	-127				
0.2	0.200	0.207	0.007	884	0.0 %	884				
2.5	2.500	2.559	0.059	9959	0.0 %	9959				
5.0	5.000	4.993	-0.007	19351	0.0 %	19351				
7.5	7.500	7.537	0.037	29169	0.0 %	29169	0			
10.0	10.000	9.960	-0.040	38516	0.0 %	38516				

LB89477 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 15 Aug 2017 09:16:11

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-127	-	0.00007471A	PPB		15 Aug 2017 09:24:05	S	Std
0.2 - 1	50.2	884	-	0.20007471A	PPB		15 Aug 2017 09:26:22	S	Std
2.5 - 1	52.5	9959	-	2.50007471A	PPB		15 Aug 2017 09:28:40	S	Std
5.0 - 1	55	19351	-	5.00007471A	PPB		15 Aug 2017 09:30:57	S	Std
7.5 - 1	57.5	29169	-	7.50007471A	PPB		15 Aug 2017 09:33:15	S	Std
10.0 - 1	510	38516	-	10.00007471A	PPB		15 Aug 2017 09:35:33	S	Std
ICV12 - 1	ICV12	15595	4.0194	-7471A	PPB		15 Aug 2017 09:38:44	U	SMPL
ICB12 - 1	ICB12	-39	-0.0325	-7471A	PPB		15 Aug 2017 09:41:00	U	SMPL
CCV65 - 1	CCV65	19030	4.9096	-7471A	PPB		15 Aug 2017 09:43:16	U	SMPL
CCB65 - 1	CCB65	-2	-0.0229	-7471A	PPB		15 Aug 2017 09:48:07	U	SMPL
CRA - 1	CRA	951	0.2241	-7471A	PPB		15 Aug 2017 09:50:22	U	SMPL
HighStd - 1	HighStd	38966	10.0765	-7471A	PPB		15 Aug 2017 09:52:38	U	SMPL
ChkStd - 1	ChkStd	26823	6.9293	-7471A	PPB		15 Aug 2017 09:54:54	U	SMPL
PB101541BL - 1	PBS	27	-0.0154	-7471A	PPB		15 Aug 2017 09:59:46	U	SMPL
PB101541BS - 1	LCSS	14961	3.8550	-7471A	PPB		15 Aug 2017 10:02:02	U	SMPL
I4768-01 - 1	W-137TH-SLUDGE-52	2928	0.7364	-7471A	PPB		15 Aug 2017 10:04:19	U	SMPL
I4768-03 - 1	HAMILTON-PL-SOIL-66	1354	0.3285	-7471A	PPB		15 Aug 2017 10:06:36	U	SMPL
I4773-01 - 1	E2-J10-ESW	596673	HIGH	-7471A	PPB		15 Aug 2017 10:08:54	X	
I4773-02 - 1	E2-J10-BASE	113658	29.4345	-7471A	PPB		15 Aug 2017 10:10:46	U	SMPL
I4773-03 - 1	E2-J10-NSW	71641	18.5449	-7471A	PPB		15 Aug 2017 10:13:28	U	SMPL
CCV66 - 1	CCV66	18706	4.8256	-7471A	PPB		15 Aug 2017 10:15:49	U	SMPL
CCB66 - 1	CCB66	-227	-0.0813	-7471A	PPB		15 Aug 2017 10:20:40	U	SMPL
I4773-04 - 1	E2-J10-SSW	3426	0.8655	-7471A	PPB		15 Aug 2017 10:22:55	U	SMPL
I4783-01 - 1	TP-106-J	5250	1.3382	-7471A	PPB		15 Aug 2017 10:25:11	U	SMPL
I4793-01 - 1	EO-02-081417-A	4490	1.1413	-7471A	PPB		15 Aug 2017 10:27:27	U	SMPL
I4793-04 - 1	EO-02-081417-B	3029	0.7626	-7471A	PPB		15 Aug 2017 10:29:44	U	SMPL
I4793-07 - 1	EO-02-081417-C	8799	2.2580	-7471A	PPB		15 Aug 2017 10:32:01	U	SMPL
I4794-01 - 1	CEMENT-BLOCK	2049	0.5086	-7471A	PPB		15 Aug 2017 10:34:18	U	SMPL
I4794-03 - 1	CONCRETE-FLOOR	0	-0.0224	-7471A	PPB		15 Aug 2017 10:36:35	U	SMPL
I4794-05 - 1	FACADE-BRICK-WALL	86	-0.0001	-7471A	PPB		15 Aug 2017 10:38:53	U	SMPL
I4795-01 - 1	SP-1-COMP	484	0.1030	-7471A	PPB		15 Aug 2017 10:41:11	U	SMPL
I4795-01DUP - 1	SP-1-COMPDUP	238	0.0393	-7471A	PPB		15 Aug 2017 10:49:27	U	SMPL
CCV67 - 1	CCV67	19073	4.9208	-7471A	PPB		15 Aug 2017 10:51:43	U	SMPL
CCB67 - 1	CCB67	-214	-0.0779	-7471A	PPB		15 Aug 2017 10:53:59	U	SMPL
I4795-01MS - 1	SP-1-COMPMS	15862	4.0886	-7471A	PPB		15 Aug 2017 10:56:16	U	SMPL
I4795-01MSD - 1	SP-1-COMPMSD	16027	4.1313	-7471A	PPB		15 Aug 2017 10:58:33	U	SMPL
PB101542BL - 1	PBS	-176	-0.0680	-7471A	PPB		15 Aug 2017 11:00:49	U	SMPL
PB101542BS - 1	LCSS	14380	3.7045	-7471A	PPB		15 Aug 2017 11:03:06	U	SMPL
I4797-01 - 1	E2-J9-BASE	167	0.0209	-7471A	PPB		15 Aug 2017 11:05:23	U	SMPL
I4797-02 - 1	E2-J9-WSW	160977	41.6982	-7471A	PPB		15 Aug 2017 11:07:39	U	SMPL
I4797-03 - 1	E2-J9-NSW	39766	10.2838	-7471A	PPB		15 Aug 2017 11:09:57	U	SMPL
I4797-04 - 1	E2-J9-SSW	18861	4.8658	-7471A	PPB		15 Aug 2017 11:12:20	U	SMPL
I4797-05 - 1	E2-K10-ESW	87168	22.5690	-7471A	PPB		15 Aug 2017 11:14:38	U	SMPL
I4797-06 - 1	E2-K9-BASE	124746	32.3082	-7471A	PPB		15 Aug 2017 11:16:55	U	SMPL
CCV68 - 1	CCV68	18457	4.7611	-7471A	PPB		15 Aug 2017 11:19:14	U	SMPL
CCB68 - 1	CCB68	-301	-0.1004	-7471A	PPB		15 Aug 2017 11:21:35	U	SMPL
I4797-07 - 1	E2-K10-BASE	144486	37.4242	-7471A	PPB		15 Aug 2017 11:23:52	U	SMPL
I4797-07DUP - 1	E2-K10-BASEDUP	142934	37.0220	-7471A	PPB		15 Aug 2017 11:26:09	U	SMPL
I4797-07MS - 1	E2-K10-BASEMS	158722	41.1138	-7471A	PPB		15 Aug 2017 11:31:24	U	SMPL
I4797-07MSD - 1	E2-K10-BASEMSD	157945	40.9124	-7471A	PPB		15 Aug 2017 11:36:21	U	SMPL
I4798-04 - 1	DRUM-1-6-COMP	501	0.1074	-7471A	PPB		15 Aug 2017 11:41:28	U	SMPL
I4795-01LX5 - 1		71	-0.0040	-7471A	PPB		15 Aug 2017 11:43:44	U	SMPL
I4795-01A - 1		14395	3.7084	-7471A	PPB		15 Aug 2017 11:46:02	U	SMPL
I4797-07LX5 - 1		26289	6.7909	-7471A	PPB		15 Aug 2017 11:48:19	U	SMPL
I4797-07A - 1		158374	41.0236	-7471A	PPB		15 Aug 2017 11:53:26	U	SMPL
I4773-01DLX10 - 1		103409	26.7782	-7471A	PPB		15 Aug 2017 11:55:44	U	SMPL
CCV69 - 1	CCV69	18813	4.8534	-7471A	PPB		15 Aug 2017 11:58:04	U	SMPL
CCB69 - 1	CCB69	-309	-0.1025	-7471A	PPB		15 Aug 2017 12:00:23	U	SMPL
I4773-01DL2X50 - 1		21345	5.5096	-7471A	PPB		15 Aug 2017 12:02:40	U	SMPL
I4773-03DLX5 - 1		14386	3.7060	-7471A	PPB		15 Aug 2017 12:07:14	U	SMPL
I4797-02DLX10 - 1		17171	4.4278	-7471A	PPB		15 Aug 2017 12:09:31	U	SMPL
I4797-03DLX2 - 1		20203	5.2136	-7471A	PPB		15 Aug 2017 12:11:47	U	SMPL
I4797-05DLX5 - 1		17022	4.3892	-7471A	PPB		15 Aug 2017 12:14:04	U	SMPL
I4797-06DLX5 - 1		24973	6.4499	-7471A	PPB		15 Aug 2017 12:16:21	U	SMPL
CCV70 - 1	CCV70	19371	4.9980	-7471A	PPB		15 Aug 2017 12:18:39	U	SMPL
CCB70 - 1	CCB70	-18	-0.0271	-7471A	PPB		15 Aug 2017 12:20:57	U	SMPL
I4797-07DLX10 - 1		13477	3.4704	-7471A	PPB		15 Aug 2017 12:23:14	U	SMPL
I4797-07DUPDLX10 - 1		13647	3.5145	-7471A	PPB		15 Aug 2017 12:25:31	U	SMPL
I4797-07MSDLX10 - 1		15384	3.9647	-7471A	PPB		15 Aug 2017 12:27:48	U	SMPL
I4797-07MSDDLX10 - 1		15002	3.8657	-7471A	PPB		15 Aug 2017 12:30:04	U	SMPL
I4797-07LDLX50 - 1		2660	0.6670	-7471A	PPB		15 Aug 2017 12:32:21	U	SMPL
I4797-07ADLX10 - 1		15495	3.9934	-7471A	PPB		15 Aug 2017 12:34:38	U	SMPL
I4773-02DLX5 - 1		24447	6.3135	-7471A	PPB		15 Aug 2017 12:36:54	U	SMPL
CCV71 - 1	CCV71	19181	4.9487	-7471A	PPB		15 Aug 2017 12:39:11	U	SMPL
CCB71 - 1	CCB71	-194	-0.0727	-7471A	PPB		15 Aug 2017 12:41:29	U	SMPL