

A analytical Summary Report

Analysis Method: SM5540 C

ANALYST: AMANDEEP

Parameter: MBAS

SUPERVISOR REVIEW BY: apatel

Run Number: LB89481

Reagent/Standard	Lot/Loq #
MBAS CAL STD, 0.5PPM	WP57258
MBAS CCV STD, 1.0PPM	WP57264
MBAS ICV-LCS std 1.0PPM	WP57265
MBAS DAILY CHECK STD, 0.01PPM	WP57266
MBAS DAILY CHECK STD, 2.0PPM	WP57267
MBAS CAL STD 2 PPM	WP57255
MBAS CAL STD 1 PPM	WP57257
MBAS CAL STD 0.1 PPM	WP57261
mbas cal std 0.05 ppm	WP57262
MBAS CAL STD 0.01 PPM	WP57263
MBAS BLANK 0 PPM	WP57254
MBAS CAL STD 0.4 PPM	WP57259
MBAS CAL STD 0.2 PPM	WP57260
Chloroform, 4X4L	W2175
MBAS Wash Solution	WP57273

Intercept: -0

Slope: 0.2275

Regression: 1

Seq	Lab ID	True Value (mg/L)	Dilution	Reading (652 nm)	Anal Date	Anal Time
1	CAL1	0	1	0.000	08/15/2017	12:01
2	CAL2	0.01	1	0.002	08/15/2017	12:01
3	CAL3	0.05	1	0.011	08/15/2017	12:02
4	CAL4	0.1	1	0.024	08/15/2017	12:02
5	CAL5	0.2	1	0.044	08/15/2017	12:03
6	CAL6	0.4	1	0.090	08/15/2017	12:04
7	CAL7	0.5	1	0.115	08/15/2017	12:04
8	CAL8	1	1	0.227	08/15/2017	12:04
9	CAL9	2	1	0.455	08/15/2017	12:05



Analytical Summary Report

Analysis Method: SM5540 C

ANALYST: AMANDEEP

Parameter: MBAS

SUPERVISOR REVIEW BY: apatel

Run Number: LB89481

Seq	Lab ID	True Value (mg/L)	Dilution	Reading (652 nm)	Result (mg/L)	AnalDate	Anal Time
1	ICV	1	1	0.219	0.963	08/15/2017	12:06
2	ICB		1	0.000	0.000	08/15/2017	12:06
3	CHK1	0.01	1	0.003	0.013	08/15/2017	12:07
4	CHK2	2	1	0.453	1.991	08/15/2017	12:07
5	CCV1	1	1	0.220	0.967	08/15/2017	12:08
6	CCB1		1	0.000	0.000	08/15/2017	12:08
7	LB89481BL		1	0.000	0.000	08/15/2017	12:09
8	LB89481BS	1	1	0.218	0.958	08/15/2017	12:09
9	I4346-26		1	0.084	0.369	08/15/2017	12:10
10	CCV2	1	1	0.226	0.993	08/15/2017	12:10
11	CCB2		1	0.000	0.000	08/15/2017	12:11