



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB89488

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	08/15/2017	09:35
2	I4736-13		1	>100.0	>212.0	08/15/2017	10:05
3	I4736-13DUP		1	>100.0	>212.0	08/15/2017	10:32
4	I4736-14		1	>100.0	>212.0	08/15/2017	10:55
5	I4792-11		1	>100.0	>212.0	08/15/2017	11:20

Result = (Celsius * 1.8) + 32

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP81517

WorkList ID : 101897

Date : 8/15/2017 8:49:13 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I4736-13	Flash Point	Cool 4 deg C	LIRO01	L11	TWP-37	08/10/2017	1010A
	Water	I4736-14	Flash Point	Cool 4 deg C	LIRO01	L11	TWP-32	08/10/2017	1010A
	Water	I4792-11	Flash Point	Cool 4 deg C	LIRO01	L63	TWP-01	08/14/2017	1010A

Date/Time 08/15/17 9:00 AM
 Received by: KS
 Relinquished by: CP

Date/Time 08/15/17 11:25 AM
 Received by: CP
 Relinquished by: KS