


7470A

LB 89491
INSTRUMENT ID: CV2

Linear

 μ Abs.:

35168



A= 0.0000e+000

B= 2.8554e-004

C= -2.5324e-002

Rho= 0.9998868

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.011	-0.011	50	0.000	50				
0.2	0.200	0.237	0.037	918	0.0 %	918				
2.5	2.500	2.534	0.034	8962	0.0 %	8962				
5.0	5.000	4.882	-0.118	17187	0.0 %	17187				
7.5	7.500	7.542	0.042	26501	0.0 %	26501				
10.0	10.000	10.017	0.017	35168	0.0 %	35168				

LB89491 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 15 Aug 2017 14:35:04

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	50	-	0.0000	7470A	PPB	15 Aug 2017 14:41:35	S	Std
0.2 - 1	50.2	918	-	0.2000	7470A	PPB	15 Aug 2017 14:43:43	S	Std
2.5 - 1	52.5	8962	-	2.5000	7470A	PPB	15 Aug 2017 14:45:52	S	Std
5.0 - 1	55	17187	-	5.0000	7470A	PPB	15 Aug 2017 14:48:00	S	Std
7.5 - 1	57.5	26501	-	7.5000	7470A	PPB	15 Aug 2017 14:50:08	S	Std
10.0 - 1	510	35168	-	10.0000	7470A	PPB	15 Aug 2017 14:52:15	S	Std
ICV13 - 1	ICV13	13606	3.8597	-7470A	PPB	15 Aug 2017 14:54:47	U	SMPL	
ICB13 - 1	ICB13	-52	-0.0402	-7470A	PPB	15 Aug 2017 14:56:53	U	SMPL	
CCV72 - 1	CCV72	17228	4.8940	-7470A	PPB	15 Aug 2017 14:59:00	U	SMPL	
CCB72 - 1	CCB72	-46	-0.0385	-7470A	PPB	15 Aug 2017 15:01:06	U	SMPL	
CRA - 1	CRA	870	0.2231	-7470A	PPB	15 Aug 2017 15:03:13	U	SMPL	
HighStd - 1	HighStd	35475	10.1042	-7470A	PPB	15 Aug 2017 15:05:19	U	SMPL	
ChkStd - 1	ChkStd	24277	6.9067	-7470A	PPB	15 Aug 2017 15:07:26	U	SMPL	
PB101533BL - 1	PBW	-217	-0.0873	-7470A	PPB	15 Aug 2017 15:09:35	U	SMPL	
PB101533BS - 1	LCSW	13425	3.8080	-7470A	PPB	15 Aug 2017 15:11:43	U	SMPL	
PB101533TB - 1	PB101464TB	-119	-0.0593	-7470A	PPB	15 Aug 2017 15:13:51	U	SMPL	
I4739-08 - 1	S11-WC-1	-66	-0.0442	-7470A	PPB	15 Aug 2017 15:15:59	U	SMPL	
I4739-09 - 1	S11-WC-2	-52	-0.0402	-7470A	PPB	15 Aug 2017 15:18:07	U	SMPL	
I4739-10 - 1	S11-WC-3	-50	-0.0396	-7470A	PPB	15 Aug 2017 15:20:16	U	SMPL	
I4739-11 - 1	S11-WC-4	8	-0.0230	-7470A	PPB	15 Aug 2017 15:22:22	U	SMPL	
CCV73 - 1	CCV73	17213	4.8897	-7470A	PPB	15 Aug 2017 15:24:29	U	SMPL	
CCB73 - 1	CCB73	-48	-0.0390	-7470A	PPB	15 Aug 2017 15:26:36	U	SMPL	
I4739-12 - 1	S11-WC-5	-9	-0.0279	-7470A	PPB	15 Aug 2017 15:28:42	U	SMPL	
I4739-13 - 1	S11-WC-6	-51	-0.0399	-7470A	PPB	15 Aug 2017 15:30:49	U	SMPL	
I4739-14 - 1	S11-WC-7	-34	-0.0350	-7470A	PPB	15 Aug 2017 15:32:56	U	SMPL	
I4750-03 - 1	B-15(0-10)COMP	-59	-0.0422	-7470A	PPB	15 Aug 2017 15:35:03	U	SMPL	
I4753-03 - 1	DRUMS-COMP	-37	-0.0359	-7470A	PPB	15 Aug 2017 15:37:11	U	SMPL	
I4755-03 - 1	SA-06-081117-A	-60	-0.0425	-7470A	PPB	15 Aug 2017 15:39:18	U	SMPL	
I4755-06 - 1	SA-06-081117-B	-54	-0.0407	-7470A	PPB	15 Aug 2017 15:41:26	U	SMPL	
I4758-03 - 1	SU-02-081117	-69	-0.0450	-7470A	PPB	15 Aug 2017 15:43:34	U	SMPL	
I4760-03 - 1	SP-1-COMP	-51	-0.0399	-7470A	PPB	15 Aug 2017 15:45:42	U	SMPL	
I4768-04 - 1	W-137TH-SLUDGE-52	-36	-0.0356	-7470A	PPB	15 Aug 2017 15:47:49	U	SMPL	
CCV74 - 1	CCV74	17205	4.8874	-7470A	PPB	15 Aug 2017 15:49:56	U	SMPL	
CCB74 - 1	CCB74	-72	-0.0459	-7470A	PPB	15 Aug 2017 15:52:03	U	SMPL	
I4768-05 - 1	HAMILTON-PL-SOIL-66	-37	-0.0359	-7470A	PPB	15 Aug 2017 15:54:10	U	SMPL	
I4768-05DUP - 1	HAMILTON-PL-SOIL-66DUP	-35	-0.0353	-7470A	PPB	15 Aug 2017 15:56:17	U	SMPL	
I4768-05MS - 1	HAMILTON-PL-SOIL-66MS	13027	3.6944	-7470A	PPB	15 Aug 2017 15:58:24	U	SMPL	
I4768-05MSD - 1	HAMILTON-PL-SOIL-66MSD	12260	3.4754	-7470A	PPB	15 Aug 2017 16:00:31	U	SMPL	
PB101534BL - 1	PBW	-115	-0.0582	-7470A	PPB	15 Aug 2017 16:02:38	U	SMPL	
PB101534BS - 1	LCSW	13641	3.8697	-7470A	PPB	15 Aug 2017 16:04:46	U	SMPL	
PB101534TB - 1	PB101487TB	-127	-0.0616	-7470A	PPB	15 Aug 2017 16:06:54	U	SMPL	
I4751-17 - 1	SB-30-COMP	17	-0.0205	-7470A	PPB	15 Aug 2017 16:09:03	U	SMPL	
I4751-18 - 1	SB-29-COMP	25	-0.0182	-7470A	PPB	15 Aug 2017 16:11:11	U	SMPL	
I4751-19 - 1	SB-38-COMP	0	-0.0253	-7470A	PPB	15 Aug 2017 16:13:18	U	SMPL	
CCV75 - 1	CCV75	17325	4.9217	-7470A	PPB	15 Aug 2017 16:15:25	U	SMPL	
CCB75 - 1	CCB75	-60	-0.0425	-7470A	PPB	15 Aug 2017 16:17:32	U	SMPL	
I4751-20 - 1	SB-28-COMP	1	-0.0250	-7470A	PPB	15 Aug 2017 16:19:39	U	SMPL	
I4751-21 - 1	SB-23-COMP	-11	-0.0285	-7470A	PPB	15 Aug 2017 16:21:46	U	SMPL	
I4751-22 - 1	SB-25-COMP	-11	-0.0285	-7470A	PPB	15 Aug 2017 16:23:53	U	SMPL	
I4751-23 - 1	SB-26-COMP	-41	-0.0370	-7470A	PPB	15 Aug 2017 16:26:00	U	SMPL	
I4751-24 - 1	SB-24-COMP	-14	-0.0293	-7470A	PPB	15 Aug 2017 16:28:07	U	SMPL	
I4751-24DUP - 1	SB-24-COMP	-5	-0.0268	-7470A	PPB	15 Aug 2017 16:30:15	U	SMPL	
I4751-24MS - 1	SB-24-COMPMS	12543	3.5562	-7470A	PPB	15 Aug 2017 16:32:23	U	SMPL	
I4751-24MSD - 1	SB-24-COMPMSD	12436	3.5257	-7470A	PPB	15 Aug 2017 16:34:31	U	SMPL	
CCV76 - 1	CCV76	17255	4.9017	-7470A	PPB	15 Aug 2017 16:50:28	U	SMPL	
CCB76 - 1	CCB76	-81	-0.0485	-7470A	PPB	15 Aug 2017 16:52:34	U	SMPL	
I4768-05LX5 - 1		0	-0.0253	-7470A	PPB	15 Aug 2017 16:54:42	U	SMPL	
I4768-05A - 1		13663	3.8760	-7470A	PPB	15 Aug 2017 16:56:49	U	SMPL	
I4751-24LX5 - 1		-68	-0.0447	-7470A	PPB	15 Aug 2017 16:58:56	U	SMPL	
I4751-24A - 1		13898	3.9431	-7470A	PPB	15 Aug 2017 17:01:03	U	SMPL	
CCV77 - 1	CCV77	16919	4.8057	-7470A	PPB	15 Aug 2017 17:03:10	U	SMPL	
CCB77 - 1	CCB77	-93	-0.0519	-7470A	PPB	15 Aug 2017 17:05:18	U	SMPL	