



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB89516

Temp.: 22.1 °C

BalanceID: WC SC-4

SLOPE.: 98.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.00	08/15/2017	15:25
2	CAL2	1	Water			7.01	08/15/2017	15:26
3	CAL3	1	Water			10.03	08/15/2017	15:27
4	ICV	1	Water			7.00	08/15/2017	15:30
5	CCV1	1	Water			2.01	08/15/2017	15:31
6	I4799-03	1	Solid	20.02	20	8.04	08/15/2017	15:32
7	I4799-06	1	Solid	20.03	20	8.06	08/15/2017	15:33
8	I4814-02	1	Solid	20.04	20	7.95	08/15/2017	15:35
9	I4815-03	1	Solid	20.01	20	8.82	08/15/2017	15:36
10	I4815-06	1	Solid	20.03	20	8.21	08/15/2017	15:37
11	I4817-03	1	Solid	20.04	20	9.23	08/15/2017	15:38
12	I4825-02	1	Solid	20.01	20	8.20	08/15/2017	15:39
13	I4825-02DUP	1	Solid	20.03	20	8.21	08/15/2017	15:40
14	CCV2	1	Water			12.01	08/15/2017	15:41

2089516

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY 14799

WorkList ID : 101913

Date : 8/15/2017 11:59:17 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/28/2017	Solid	14814-02	Corrosivity	Cool 4 deg C	NATU02	L53	WHITE-TWP-COMPOST	08/14/2017	9045C
08/18/2017	Solid	14825-02	Corrosivity	Cool 4 deg C	PSEG01	L62	TP-A-B-COMP	08/15/2017	9045C
08/18/2017	Solid	14799-03	Corrosivity	Cool 4 deg C	PSEG05	L52	AU-05-081417-A	08/14/2017	9045C
08/18/2017	Solid	14799-06	Corrosivity	Cool 4 deg C	PSEG05	L52	AU-05-081417-B	08/14/2017	9045C
	Solid	14815-03	Corrosivity	Cool 4 deg C	PSEG05	L52	TR-05-081517-A	08/15/2017	9045C
	Solid	14815-06	Corrosivity	Cool 4 deg C	PSEG05	L52	TR-05-081517-B	08/15/2017	9045C
	Solid	14817-03	Corrosivity	Cool 4 deg C	PSEG05	L53	SU-03-081517	08/15/2017	9045C

Date/Time

Received by:

Relinquished by:

08-15-17 12:00 PM

JD

JR

Date/Time

Received by:

Relinquished by:

08-15-17 2:00 PM

JP

JP