

ISM02.4


LB 89518

INSTRUMENT ID: CV1

Linear

 μ Abs.:

38582



A= 0.0000e+000

B= 2.5831e-004

C= 9.7115e-003

Rho= 0.9999783

Accept=Accepted

slope
 $y = 10161x$

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	QVD
0.00	0.000	0.004	0.004	-23	0.000	-23					-
0.2	0.200	0.192	-0.008	705	0.0 %	705					-4
2.5	2.500	2.508	0.008	9670	0.0 %	9670					0
5.0	5.000	4.975	-0.025	19221	0.0 %	19221					-1
7.5	7.500	7.547	0.047	29178	0.0 %	29178					1
10.0	10.000	9.976	-0.024	38582	0.0 %	38582					0

MSB
 81161x

LB89518
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	-23	-		0 ISM02.4	PPB	8/16/2017 10:48	Std
	0.2 S0.2	705	-		0.2 ISM02.4	PPB	8/16/2017 10:50	Std
	2.5 S2.5	9670	-		2.5 ISM02.4	PPB	8/16/2017 10:52	Std
	5 S5	19221	-		5 ISM02.4	PPB	8/16/2017 10:55	Std
	7.5 S7.5	29178	-		7.5 ISM02.4	PPB	8/16/2017 10:57	Std
	10 S10	38582	-		10 ISM02.4	PPB	8/16/2017 10:59	Std
ICV	ICV	15167	3.9274 -		ISM02.4	PPB	8/16/2017 11:02	SMPL
ICB	ICB	-390	-0.091 -		ISM02.4	PPB	8/16/2017 11:04	SMPL
CCV082	CCV082	19450	5.0338 -		ISM02.4	PPB	8/16/2017 11:06	SMPL
CCB082	CCB082	-310	-0.0704 -		ISM02.4	PPB	8/16/2017 11:09	SMPL
PB101565BL	PBW003	-202	-0.0425 -		ISM02.4	PPB	8/16/2017 11:11	SMPL
I4724-01	MJJ3S9	-320	-0.0729 -		ISM02.4	PPB	8/16/2017 11:13	SMPL
I4724-02	MJJ3T0	-320	-0.0729 -		ISM02.4	PPB	8/16/2017 11:16	SMPL
I4724-03	MJJ3T1	-325	-0.0742 -		ISM02.4	PPB	8/16/2017 11:18	SMPL
I4724-04	MJJ3T2	-331	-0.0758 -		ISM02.4	PPB	8/16/2017 11:20	SMPL
I4724-05	MJJ3T3	-312	-0.0709 -		ISM02.4	PPB	8/16/2017 11:22	SMPL
I4724-06	MJJ3T4	-284	-0.0636 -		ISM02.4	PPB	8/16/2017 11:25	SMPL
I4724-07	MJJ3T5	-312	-0.0709 -		ISM02.4	PPB	8/16/2017 11:27	SMPL
I4724-08	MJJ3T6	-303	-0.0686 -		ISM02.4	PPB	8/16/2017 11:29	SMPL
I4724-09	MJJ3T6D	-288	-0.0647 -		ISM02.4	PPB	8/16/2017 11:32	SMPL
I4724-10	MJJ3T6S	3801	0.9915 -		ISM02.4	PPB	8/16/2017 11:34	SMPL
I4724-11	MJJ3T7	-273	-0.0608 -		ISM02.4	PPB	8/16/2017 11:36	SMPL
I4724-12	MJJ3T8	-306	-0.0693 -		ISM02.4	PPB	8/16/2017 11:38	SMPL
I4724-13	MJJ3T9	-284	-0.0636 -		ISM02.4	PPB	8/16/2017 11:41	SMPL
I4724-14	MJJ3W0	-293	-0.066 -		ISM02.4	PPB	8/16/2017 11:43	SMPL
I4724-15	MJJ3W1	-281	-0.0629 -		ISM02.4	PPB	8/16/2017 11:45	SMPL
PB101570BL	PBW003	-65	-0.0071 -		ISM02.4	PPB	8/16/2017 11:47	SMPL
I4722-01	ME7GE2	-201	-0.0422 -		ISM02.4	PPB	8/16/2017 11:50	SMPL
I4722-02	ME7GE3	-267	-0.0593 -		ISM02.4	PPB	8/16/2017 11:52	SMPL
I4722-03	ME7GE2D	-201	-0.0422 -		ISM02.4	PPB	8/16/2017 11:54	SMPL
CCV083	CCV083	19620	5.0777 -		ISM02.4	PPB	8/16/2017 11:57	SMPL
CCB083	CCB083	-301	-0.068 -		ISM02.4	PPB	8/16/2017 11:59	SMPL
I4722-04	ME7GE2S	4111	1.0716 -		ISM02.4	PPB	8/16/2017 12:01	SMPL
CCV084	CCV084	19667	5.0898 -		ISM02.4	PPB	8/16/2017 12:03	SMPL
CCB084	CCB084	-305	-0.0691 -		ISM02.4	PPB	8/16/2017 12:06	SMPL