



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB89540

Temp.: 21.2 °C

SLOPE.: 98.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	08/16/2017	17:00
2	CAL2	1	Water			7.01	08/16/2017	17:01
3	CAL3	1	Water			10.03	08/16/2017	17:02
4	ICV	1	Water			7.00	08/16/2017	17:03
5	CCV1	1	Water			2.02	08/16/2017	17:05
6	I4848-01	1	Water			7.24	08/16/2017	17:06
7	I4848-01DUP	1	Water			7.25	08/16/2017	17:07
8	CCV2	1	Water			12.01	08/16/2017	17:08

2B 89540

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH 14848

WorkList ID : 101993

Date : 8/16/2017 4:42:57 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/16/2017	Water	14848-01-H	pH	Cool 4 deg C	PSEG01	M33	A2882	08/16/2017	9045C

Date/Time 08-16-17 4:41 PM
Received by: JS
Relinquished by: JS

Date/Time 08-16-17 6:30 PM
Received by: CP
Relinquished by: JD