



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB89550
Analysis Date: 08/17/2017
BalanceID: WC SC-3
OvenID: EX Oven-2

ANALYST: JIGNESH
REVIEWED BY: apatel
Extraction Date: 08/17/2017
Extraction IN Time: 08:30
Extraction OUT Time: 09:15

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB89550BL	LB89550BL	WATER	2	1000	100	3.0644	3.0644	3.02	3.0644	3.0644	0	0
2	LB89550BS	LB89550BS	WATER	2	1000	100	3.1477	3.1477	3.03	3.1667	3.1667	0.019	19
3	LB89550BSD	LB89550BSD	WATER	2	1000	100	3.0628	3.0628	3.04	3.0815	3.0815	0.0187	18.7
4	I4848-01	A2882	WATER	2	1000	100	3.1093	3.1093	3.03	3.5686	3.5686	0.4593	459.3



QC Batch# LB89550 Test: TPH Analysis Date: 08/17/2017

Chemicals Used:

Chemical Name	Chemical Lot #
Hexane	W2254
pH Paper 0-14	W2248
Sodium Sulfate	EP1762
1:1 HCL	WP54437
Silica Gel	W2199
Sand	

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP56784
LCSWD	10.00 ML	WP54439
MS/MSD		

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance: 0.0018 In OVEN TEMP : 71 Dessicator Time In : 10:25
1.0000 gram Balance: 1.0003 In Time: 09:40
Bal Check Time: 08:40

After Analysis

0.0020 gram Balance: 0.0021 Out OVEN TEMP: 70 Dessicator Time Out: 12:00
1.0000 gram Balance: 1.0005 Out Time: 11:25
Bal Check Time: 12:30

2B89550

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TPH I4848

WorkList ID : 102010

Date : 8/17/2017 8:25:39 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/21/2017	Water	I4848-01-5	TPH	Cool 4 deg C	PSEG01	M33	A2882	08/16/2017	1664A

Date/Time 08-17-17 8:25 AM
Received by: JD
Relinquished by: JD

Date/Time 08-17-17 10:00 AM
Received by: CR
Relinquished by: JD