


7470A

LB89552
INSTRUMENT ID: CV1

Linear

 μ Abs.:

35627



A= 0.0000e+000

B= 2.8136e-004

C= 8.4550e-003

Rho= 0.9999806

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.000	0.000	-29	0.000	-29				
0.2	0.200	0.219	0.019	748	0.0 %	748				
2.5	2.500	2.506	0.006	8877	0.0 %	8877				
5.0	5.000	4.965	-0.035	17618	0.0 %	17618				
7.5	7.500	7.477	-0.023	26545	0.0 %	26545				
10.0	10.000	10.032	0.032	35627	0.0 %	35627				

LB89552 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 17 Aug 2017 10:02:13

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-29	-	0.0000	7470A	PPB	17 Aug 2017 10:05:39	S	Std
0.2 - 1	80.2	748	-	0.2000	7470A	PPB	17 Aug 2017 10:07:47	S	Std
2.5 - 1	82.5	8877	-	2.5000	7470A	PPB	17 Aug 2017 10:09:55	S	Std
5.0 - 1	85	17618	-	5.0000	7470A	PPB	17 Aug 2017 10:12:04	S	Std
7.5 - 1	87.5	26545	-	7.5000	7470A	PPB	17 Aug 2017 10:16:36	S	Std
10.0 - 1	810	35627	-	10.0000	7470A	PPB	17 Aug 2017 10:21:08	S	Std
ICV16 - 1	ICV16	13362	3.7679	-	7470A	PPB	17 Aug 2017 10:23:36	U	SMPL
ICB16 - 1	ICB16	-76	-0.0129	-	7470A	PPB	17 Aug 2017 10:25:43	U	SMPL
CCV93 - 1	CCV93	17598	4.9598	-	7470A	PPB	17 Aug 2017 10:27:50	U	SMPL
CCB93 - 1	CCB93	-186	-0.0439	-	7470A	PPB	17 Aug 2017 10:29:56	U	SMPL
CRA - 1	CRA	644	0.1896	-	7470A	PPB	17 Aug 2017 10:32:04	U	SMPL
HighStd - 1	HighStd	36203	10.1944	-	7470A	PPB	17 Aug 2017 10:34:10	U	SMPL
ChkStd - 1	ChkStd	24654	6.9450	-	7470A	PPB	17 Aug 2017 10:36:17	U	SMPL
PB101611BL - 1	PBW	-246	-0.0608	-	7470A	PPB	17 Aug 2017 10:38:26	U	SMPL
PB101611BS - 1	LCSW	12986	3.6621	-	7470A	PPB	17 Aug 2017 10:40:35	U	SMPL
PB101611TB - 1	PB101556TB	-115	-0.0239	-	7470A	PPB	17 Aug 2017 10:42:42	U	SMPL
I4791-11 - 1	SB-27-COMP	-19	0.0031	-	7470A	PPB	17 Aug 2017 10:44:50	U	SMPL
I4791-12 - 1	SB-33-COMP	-49	-0.0053	-	7470A	PPB	17 Aug 2017 10:46:58	U	SMPL
I4791-13 - 1	SB-22-COMP	-93	-0.0177	-	7470A	PPB	17 Aug 2017 10:49:07	U	SMPL
I4791-14 - 1	SB-21-COMP	-103	-0.0205	-	7470A	PPB	17 Aug 2017 10:51:14	U	SMPL
CCV94 - 1	CCV94	17612	4.9637	-	7470A	PPB	17 Aug 2017 10:53:20	U	SMPL
CCB94 - 1	CCB94	-168	-0.0388	-	7470A	PPB	17 Aug 2017 10:55:27	U	SMPL
I4791-15 - 1	SB-19-COMP	-89	-0.0166	-	7470A	PPB	17 Aug 2017 10:57:35	U	SMPL
I4792-14 - 1	SB-06-COMP	-59	-0.0081	-	7470A	PPB	17 Aug 2017 10:59:41	U	SMPL
I4792-15 - 1	SB-05-COMP	-6	0.0068	-	7470A	PPB	17 Aug 2017 11:01:48	U	SMPL
I4792-16 - 1	SB-01-COMP	-9	0.0059	-	7470A	PPB	17 Aug 2017 11:03:55	U	SMPL
I4792-17 - 1	SB-02-COMP	-8	0.0062	-	7470A	PPB	17 Aug 2017 11:06:02	U	SMPL
I4792-18 - 1	SB-03-COMP	-103	-0.0205	-	7470A	PPB	17 Aug 2017 11:08:10	U	SMPL
I4792-19 - 1	SB-16-COMP	-92	-0.0174	-	7470A	PPB	17 Aug 2017 11:10:18	U	SMPL
I4799-03 - 1	AU-05-081417-A	-91	-0.0171	-	7470A	PPB	17 Aug 2017 11:12:27	U	SMPL
I4799-06 - 1	AU-05-081417-B	-101	-0.0200	-	7470A	PPB	17 Aug 2017 11:14:35	U	SMPL
I4814-02 - 1	WHITE-TWP-COMPOST	-86	-0.0157	-	7470A	PPB	17 Aug 2017 11:16:42	U	SMPL
CCV95 - 1	CCV95	17609	4.9628	-	7470A	PPB	17 Aug 2017 11:18:49	U	SMPL
CCB95 - 1	CCB95	-300	-0.0760	-	7470A	PPB	17 Aug 2017 11:20:56	U	SMPL
I4815-03 - 1	TR-05-081517-A	-222	-0.0540	-	7470A	PPB	17 Aug 2017 11:23:03	U	SMPL
I4815-06 - 1	TR-05-081517-B	-225	-0.0549	-	7470A	PPB	17 Aug 2017 11:25:10	U	SMPL
I4817-03 - 1	SU-03-081517	-211	-0.0509	-	7470A	PPB	17 Aug 2017 11:27:17	U	SMPL
I4825-02 - 1	TP-A-B-COMP	-230	-0.0563	-	7470A	PPB	17 Aug 2017 11:29:23	U	SMPL
I4825-02DUP - 1	TP-A-B-COMPDUP	-208	-0.0501	-	7470A	PPB	17 Aug 2017 11:31:31	U	SMPL
I4825-02MS - 1	TP-A-B-COMPMS	13159	3.7108	-	7470A	PPB	17 Aug 2017 11:33:38	U	SMPL
I4825-02MSD - 1	TP-A-B-COMPMSD	12901	3.6382	-	7470A	PPB	17 Aug 2017 11:35:47	U	SMPL
I4825-02LX5 - 1		-296	-0.0748	-	7470A	PPB	17 Aug 2017 11:37:55	U	SMPL
I4825-02A - 1		14028	3.9553	-	7470A	PPB	17 Aug 2017 11:40:04	U	SMPL
CCV96 - 1	CCV96	17815	5.0208	-	7470A	PPB	17 Aug 2017 11:42:11	U	SMPL
CCB96 - 1	CCB96	-320	-0.0816	-	7470A	PPB	17 Aug 2017 11:44:18	U	SMPL