


7471A

LB89554
INSTRUMENT ID: CV1

Linear

 μ Abs.:

41732



A= 0.0000e+000

B= 2.3951e-004

C= 1.8176e-002

Rho= 0.9999176

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.005	0.005	-57	0.000	-57				
0.2	0.200	0.234	0.034	903	0.0 %	903				
2.5	2.500	2.502	0.002	10372	0.0 %	10372				
5.0	5.000	4.900	-0.100	20382	0.0 %	20382				
7.5	7.500	7.545	0.045	31427	0.0 %	31427	0			
10.0	10.000	10.013	0.013	41732	0.0 %	41732				

LB89554 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 17 Aug 2017 11:53:01

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc/Method	Units	Date	Type	Type
0.0 - 1	SD	-57	-	0.00007471A	PPB	17 Aug 2017 11:54:24	S	Std
0.2 - 1	SD.2	903	-	0.20007471A	PPB	17 Aug 2017 11:56:41	S	Std
2.5 - 1	SD.2	10372	-	2.50007471A	PPB	17 Aug 2017 11:58:59	S	Std
5.0 - 1	SD	20382	-	5.00007471A	PPB	17 Aug 2017 12:01:17	S	Std
7.5 - 1	SD.2	31427	-	7.50007471A	PPB	17 Aug 2017 12:03:34	S	Std
10.0 - 1	SD	41732	-	10.00007471A	PPB	17 Aug 2017 12:05:51	S	Std
ICV17 - 1	ICV17	16704	4.0190	-7471A	PPB	17 Aug 2017 12:08:37	U	SMPL
ICB17 - 1	ICB17	-148	-0.0173	-7471A	PPB	17 Aug 2017 12:10:52	U	SMPL
CCV97 - 1	CCV97	20430	4.9114	-7471A	PPB	17 Aug 2017 12:13:08	U	SMPL
CCB97 - 1	CCB97	-152	-0.0182	-7471A	PPB	17 Aug 2017 12:15:24	U	SMPL
CRA - 1	CRA	873	0.2273	-7471A	PPB	17 Aug 2017 12:17:39	U	SMPL
HighStd - 1	HighStd	42320	10.1543	-7471A	PPB	17 Aug 2017 12:19:55	U	SMPL
ChkStd - 1	ChkStd	28765	6.9077	-7471A	PPB	17 Aug 2017 12:22:11	U	SMPL
PB101619BL - 1	PBS	-139	-0.0151	-7471A	PPB	17 Aug 2017 12:24:29	U	SMPL
PB101619BS - 1	LCSS	14875	3.5809	-7471A	PPB	17 Aug 2017 12:26:46	U	SMPL
I4838-01 - 1	SASP2P-TP-106-F(0-5)COMP	2245	0.5559	-7471A	PPB	17 Aug 2017 12:29:03	U	SMPL
I4844-01 - 1	OR-03-081617	2099	0.5209	-7471A	PPB	17 Aug 2017 12:31:20	U	SMPL
I4845-01 - 1	OK-03-81617	1187	0.3025	-7471A	PPB	17 Aug 2017 12:33:38	U	SMPL
I4847-01 - 1	PVSC-A-COMP	29217	7.0160	-7471A	PPB	17 Aug 2017 12:35:56	U	SMPL
I4847-06 - 1	PVSC-B-COMP	33105	7.9472	-7471A	PPB	17 Aug 2017 12:38:11	U	SMPL
CCV98 - 1	CCV98	20031	4.8158	-7471A	PPB	17 Aug 2017 12:40:29	U	SMPL
CCB98 - 1	CCB98	-204	-0.0307	-7471A	PPB	17 Aug 2017 12:42:46	U	SMPL
I4847-11 - 1	RO-152-ETGI	-142	-0.0158	-7471A	PPB	17 Aug 2017 12:45:01	U	SMPL
I4852-01 - 1	VNJ-207	9148	2.2092	-7471A	PPB	17 Aug 2017 12:47:17	U	SMPL
I4852-04 - 1	VNJ-242	2778	0.6835	-7471A	PPB	17 Aug 2017 12:49:33	U	SMPL
I4852-04DUP - 1	VNJ-242DUP	3120	0.7655	-7471A	PPB	17 Aug 2017 12:51:49	U	SMPL
I4852-04MS - 1	VNJ-242MS	18258	4.3912	-7471A	PPB	17 Aug 2017 12:54:06	U	SMPL
I4852-04MSD - 1	VNJ-242MSD	18353	4.4139	-7471A	PPB	17 Aug 2017 12:56:23	U	SMPL
I4852-04LX5 - 1		348	0.1015	-7471A	PPB	17 Aug 2017 12:58:41	U	SMPL
I4852-04A - 1		19021	4.5739	-7471A	PPB	17 Aug 2017 13:00:58	U	SMPL
CCV99 - 1	CCV99	20301	4.8805	-7471A	PPB	17 Aug 2017 13:03:16	U	SMPL
CCB99 - 1	CCB99	-96	-0.0048	-7471A	PPB	17 Aug 2017 13:05:32	U	SMPL