



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: apatel

Parameter: Flash Point

Supervisor Review By: AMANDEEP

Run Number: LB89585

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	28.0	82.4	08/17/2017	15:34
2	I4848-02		1	>100.0	>212.0	08/17/2017	15:50
3	I4848-02DUP		1	>100.0	>212.0	08/17/2017	16:02

Result = (Celsius * 1.8) + 32

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP081717

WorkList ID : 102022

Date : 8/17/2017 10:27:22 AM

Due Date	Matrix	Sample	Test	Flash Point	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/21/2017	Water	14848-02	DP	Flash Point	Cool 4 deg C	PSEG01	M33	A2882	08/16/2017	1010A

Date/Time 8/17/17 2:30 PM
 Received by: AS
 Relinquished by: AS

Date/Time 8/17/17 4:20 PM
 Received by: CP
 Relinquished by: AS