

245.1


LB89595

INSTRUMENT ID: CX1

Linear

 μ Abs.:

30692



A= 0.0000e+000

B= 3.2596e-004

C= 3.6436e-002

Rho= 0.9998854

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.016	0.016	-63	0.000	-63				
0.2	0.200	0.223	0.023	571	0.0 %	571				
2.5	2.500	2.529	0.029	7647	0.0 %	7647				
5.0	5.000	4.877	-0.123	14849	0.0 %	14849				
7.5	7.500	7.515	0.015	22944	0.0 %	22944				
10.0	10.000	10.041	0.041	30692	0.0 %	30692				

LB89595 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 18 Aug 2017 15:01:52

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-63	-	0.0000	245.1	PPB	18 Aug 2017 15:14:41	S	Std
0.2 - 1	50.2	571	-	0.2000	245.1	PPB	18 Aug 2017 15:16:49	S	Std
2.5 - 1	52.5	7647	-	2.5000	245.1	PPB	18 Aug 2017 15:18:57	S	Std
5.0 - 1	55	14849	-	5.0000	245.1	PPB	18 Aug 2017 15:21:06	S	Std
7.5 - 1	57.5	22944	-	7.5000	245.1	PPB	18 Aug 2017 15:23:13	S	Std
10.0 - 1	510	30692	-	10.0000	245.1	PPB	18 Aug 2017 15:25:21	S	Std
ICV19 - 1	ICV19	11883	3.9098	-245.1	PPB	PPB	18 Aug 2017 15:27:52	U	SMPL
ICB19 - 1	ICB19	-180	-0.0222	-245.1	PPB	PPB	18 Aug 2017 15:29:58	U	SMPL
CCV10 - 1	CCV10	15147	4.9737	-245.1	PPB	PPB	18 Aug 2017 15:32:06	U	SMPL
CCB10 - 1	CCB10	-318	-0.0672	-245.1	PPB	PPB	18 Aug 2017 15:34:12	U	SMPL
CRA - 1	CRA	451	0.1834	-245.1	PPB	PPB	18 Aug 2017 15:36:19	U	SMPL
HighStd - 1	HighStd	30640	10.0238	-245.1	PPB	PPB	18 Aug 2017 15:38:25	U	SMPL
ChkStd - 1	ChkStd	20309	6.6563	-245.1	PPB	PPB	18 Aug 2017 15:40:32	U	SMPL
PB101650BL - 1	PBW	-268	-0.0509	-245.1	PPB	PPB	18 Aug 2017 15:42:41	U	SMPL
PB101650BS - 1	LCSW	13265	4.3603	-245.1	PPB	PPB	18 Aug 2017 15:44:50	U	SMPL
I4792-11 - 1	TWP-01	-403	-0.0949	-245.1	PPB	PPB	18 Aug 2017 15:46:57	U	SMPL
I4792-11DUP - 1	TWP-01DUP	-312	-0.0653	-245.1	PPB	PPB	18 Aug 2017 15:49:05	U	SMPL
I4792-11MS - 1	TWP-01MS	13772	4.5255	-245.1	PPB	PPB	18 Aug 2017 15:51:14	U	SMPL
I4792-11MSD - 1	TWP-01MSD	12717	4.1816	-245.1	PPB	PPB	18 Aug 2017 15:53:22	U	SMPL
CCV11 - 1	CCV11	15252	5.0079	-245.1	PPB	PPB	18 Aug 2017 15:55:30	U	SMPL
CCB11 - 1	CCB11	-285	-0.0565	-245.1	PPB	PPB	18 Aug 2017 15:57:37	U	SMPL
I4792-11LX5 - 1		-211	-0.0323	-245.1	PPB	PPB	18 Aug 2017 15:59:44	U	SMPL
I4792-11A - 1		13530	4.4466	-245.1	PPB	PPB	18 Aug 2017 16:01:51	U	SMPL
CCV12 - 1	CCV12	15029	4.9353	-245.1	PPB	PPB	18 Aug 2017 16:03:57	U	SMPL
CCB12 - 1	CCB12	-298	-0.0607	-245.1	PPB	PPB	18 Aug 2017 16:15:16	U	SMPL