


7471A

LB89598
INSTRUMENT ID: CV2

 μ Abs.:

38834



Linear

A= 0.0000e+000

B= 2.5919e-004

C= 3.3051e-002

Rho= 0.9997554

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.001	0.001	-125	0.000	-125				
0.2	0.200	0.215	0.015	701	0.0 %	701				
2.5	2.500	2.515	0.015	9576	0.0 %	9576				
5.0	5.000	5.039	0.039	19315	0.0 %	19315				
7.5	7.500	7.332	-0.168	28160	0.0 %	28160	0			
10.0	10.000	10.098	0.098	38834	0.0 %	38834				

LB89598 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 18 Aug 2017 16:51:40

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc/Method	Units	Date	Type	Type
0.0 - 1	50	-125	-	0.00007471A	PPB	18 Aug 2017 16:53:58	S	Std
0.2 - 1	50.2	701	-	0.20007471A	PPB	18 Aug 2017 16:56:15	S	Std
2.5 - 1	52.5	9576	-	2.50007471A	PPB	18 Aug 2017 16:58:33	S	Std
5.0 - 1	55	19315	-	5.00007471A	PPB	18 Aug 2017 17:00:50	S	Std
7.5 - 1	57.5	28160	-	7.50007471A	PPB	18 Aug 2017 17:03:07	S	Std
10.0 - 1	510	38834	-	10.00007471A	PPB	18 Aug 2017 17:05:25	S	Std
ICV20 - 1	ICV20	15758	4.1174	-7471A	PPB	18 Aug 2017 17:08:04	U	SMPL
ICB20 - 1	ICB20	-210	-0.0214	-7471A	PPB	18 Aug 2017 17:10:20	U	SMPL
CCV13 - 1	CCV13	19412	5.0644	-7471A	PPB	18 Aug 2017 17:12:36	U	SMPL
CCB13 - 1	CCB13	-272	-0.0374	-7471A	PPB	18 Aug 2017 17:14:51	U	SMPL
CRA - 1	CRA	644	0.2000	-7471A	PPB	18 Aug 2017 17:17:07	U	SMPL
HighStd - 1	HighStd	38810	10.0922	-7471A	PPB	18 Aug 2017 17:19:23	U	SMPL
ChkStd - 1	ChkStd	27529	7.1683	-7471A	PPB	18 Aug 2017 17:21:39	U	SMPL
PB101671BL - 1	PBS	-387	-0.0673	-7471A	PPB	18 Aug 2017 17:23:57	U	SMPL
PB101671BS - 1	LCSS	14883	3.8906	-7471A	PPB	18 Aug 2017 17:26:14	U	SMPL
I4849-01 - 1	CAPTAINSCOVE-WC-002	3370	0.9065	-7471A	PPB	18 Aug 2017 17:28:31	U	SMPL
I4869-01 - 1	SASP2P-TP-106-A(0-5)COMP	2536	0.6904	-7471A	PPB	18 Aug 2017 17:30:48	U	SMPL
I4869-03 - 1	SASP2P-TP-106-B(0-5)COMP	9503	2.4961	-7471A	PPB	18 Aug 2017 17:33:06	U	SMPL
I4873-01 - 1	TEC-BE3-1-3-COMP	17094	4.4636	-7471A	PPB	18 Aug 2017 17:35:24	U	SMPL
I4873-04 - 1	TEC-BE3-3-10-COMP	1618	0.4524	-7471A	PPB	18 Aug 2017 17:37:40	U	SMPL
CCV14 - 1	CCV14	19269	5.0274	-7471A	PPB	18 Aug 2017 17:39:55	U	SMPL
CCB14 - 1	CCB14	-243	-0.0299	-7471A	PPB	18 Aug 2017 17:42:11	U	SMPL
I4875-01 - 1	CL-02-081717-A	4191	1.1193	-7471A	PPB	18 Aug 2017 17:44:27	U	SMPL
I4875-01DUP - 1	CL-02-081717-ADUP	4266	1.1388	-7471A	PPB	18 Aug 2017 17:46:43	U	SMPL
I4875-01MS - 1	CL-02-081717-AMS	20364	5.3112	-7471A	PPB	18 Aug 2017 17:48:59	U	SMPL
I4875-01MSD - 1	CL-02-081717-AMSD	20205	5.2700	-7471A	PPB	18 Aug 2017 17:51:16	U	SMPL
I4875-04 - 1	CL-02-081717-B	1988	0.5483	-7471A	PPB	18 Aug 2017 17:53:33	U	SMPL
I4875-07 - 1	CL-02-081717-C	1421	0.4014	-7471A	PPB	18 Aug 2017 17:55:50	U	SMPL
I4875-10 - 1	CL-02-081717-D	1378	0.3902	-7471A	PPB	18 Aug 2017 17:58:07	U	SMPL
I4875-13 - 1	CL-02-081717-E	6529	1.7253	-7471A	PPB	18 Aug 2017 18:00:25	U	SMPL
I4879-01 - 1	6706-AA-081717	4032	1.0781	-7471A	PPB	18 Aug 2017 18:02:43	U	SMPL
CCV15 - 1	CCV15	19361	5.0512	-7471A	PPB	18 Aug 2017 18:04:59	U	SMPL
CCB15 - 1	CCB15	-270	-0.0369	-7471A	PPB	18 Aug 2017 18:07:15	U	SMPL
I4875-01LX5 - 1		676	0.2083	-7471A	PPB	18 Aug 2017 18:09:32	U	SMPL
I4875-01A - 1		20670	5.3905	-7471A	PPB	18 Aug 2017 18:11:48	U	SMPL
CCV16 - 1	CCV16	19298	5.0349	-7471A	PPB	18 Aug 2017 18:14:04	U	SMPL
CCB16 - 1	CCB16	-306	-0.0463	-7471A	PPB	18 Aug 2017 18:16:22	U	SMPL