



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB89605

Temp.: 23.3 °C

BalanceID: WC SC-4

SLOPE.: 98.5

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	08/18/2017	16:10
2	CAL2	1	Water			7.00	08/18/2017	16:11
3	CAL3	1	Water			10.03	08/18/2017	16:12
4	ICV	1	Water			7.00	08/18/2017	16:13
5	CCV1	1	Water			2.01	08/18/2017	16:15
6	I4869-05	1	Solid	20.02	20	8.54	08/18/2017	16:17
7	I4869-05DUP	1	Solid	20.03	20	8.55	08/18/2017	16:18
8	I4869-06	1	Solid	20.05	20	8.43	08/18/2017	16:20
9	I4889-05	1	Solid	20.04	20	8.17	08/18/2017	16:21
10	I4892-03	1	Solid	20.02	20	8.58	08/18/2017	16:22
11	I4898-03	1	Solid	20.03	20	9.00	08/18/2017	16:25
12	I4898-06	1	Solid	20.01	20	7.63	08/18/2017	16:26
13	I4900-16	1	Solid	20.02	20	6.60	08/18/2017	16:27
14	I4900-18	1	Solid	20.03	20	6.50	08/18/2017	16:30
15	CCV2	1	Water			12.01	08/18/2017	16:31

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I4892

WorkList ID : 102101

Date : 8/18/2017 3:53:31 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/22/2017	Solid	I4869-05	Corrosivity	Cool 4 deg C	THEP04	L11	SASP2P-TP-106-A(0-5)COMF	08/16/2017	9045C
08/22/2017	Solid	I4869-06	Corrosivity	Cool 4 deg C	THEP04	L11	SASP2P-TP-106-B(0-5)COMF	08/16/2017	9045C
	Solid	I4889-05	Corrosivity	Cool 4 deg C	THEP04	N41	SASP2P-TP-106-L(0-5)COMP	08/17/2017	9045C
08/23/2017	Solid	I4892-03	Corrosivity	Cool 4 deg C	PSEG05	N11	HR-06-081817	08/18/2017	9045C
08/23/2017	Solid	I4898-03	Corrosivity	Cool 4 deg C	PSEG05	N11	NB-08-081817-A	08/18/2017	9045C
08/23/2017	Solid	I4898-06	Corrosivity	Cool 4 deg C	PSEG05	N11	NB-08-081817-B	08/18/2017	9045C
08/25/2017	Solid	I4900-16	Corrosivity	Cool 4 deg C	PSEG01	M43	220-1-STONE	08/18/2017	9045C
08/25/2017	Solid	I4900-18	Corrosivity	Cool 4 deg C	PSEG01	M43	220-2-STONE	08/18/2017	9045C

Date/Time 08/18/17 4:00 PM
 Received by: JP
 Relinquished by: JP

Date/Time 08/18/17 7:00 AM
 Received by: CP
 Relinquished by: JP