


7470A

LB89618
INSTRUMENT ID: CV1

Linear

 μ Abs.:

35349



A= 0.0000e+000

B= 2.8459e-004

C= 5.8028e-002

Rho= 0.9997001

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.002	-0.002	-212	0.000	-212				
0.2	0.200	0.232	0.032	612	0.0 %	612				
2.5	2.500	2.584	0.084	8876	0.0 %	8876				
5.0	5.000	4.882	-0.118	16952	0.0 %	16952				
7.5	7.500	7.385	-0.115	25747	0.0 %	25747				
10.0	10.000	10.118	0.118	35349	0.0 %	35349				

LB89618 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 21 Aug 2017 11:54:57

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-212	-	0.0000	7470A	PPB	21 Aug 2017 12:12:16	S	Std
0.2 - 1	50.2	612	-	0.2000	7470A	PPB	21 Aug 2017 12:14:24	S	Std
2.5 - 1	50.5	8876	-	2.5000	7470A	PPB	21 Aug 2017 12:16:33	S	Std
5.0 - 1	55	16952	-	5.0000	7470A	PPB	21 Aug 2017 12:18:41	S	Std
7.5 - 1	57.5	25747	-	7.5000	7470A	PPB	21 Aug 2017 12:20:49	S	Std
10.0 - 1	510	35349	-	10.0000	7470A	PPB	21 Aug 2017 12:22:56	S	Std
ICV21 - 1	ICV21	13059	3.7745	-	7470A	PPB	21 Aug 2017 12:25:35	U	SMPL
ICB21 - 1	ICB21	-332	-0.0365	-	7470A	PPB	21 Aug 2017 12:27:41	U	SMPL
CCV17 - 1	CCV17	16822	4.8455	-	7470A	PPB	21 Aug 2017 12:29:48	U	SMPL
CCB17 - 1	CCB17	-346	-0.0404	-	7470A	PPB	21 Aug 2017 12:31:55	U	SMPL
CRA - 1	CRA	594	0.2271	-	7470A	PPB	21 Aug 2017 12:34:02	U	SMPL
HighStd - 1	HighStd	34882	9.9852	-	7470A	PPB	21 Aug 2017 12:36:09	U	SMPL
ChkStd - 1	ChkStd	23545	6.7588	-	7470A	PPB	21 Aug 2017 12:38:16	U	SMPL
PB101684BL - 1	PBW	-291	-0.0248	-	7470A	PPB	21 Aug 2017 12:43:01	U	SMPL
PB101684BS - 1	LCSW	13782	3.9803	-	7470A	PPB	21 Aug 2017 12:45:08	U	SMPL
PB101684TB - 1	PB101653TB	-331	-0.0362	-	7470A	PPB	21 Aug 2017 12:47:16	U	SMPL
I4849-05 - 1	CAPTAINSCOVE-WC-002	-270	-0.0188	-	7470A	PPB	21 Aug 2017 12:49:24	U	SMPL
I4875-03 - 1	CL-02-081717-A	-255	-0.0145	-	7470A	PPB	21 Aug 2017 12:51:32	U	SMPL
I4875-06 - 1	CL-02-081717-B	-239	-0.0100	-	7470A	PPB	21 Aug 2017 12:53:41	U	SMPL
I4875-09 - 1	CL-02-081717-C	-229	-0.0071	-	7470A	PPB	21 Aug 2017 12:55:48	U	SMPL
CCV18 - 1	CCV18	16638	4.7931	-	7470A	PPB	21 Aug 2017 12:57:55	U	SMPL
CCB18 - 1	CCB18	-366	-0.0461	-	7470A	PPB	21 Aug 2017 13:00:02	U	SMPL
I4875-12 - 1	CL-02-081717-D	-249	-0.0128	-	7470A	PPB	21 Aug 2017 13:02:08	U	SMPL
I4875-15 - 1	CL-02-081717-E	-241	-0.0106	-	7470A	PPB	21 Aug 2017 13:04:15	U	SMPL
I4879-03 - 1	6706-AA-081717	-244	-0.0114	-	7470A	PPB	21 Aug 2017 13:06:21	U	SMPL
I4879-03DUP - 1	6706-AA-081717DUP	-224	-0.0057	-	7470A	PPB	21 Aug 2017 13:08:28	U	SMPL
I4879-03MS - 1	6706-AA-081717MS	13972	4.0344	-	7470A	PPB	21 Aug 2017 13:10:36	U	SMPL
I4879-03MSD - 1	6706-AA-081717MSD	13896	4.0127	-	7470A	PPB	21 Aug 2017 13:12:44	U	SMPL
PB101685BL - 1	PBW	-306	-0.0291	-	7470A	PPB	21 Aug 2017 13:14:52	U	SMPL
PB101685BS - 1	LCSW	14349	4.1417	-	7470A	PPB	21 Aug 2017 13:17:00	U	SMPL
I4873-05 - 1	TEC-BIMW	2459	0.7578	-	7470A	PPB	21 Aug 2017 13:19:09	U	SMPL
I4873-05DUP - 1	TEC-BIMWDUP	2201	0.6844	-	7470A	PPB	21 Aug 2017 13:21:16	U	SMPL
CCV19 - 1	CCV19	17389	5.0068	-	7470A	PPB	21 Aug 2017 13:23:23	U	SMPL
CCB19 - 1	CCB19	-326	-0.0347	-	7470A	PPB	21 Aug 2017 13:25:30	U	SMPL
I4873-05MS - 1	TEC-BIMWMS	18881	5.4314	-	7470A	PPB	21 Aug 2017 13:27:38	U	SMPL
I4873-05MSD - 1	TEC-BIMWMSD	18001	5.1810	-	7470A	PPB	21 Aug 2017 13:29:45	U	SMPL
I4873-06 - 1	TEC-BIMW	-506	-0.0860	-	7470A	PPB	21 Aug 2017 13:31:52	U	SMPL
I4877-01 - 1	VL-WWTP-005	118	0.0916	-	7470A	PPB	21 Aug 2017 13:34:00	U	SMPL
I4879-03LX5 - 1		-207	-0.0009	-	7470A	PPB	21 Aug 2017 13:36:08	U	SMPL
I4879-03A - 1		14511	4.1878	-	7470A	PPB	21 Aug 2017 13:38:15	U	SMPL
I4873-05LX5 - 1		156	0.1024	-	7470A	PPB	21 Aug 2017 13:40:24	U	SMPL
I4873-05A - 1		18994	5.4636	-	7470A	PPB	21 Aug 2017 13:42:33	U	SMPL
CCV20 - 1	CCV20	17458	5.0265	-	7470A	PPB	21 Aug 2017 13:44:42	U	SMPL
CCB20 - 1	CCB20	-345	-0.0402	-	7470A	PPB	21 Aug 2017 13:46:49	U	SMPL