

TitriSoft Report

Ident	Amount	DateTime	Alkalinity	Initial pH (sample)	slope	offset
LB89623BLW	100	21/08/2017 13:40	0.00	3.74	104.2	6.95
LB89623BSW	100	21/08/2017 13:55	47.76	8.42	104.2	6.97
I4832-02	50	21/08/2017 14:10	218.80	6.69	104.2	6.97
I4832-05	50	21/08/2017 14:14	224.80	7.08	104.2	6.97
I4832-06	50	21/08/2017 14:20	180.96	6.83	104.2	6.97
I4832-07	50	21/08/2017 14:25	211.92	7.07	104.2	6.97
I4832-08	50	21/08/2017 14:30	204.64	6.89	104.2	6.97
I4832-09	50	21/08/2017 14:36	207.28	7.20	104.2	6.97
I4832-10	50	21/08/2017 14:41	211.92	7.14	104.2	6.97
I4832-11	50	21/08/2017 14:47	208.40	7.18	104.2	6.97
I4832-12	50	21/08/2017 14:52	217.28	7.28	104.2	6.97
I4832-13	50	21/08/2017 14:58	229.84	7.25	104.2	6.97
I4832-14	50	21/08/2017 15:03	233.68	7.18	104.2	6.97
I4833-01	50	21/08/2017 15:08	205.76	6.98	104.2	6.97
I4833-02	50	21/08/2017 15:14	192.72	6.73	104.2	6.97
I4835-01	50	21/08/2017 15:21	46.40	6.40	104.2	6.97
I4835-02	50	21/08/2017 15:26	16.48	5.34	104.2	6.97
I4835-03	50	21/08/2017 15:29	93.12	7.59	104.2	6.97
I4836-01	50	21/08/2017 15:33	235.28	6.55	104.2	6.97
I4836-02	50	21/08/2017 15:38	230.80	6.60	104.2	6.97
I4836-03	50	21/08/2017 15:43	104.08	6.56	104.2	6.97
I4836-03DUP	50	21/08/2017 15:53	102.16	6.42	104.2	6.97
I4835-02	100	21/08/2017 16:07	16.64	5.42	104.2	6.97

08/21/17 IB