


7471A

L389624
INSTRUMENT ID: CV1

Linear

 μ Abs.:

4177E



A= 0.0000e+000

B= 2.3776e-004

C= -1.2715e-002

Rho= 0.9998677

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.017	-0.017	-19	0.000	-19				
0.2	0.200	0.210	0.010	935	0.0 %	935				
2.5	2.500	2.484	-0.016	10501	0.0 %	10501				
5.0	5.000	4.984	-0.016	21017	0.0 %	21017				
7.5	7.500	7.620	0.120	32101	0.0 %	32101	0			
10.0	10.000	9.920	-0.080	41776	0.0 %	41776				

LB89624 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 21 Aug 2017 15:00:41

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-19	-	0.00007471A	PPB		21 Aug 2017 15:04:25	S	Std
0.2 - 1	50.2	935	-	0.20007471A	PPB		21 Aug 2017 15:06:43	S	Std
2.5 - 1	50.5	10501	-	2.50007471A	PPB		21 Aug 2017 15:09:00	S	Std
5.0 - 1	55	21017	-	5.00007471A	PPB		21 Aug 2017 15:11:18	S	Std
7.5 - 1	57.5	32101	-	7.50007471A	PPB		21 Aug 2017 15:13:34	S	Std
10.0 - 1	510	41776	-	10.00007471A	PPB		21 Aug 2017 15:21:06	S	Std
ICV22 - 1	ICV22	16302	3.8632	-7471A	PPB		21 Aug 2017 15:23:52	U	SMPL
ICB22 - 1	ICB22	-244	-0.0707	-7471A	PPB		21 Aug 2017 15:26:07	U	SMPL
CCV21 - 1	CCV21	20952	4.9688	-7471A	PPB		21 Aug 2017 15:28:23	U	SMPL
CCB21 - 1	CCB21	-265	-0.0757	-7471A	PPB		21 Aug 2017 15:30:39	U	SMPL
CRA - 1	CRA	930	0.2084	-7471A	PPB		21 Aug 2017 15:32:54	U	SMPL
HighStd - 1	HighStd	42496	10.0910	-7471A	PPB		21 Aug 2017 15:35:10	U	SMPL
ChkStd - 1	ChkStd	27733	6.5810	-7471A	PPB		21 Aug 2017 15:37:26	U	SMPL
PB101696BL - 1	PBS	-335	-0.0924	-7471A	PPB		21 Aug 2017 15:39:44	U	SMPL
PB101696BS - 1	LCSS	15085	3.5739	-7471A	PPB		21 Aug 2017 15:42:01	U	SMPL
I4881-01 - 1	RWB-10	12867	3.0465	-7471A	PPB		21 Aug 2017 15:44:18	U	SMPL
I4881-03 - 1	RWB-9	9206	2.1761	-7471A	PPB		21 Aug 2017 15:46:35	U	SMPL
I4889-01 - 1	SASP2P-TP-106-L(0-5)COMP	5563	1.3099	-7471A	PPB		21 Aug 2017 15:48:53	U	SMPL
I4889-03 - 1	SASP2P-TP-106-E(0-5)COMP	7141	1.6851	-7471A	PPB		21 Aug 2017 15:51:11	U	SMPL
I4890-01 - 1	CLEAN-SAND-6	135	0.0194	-7471A	PPB		21 Aug 2017 15:53:27	U	SMPL
CCV22 - 1	CCV22	21144	5.0144	-7471A	PPB		21 Aug 2017 15:55:43	U	SMPL
CCB22 - 1	CCB22	-311	-0.0867	-7471A	PPB		21 Aug 2017 15:57:59	U	SMPL
I4890-01DUP - 1	CLEAN-SAND-6DUP	179	0.0298	-7471A	PPB		21 Aug 2017 16:00:15	U	SMPL
I4890-01MS - 1	CLEAN-SAND-6MS	15642	3.7063	-7471A	PPB		21 Aug 2017 16:02:31	U	SMPL
I4890-01MSD - 1	CLEAN-SAND-6MSD	15364	3.6402	-7471A	PPB		21 Aug 2017 16:04:47	U	SMPL
I4890-02 - 1	CLEAN-SAND-7	185	0.0313	-7471A	PPB		21 Aug 2017 16:07:03	U	SMPL
I4892-01 - 1	HR-06-081817	1308	0.2983	-7471A	PPB		21 Aug 2017 16:09:20	U	SMPL
I4898-01 - 1	NB-08-081817-A	2195	0.5092	-7471A	PPB		21 Aug 2017 16:11:37	U	SMPL
I4898-04 - 1	NB-08-081817-B	1334	0.3045	-7471A	PPB		21 Aug 2017 16:13:55	U	SMPL
I4900-01 - 1	CONTROL-HOUSE-RAMP	16	-0.0089	-7471A	PPB		21 Aug 2017 16:16:12	U	SMPL
I4900-03 - 1	CONTROL-HOUSE-STAIR	0	-0.0127	-7471A	PPB		21 Aug 2017 16:18:30	U	SMPL
I4900-05 - 1	CONTROL-LOADING-DOCK	-289	-0.0814	-7471A	PPB		21 Aug 2017 16:20:46	U	SMPL
CCV23 - 1	CCV23	19817	4.6989	-7471A	PPB		21 Aug 2017 16:23:02	U	SMPL
CCB23 - 1	CCB23	-265	-0.0757	-7471A	PPB		21 Aug 2017 16:25:19	U	SMPL
I4900-07 - 1	220-1-WALL	93	0.0094	-7471A	PPB		21 Aug 2017 16:27:35	U	SMPL
I4900-09 - 1	220-1-CONCRETE	101	0.0113	-7471A	PPB		21 Aug 2017 16:29:51	U	SMPL
I4900-11 - 1	DEAD MEN	139	0.0203	-7471A	PPB		21 Aug 2017 16:32:08	U	SMPL
I4900-13 - 1	TRENWA	44	-0.0023	-7471A	PPB		21 Aug 2017 16:34:24	U	SMPL
I4900-15 - 1	220-1-STONE	285	0.0550	-7471A	PPB		21 Aug 2017 16:36:41	U	SMPL
I4900-17 - 1	220-2-STONE	3131	0.7317	-7471A	PPB		21 Aug 2017 16:38:58	U	SMPL
I4900-19 - 1	220-2-CONCRETE	137	0.0199	-7471A	PPB		21 Aug 2017 16:41:15	U	SMPL
I4900-21 - 1	PUMP-HOUSE	19	-0.0082	-7471A	PPB		21 Aug 2017 16:43:33	U	SMPL
I4890-01LX5 - 1		-40	-0.0222	-7471A	PPB		21 Aug 2017 16:45:50	U	SMPL
I4890-01A - 1		16619	3.9386	-7471A	PPB		21 Aug 2017 16:48:07	U	SMPL
CCV24 - 1	CCV24	20563	4.8763	-7471A	PPB		21 Aug 2017 16:50:23	U	SMPL
CCB24 - 1	CCB24	-200	-0.0603	-7471A	PPB		21 Aug 2017 16:52:39	U	SMPL