



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB89633

Temp.: 21.4 °C

BalanceID: WC SC-4

SLOPE.: 98.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	08/21/2017	15:30
2	CAL2	1	Water			7.00	08/21/2017	15:31
3	CAL3	1	Water			10.04	08/21/2017	15:33
4	ICV	1	Water			6.99	08/21/2017	15:34
5	CCV1	1	Water			2.01	08/21/2017	15:35
6	I4908-03	1	Solid	20.02	20	7.10	08/21/2017	15:36
7	I4908-03DUP	1	Solid	20.04	20	7.11	08/21/2017	15:37
8	I4908-06	1	Solid	20.01	20	7.60	08/21/2017	15:38
9	I4909-03	1	Solid	20.03	20	8.46	08/21/2017	15:39
10	I4914-03	1	Solid	20.00	20	7.02	08/21/2017	15:40
11	I4914-06	1	Solid	20.02	20	6.34	08/21/2017	15:41
12	I4915-03	1	Solid	20.04	20	8.77	08/21/2017	15:43
13	I4915-06	1	Solid	20.02	20	6.44	08/21/2017	15:45
14	CCV2	1	Water			12.01	08/21/2017	15:46

1389633

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I4909 WorkList ID : 102153 Date : 8/21/2017 4:28:51 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/24/2017	Solid	I4908-03	Corrosivity		PSEG05		TR-04-082117-A	08/21/2017	9045C
08/24/2017	Solid	I4908-06	Corrosivity		PSEG05		TR-04-082117-B	08/21/2017	9045C
08/24/2017	Solid	I4909-03	Corrosivity		PSEG05		SU-04-82117	08/21/2017	9045C
08/24/2017	Solid	I4914-03	Corrosivity	Cool 4 deg C	PSEG01		289-COMP	08/21/2017	9045C
08/24/2017	Solid	I4914-06	Corrosivity	Cool 4 deg C	PSEG01		165-COMP	08/21/2017	9045C
08/24/2017	Solid	I4915-03	Corrosivity	Cool 4 deg C	PSEG01	M53	72-11991-GRAVEL-COMP	08/21/2017	9045C
08/24/2017	Solid	I4915-06	Corrosivity	Cool 4 deg C	PSEG01	M53	72-11991-DEBRIS-COMP	08/21/2017	9045C

Date/Time 08/21/17 6:32
Received by: JP
Relinquished by: HL

Date/Time 08/21/17 7:00 PM
Received by: HL
Relinquished by: JP