



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB89654

Temp.: 21.8 °C

SLOPE.: 98.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	08/22/2017	16:10
2	CAL2	1	Water			7.01	08/22/2017	16:11
3	CAL3	1	Water			10.02	08/22/2017	16:12
4	ICV	1	Water			7.00	08/22/2017	16:15
5	CCV1	1	Water			2.01	08/22/2017	16:16
6	I4923-01	1	Water			6.98	08/22/2017	16:17
7	I4923-01DUP	1	Water			6.99	08/22/2017	16:18
8	I4923-02	1	Water			6.98	08/22/2017	16:19
9	CCV2	1	Water			12.01	08/22/2017	16:21

2389654

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 14923

WorkList ID : 102194

Date : 8/22/2017 4:01:44 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	14923-01-1	pH	Cool 4 deg C	LIRO01	N41	SB04-GW	08/22/2017	SM 4500-PH B
	Water	14923-02-1	pH	Cool 4 deg C	LIRO01	N41	SB01-GW	08/22/2017	SM 4500-PH B

Date/Time 08/22/17 4:01 PM
Received by: [Signature]
Relinquished by: [Signature]

Date/Time 08/22/17 5:00 PM
Received by: [Signature]
Relinquished by: [Signature]