


7470A

LB 89 664
INSTRUMENT ID: CV1

Linear

 μ Abs.:

34063



A= 0.0000e+000

B= 2.9217e-004

C= 5.1157e-002

Rho= 0.9999714

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.019	0.019	-111	0.000	-111				
0.2	0.200	0.217	0.017	569	0.0 %	569				
2.5	2.500	2.440	-0.060	8175	0.0 %	8175				
5.0	5.000	5.019	0.019	17003	0.0 %	17003				
7.5	7.500	7.502	0.002	25501	0.0 %	25501				
10.0	10.000	10.003	0.003	34063	0.0 %	34063				

LB89664 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 23 Aug 2017 10:14:46

Sample ID	Extended ID	µ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	-111	-	0.00007470A	PPB	23 Aug 2017 10:23:12	S	Std
0.2 - 1	50.2	569	-	0.20007470A	PPB	23 Aug 2017 10:25:21	S	Std
2.5 - 1	52.5	8175	-	2.50007470A	PPB	23 Aug 2017 10:27:29	S	Std
5.0 - 1	55	17003	-	5.00007470A	PPB	23 Aug 2017 10:29:37	S	Std
7.5 - 1	57.5	25501	-	7.50007470A	PPB	23 Aug 2017 10:31:45	S	Std
10.0 - 1	510	34063	-	10.00007470A	PPB	23 Aug 2017 10:33:53	S	Std
ICV25 - 1	ICV25	13306	3.9388	-7470A	PPB	23 Aug 2017 10:36:24	U	SMPL
ICB25 - 1	ICB25	-179	-0.0011	-7470A	PPB	23 Aug 2017 10:38:31	U	SMPL
CCV35 - 1	CCV35	17154	5.0631	-7470A	PPB	23 Aug 2017 10:40:38	U	SMPL
CCB35 - 1	CCB35	-281	-0.0309	-7470A	PPB	23 Aug 2017 10:42:45	U	SMPL
CRA - 1	CRA	516	0.2019	-7470A	PPB	23 Aug 2017 10:44:51	U	SMPL
HighStd - 1	HighStd	34336	10.0832	-7470A	PPB	23 Aug 2017 10:46:58	U	SMPL
ChkStd - 1	ChkStd	23587	6.9426	-7470A	PPB	23 Aug 2017 10:49:05	U	SMPL
PB101752BL - 1	PBW	-352	-0.0517	-7470A	PPB	23 Aug 2017 10:51:14	U	SMPL
PB101752BS - 1	LCSW	12953	3.8357	-7470A	PPB	23 Aug 2017 10:53:22	U	SMPL
PB101752TB - 1	PB101730TB	-191	-0.0046	-7470A	PPB	23 Aug 2017 10:55:29	U	SMPL
I4907-01 - 1	NYSEG-PH4-ESMI-1A	-97	0.0228	-7470A	PPB	23 Aug 2017 10:57:37	U	SMPL
I4908-03 - 1	TR-04-082117-A	-150	0.0073	-7470A	PPB	23 Aug 2017 10:59:45	U	SMPL
I4908-06 - 1	TR-04-082117-B	-199	-0.0070	-7470A	PPB	23 Aug 2017 11:01:54	U	SMPL
I4909-03 - 1	SU-04-082117	-171	0.0012	-7470A	PPB	23 Aug 2017 11:04:01	U	SMPL
CCV36 - 1	CCV36	17225	5.0838	-7470A	PPB	23 Aug 2017 11:06:08	U	SMPL
CCB36 - 1	CCB36	-272	-0.0283	-7470A	PPB	23 Aug 2017 11:08:15	U	SMPL
I4914-03 - 1	289-COMP	-180	-0.0014	-7470A	PPB	23 Aug 2017 11:10:23	U	SMPL
I4914-06 - 1	165-COMP	-161	0.0041	-7470A	PPB	23 Aug 2017 11:12:30	U	SMPL
I4915-03 - 1	72-11991-GRAVEL-COMP	-147	0.0082	-7470A	PPB	23 Aug 2017 11:14:36	U	SMPL
I4915-03DUP - 1	72-11991-GRAVEL-COMPDUP	-151	0.0070	-7470A	PPB	23 Aug 2017 11:16:43	U	SMPL
I4915-03MS - 1	72-11991-GRAVEL-COMPMS	12888	3.8167	-7470A	PPB	23 Aug 2017 11:18:51	U	SMPL
I4915-03MSD - 1	72-11991-GRAVEL-COMPMSD	13363	3.9555	-7470A	PPB	23 Aug 2017 11:20:59	U	SMPL
I4915-06 - 1	72-11991-DEBRIS-COMP	-181	-0.0017	-7470A	PPB	23 Aug 2017 11:23:07	U	SMPL
I4915-03LX5 - 1		-126	0.0143	-7470A	PPB	23 Aug 2017 11:25:16	U	SMPL
I4915-03A - 1		14545	4.3008	-7470A	PPB	23 Aug 2017 11:27:24	U	SMPL
CCV37 - 1	CCV37	17260	5.0940	-7470A	PPB	23 Aug 2017 11:29:31	U	SMPL
CCB37 - 1	CCB37	-252	-0.0225	-7470A	PPB	23 Aug 2017 11:31:39	U	SMPL