

ISM02.4


LB89672

INSTRUMENT ID: CV1

Linear

 μ Abs.:

43131



A= 0.0000e+000

B= 2.3295e-004

C= -2.4087e-002

Rho= 0.9998840

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.045	-0.045	-90	0.000	-90					-
0.2	0.200	0.225	0.025	1070	0.0 %	1070					13
2.5	2.500	2.504	0.004	10852	0.0 %	10852					0
5.0	5.000	5.085	0.085	21931	0.0 %	21931					2
7.5	7.500	7.408	-0.092	31902	0.0 %	31902					-1
10.0	10.000	10.023	0.023	43131	0.0 %	43131					0

NB
8/22/17

LB89672
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	-90	-		0 ISM02.4	PPB	8/23/2017 14:12	Std
	0.2 S0.2	1070	-		0.2 ISM02.4	PPB	8/23/2017 14:14	Std
	2.5 S2.5	10852	-		2.5 ISM02.4	PPB	8/23/2017 14:16	Std
	5 S5	21931	-		5 ISM02.4	PPB	8/23/2017 14:20	Std
	7.5 S7.5	31902	-		7.5 ISM02.4	PPB	8/23/2017 14:23	Std
	10 S10	43131	-		10 ISM02.4	PPB	8/23/2017 14:28	Std
ICV	ICV	17450	4.041 -		ISM02.4	PPB	8/23/2017 14:30	SMPL
ICB	ICB	-182	-0.0665 -		ISM02.4	PPB	8/23/2017 14:33	SMPL
CCV038	CCV038	22402	5.1946 -		ISM02.4	PPB	8/23/2017 14:35	SMPL
CCB038	CCB038	-261	-0.0849 -		ISM02.4	PPB	8/23/2017 14:37	SMPL
PB101754BL	PBS004	-117	-0.0513 -		ISM02.4	PPB	8/23/2017 14:39	SMPL
I4868-03	MDAGH3	823	0.1676 -		ISM02.4	PPB	8/23/2017 14:49	SMPL
I4868-04	MDAGH4	1403	0.3027 -		ISM02.4	PPB	8/23/2017 14:51	SMPL
I4868-05	MDAGH5	6586	1.5102 -		ISM02.4	PPB	8/23/2017 14:53	SMPL
I4868-06	MDAGH6	5120	1.1686 -		ISM02.4	PPB	8/23/2017 14:56	SMPL
I4868-07	MDAGH7	1346	0.2895 -		ISM02.4	PPB	8/23/2017 14:58	SMPL
I4868-08	MDAGH8	8326	1.9155 -		ISM02.4	PPB	8/23/2017 15:00	SMPL
I4868-09	MDAGH9	1188	0.2527 -		ISM02.4	PPB	8/23/2017 15:03	SMPL
I4868-10	MDAGH9D	1191	0.2534 -		ISM02.4	PPB	8/23/2017 15:05	SMPL
I4868-11	MDAGH9S	13237	3.0595 -		ISM02.4	PPB	8/23/2017 15:07	SMPL
I4868-12	MDAGJ0	928	0.1921 -		ISM02.4	PPB	8/23/2017 15:09	SMPL
I4868-01	MDAGH2	419088	HIGH -		ISM02.4	PPB	8/23/2017 15:12	
I4868-02	MDAGH1	12182	2.8138 -		ISM02.4	PPB	8/23/2017 15:16	SMPL
CCV039	CCV039	22662	5.2551 -		ISM02.4	PPB	8/23/2017 15:20	SMPL
CCB039	CCB039	-308	-0.0958 -		ISM02.4	PPB	8/23/2017 15:23	SMPL
I4868-01DLX10	MDAGH2	60158	13.99 -		ISM02.4	PPB	8/23/2017 15:27	SMPL
I4868-01DLX50	MDAGH2	12213	2.821 -		ISM02.4	PPB	8/23/2017 15:37	SMPL
CCV040	CCV040	22888	5.3078 -		ISM02.4	PPB	8/23/2017 15:39	SMPL
CCB040	CCB040	-274	-0.0879 -		ISM02.4	PPB	8/23/2017 15:41	SMPL