



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : AMANDEEP

Run Number: LB89680

Temp.: 22.2 °C

BalanceID: WC SC-4

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	08/23/2017	16:20
2	CAL2	1	Water			7.00	08/23/2017	16:21
3	CAL3	1	Water			10.03	08/23/2017	16:22
4	ICV	1	Water			7.00	08/23/2017	16:23
5	CCV1	1	Water			2.01	08/23/2017	16:25
6	I4935-01	1	Solid	20.02	20	6.77	08/23/2017	16:26
7	I4935-01DUP	1	Solid	20.04	20	6.78	08/23/2017	16:27
8	I4938-01	1	Solid	20.05	20	8.52	08/23/2017	16:30
9	CCV2	1	Water			12.01	08/23/2017	16:31

LB 89680

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH 14935

WorkList ID : 102232

Date : 8/23/2017 4:02:06 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I4935-01	pH	Cool 4 deg C	TULL02	N11	FK-PS-01-014	08/23/2017	9045C
	Solid	I4938-01	pH	Cool 4 deg C	TULL02	N11	IRON-LEAF-TOPSOIL	08/22/2017	9045C

Date/Time 08/23/17 11:02A
Received by: JD
Relinquished by: JD

Date/Time 08/23/17 5:30P
Received by: CD
Relinquished by: JD