


7471A

LB89688
INSTRUMENT ID: CV2

Linear

 μ Abs.:

41770



A= 0.0000e+000

B= 2.3857e-004

C= 9.3581e-004

Rho= 0.9998751

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.048	-0.048	-207	0.000	-207				
0.2	0.200	0.213	0.013	887	0.0 %	887				
2.5	2.500	2.580	0.080	10810	0.0 %	10810				
5.0	5.000	4.923	-0.077	20632	0.0 %	20632				
7.5	7.500	7.567	0.067	31713	0.0 %	31713	0			
10.0	10.000	9.966	-0.034	41770	0.0 %	41770				

LB89683 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 24 Aug 2017 09:39:43

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SO	-207	-	0.0000	7471A	PPB	24 Aug 2017 10:43:19	S	Std
0.2 - 1	SO-2	887	-	0.2000	7471A	PPB	24 Aug 2017 10:45:36	S	Std
2.5 - 1	SO-5	10810	-	2.5000	7471A	PPB	24 Aug 2017 10:47:54	S	Std
5.0 - 1	SO-5	20632	-	5.0000	7471A	PPB	24 Aug 2017 10:50:11	S	Std
7.5 - 1	SO-5	31713	-	7.5000	7471A	PPB	24 Aug 2017 10:52:28	S	Std
10.0 - 1	SO-5	41770	-	10.0000	7471A	PPB	24 Aug 2017 10:54:44	S	Std
ICV26 - 1	ICV26	17212	4.1072	-	7471A	PPB	24 Aug 2017 10:57:45	U	SMPL
ICB26 - 1	ICB26	-354	-0.0835	-	7471A	PPB	24 Aug 2017 11:00:01	U	SMPL
CCV41 - 1	CCV41	21523	5.1357	-	7471A	PPB	24 Aug 2017 11:02:17	U	SMPL
CCB41 - 1	CCB41	-251	-0.0589	-	7471A	PPB	24 Aug 2017 11:04:33	U	SMPL
CRA - 1	CRA	977	0.2340	-	7471A	PPB	24 Aug 2017 11:06:48	U	SMPL
HighStd - 1	HighStd	42133	10.0527	-	7471A	PPB	24 Aug 2017 11:09:04	U	SMPL
ChkStd - 1	ChkStd	28093	6.7031	-	7471A	PPB	24 Aug 2017 11:11:20	U	SMPL
PB101779BL - 1	PBS	-319	-0.0752	-	7471A	PPB	24 Aug 2017 11:13:38	U	SMPL
PB101779BS - 1	LCSS	17124	4.0862	-	7471A	PPB	24 Aug 2017 11:15:56	U	SMPL
I4919-02 - 1	SASP2P-TP-106-W(0-5)COMP	5940	1.4181	-	7471A	PPB	24 Aug 2017 11:18:13	U	SMPL
I4920-01 - 1	FK-SF-03-002-A	1552	0.3712	-	7471A	PPB	24 Aug 2017 11:20:30	U	SMPL
I4920-01DUP - 1	FK-SF-03-002-ADUP	1506	0.3602	-	7471A	PPB	24 Aug 2017 11:22:48	U	SMPL
I4920-01MS - 1	FK-SF-03-002-AMS	17363	4.1433	-	7471A	PPB	24 Aug 2017 11:25:05	U	SMPL
I4920-01MSD - 1	FK-SF-03-002-AMSD	16538	3.9464	-	7471A	PPB	24 Aug 2017 11:27:22	U	SMPL
CCV42 - 1	CCV42	21416	5.1102	-	7471A	PPB	24 Aug 2017 11:29:38	U	SMPL
CCB42 - 1	CCB42	-297	-0.0699	-	7471A	PPB	24 Aug 2017 11:31:54	U	SMPL
I4920-03 - 1	FK-SF-03-002-B	1351	0.3232	-	7471A	PPB	24 Aug 2017 11:34:12	U	SMPL
I4922-01 - 1	SA-01-082217	1676	0.4008	-	7471A	PPB	24 Aug 2017 11:36:28	U	SMPL
I4924-01 - 1	72-11975-COMP	44864	10.7042	-	7471A	PPB	24 Aug 2017 11:38:44	U	SMPL
I4925-01 - 1	RWB-6	7793	1.8601	-	7471A	PPB	24 Aug 2017 11:41:01	U	SMPL
I4925-03 - 1	RWB-1	7666	1.8298	-	7471A	PPB	24 Aug 2017 11:43:19	U	SMPL
I4925-05 - 1	RWB-2	13407	3.1995	-	7471A	PPB	24 Aug 2017 11:45:36	U	SMPL
I4925-07 - 1	RWB-7	15043	3.5898	-	7471A	PPB	24 Aug 2017 11:47:53	U	SMPL
I4925-09 - 1	RWB-8	7661	1.8286	-	7471A	PPB	24 Aug 2017 11:50:10	U	SMPL
I4925-12 - 1	RWB-5	6685	1.5958	-	7471A	PPB	24 Aug 2017 11:52:29	U	SMPL
I4925-13 - 1	RWB-4	8670	2.0694	-	7471A	PPB	24 Aug 2017 11:54:45	U	SMPL
CCV43 - 1	CCV43	21526	5.1364	-	7471A	PPB	24 Aug 2017 11:57:01	U	SMPL
CCB43 - 1	CCB43	-302	-0.0711	-	7471A	PPB	24 Aug 2017 11:59:18	U	SMPL
I4925-15 - 1	RWB-3	12480	2.9783	-	7471A	PPB	24 Aug 2017 12:01:35	U	SMPL
I4929-02 - 1	SASP2P-TP-106-S(0-5)COMP	326	0.0787	-	7471A	PPB	24 Aug 2017 12:03:51	U	SMPL
I4930-01DLX10 - 1	1708015690	515	0.1238	-	7471A	PPB	24 Aug 2017 12:06:07	U	SMPL
I4935-01 - 1	FK-PS-01-014	283	0.0685	-	7471A	PPB	24 Aug 2017 12:08:24	U	SMPL
I4938-01 - 1	IRON-LEAF-TOPSOIL	3455	0.8252	-	7471A	PPB	24 Aug 2017 12:10:41	U	SMPL
I4920-01LX5 - 1		167	0.0408	-	7471A	PPB	24 Aug 2017 12:12:57	U	SMPL
I4920-01A - 1		18492	4.4126	-	7471A	PPB	24 Aug 2017 12:15:15	U	SMPL
I4924-01DLX2 - 1		21751	5.1901	-	7471A	PPB	24 Aug 2017 12:17:33	U	SMPL
CCV44 - 1	CCV44	20986	5.0076	-	7471A	PPB	24 Aug 2017 12:19:50	U	SMPL
CCB44 - 1	CCB44	-294	-0.0692	-	7471A	PPB	24 Aug 2017 12:22:08	U	SMPL