



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : AMANDEEP

Run Number: LB89691

Temp.: 21.0 °C

SLOPE.: 97.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	08/24/2017	12:00
2	CAL2	1	Water			7.01	08/24/2017	12:01
3	CAL3	1	Water			10.02	08/24/2017	12:02
4	ICV	1	Water			6.99	08/24/2017	12:03
5	CCV1	1	Water			2.01	08/24/2017	12:05
6	I4951-01	1	Water			5.93	08/24/2017	12:05
7	I4951-01DUP	1	Water			5.94	08/24/2017	12:06
8	CCV2	1	Water			12.01	08/24/2017	12:07

2389691

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I4951

WorkList ID : 102257

Date : 8/24/2017 11:43:12 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I4951-01-A	pH	Cool 4 deg C	STER01	O11	POOLWATER	08/23/2017	SM 4500-PH B

Date/Time 08/24/17 11:43 AM
Received by: JD
Relinquished by: JD

Date/Time 08/23/17 11:00 AM
Received by: CG CD
Relinquished by: JD