


7470A

LB89692
INSTRUMENT ID: CV1

Linear

 μ Abs.:

34753



A= 0.0000e+000

B= 2.8820e-004

C= -1.0657e-002

Rho= 0.9997206

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.056	-0.056	-158	0.000	-158				
0.2	0.200	0.218	0.018	793	0.0 %	793				
2.5	2.500	2.487	-0.013	8666	0.0 %	8666				
5.0	5.000	5.166	0.166	17962	0.0 %	17962				
7.5	7.500	7.380	-0.120	25644	0.0 %	25644				
10.0	10.000	10.005	0.005	34753	0.0 %	34753				

LB89692 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 24 Aug 2017 12:35:21

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-158	-	0.00007470A	PPB		24 Aug 2017 12:36:53	S	Std
0.2 - 1	50.2	793	-	0.20007470A	PPB		24 Aug 2017 12:39:02	S	Std
2.5 - 1	52.5	8666	-	2.50007470A	PPB		24 Aug 2017 12:41:10	S	Std
5.0 - 1	55	17962	-	5.00007470A	PPB		24 Aug 2017 12:43:19	S	Std
7.5 - 1	57.5	25644	-	7.50007470A	PPB		24 Aug 2017 12:45:26	S	Std
10.0 - 1	510	34753	-	10.00007470A	PPB		24 Aug 2017 12:47:34	S	Std
ICV27 - 1	ICV27	13678	3.9314	-7470A	PPB		24 Aug 2017 12:50:14	U	SMPL
ICB27 - 1	ICB27	-234	-0.0781	-7470A	PPB		24 Aug 2017 12:52:21	U	SMPL
CCV45 - 1	CCV45	17811	5.1225	-7470A	PPB		24 Aug 2017 12:54:27	U	SMPL
CCB45 - 1	CCB45	-269	-0.0882	-7470A	PPB		24 Aug 2017 12:56:34	U	SMPL
CRA - 1	CRA	769	0.2110	-7470A	PPB		24 Aug 2017 12:58:41	U	SMPL
HighStd - 1	HighStd	34827	10.0266	-7470A	PPB		24 Aug 2017 13:00:47	U	SMPL
ChkStd - 1	ChkStd	23676	6.8129	-7470A	PPB		24 Aug 2017 13:02:54	U	SMPL
PB101768BL - 1	PBW	-178	-0.0620	-7470A	PPB		24 Aug 2017 13:07:30	U	SMPL
PB101768BS - 1	LCSW	12719	3.6550	-7470A	PPB		24 Aug 2017 13:09:37	U	SMPL
PB101768TB - 1	PB101751TB	-215	-0.0726	-7470A	PPB		24 Aug 2017 13:11:45	U	SMPL
I4912-02 - 1	FENCE-POSTS-PILED-COMP	-184	-0.0637	-7470A	PPB		24 Aug 2017 13:13:53	U	SMPL
I4922-03 - 1	SA-01-082217	-165	-0.0582	-7470A	PPB		24 Aug 2017 13:16:01	U	SMPL
I4922-03DUP - 1	SA-01-082217DUP	-181	-0.0628	-7470A	PPB		24 Aug 2017 13:18:10	U	SMPL
I4922-03MS - 1	SA-01-082217MS	13255	3.8095	-7470A	PPB		24 Aug 2017 13:20:17	U	SMPL
CCV46 - 1	CCV46	17602	5.0623	-7470A	PPB		24 Aug 2017 13:22:24	U	SMPL
CCB46 - 1	CCB46	-258	-0.0850	-7470A	PPB		24 Aug 2017 13:24:30	U	SMPL
I4922-03MSD - 1	SA-01-082217MSD	13054	3.7516	-7470A	PPB		24 Aug 2017 13:26:37	U	SMPL
PB101781BL - 1	PBW	-361	-0.1147	-7470A	PPB		24 Aug 2017 13:28:44	U	SMPL
PB101781BS - 1	LCSW	13106	3.7665	-7470A	PPB		24 Aug 2017 13:30:51	U	SMPL
I4932-03 - 1	FT-07M	-189	-0.0651	-7470A	PPB		24 Aug 2017 13:32:58	U	SMPL
I4932-03DUP - 1	FT-07MDUP	-149	-0.0536	-7470A	PPB		24 Aug 2017 13:35:05	U	SMPL
I4932-03MS - 1	FT-07MMS	13088	3.7614	-7470A	PPB		24 Aug 2017 13:37:13	U	SMPL
I4932-03MSD - 1	FT-07MMSD	12645	3.6337	-7470A	PPB		24 Aug 2017 13:39:21	U	SMPL
I4932-04 - 1	FT-09M	-480	-0.1490	-7470A	PPB		24 Aug 2017 13:41:30	U	SMPL
CCV47 - 1	CCV47	17781	5.1139	-7470A	PPB		24 Aug 2017 13:43:38	U	SMPL
CCB47 - 1	CCB47	-254	-0.0839	-7470A	PPB		24 Aug 2017 13:45:45	U	SMPL
I4922-03LX5 - 1		-182	-0.0631	-7470A	PPB		24 Aug 2017 13:47:53	U	SMPL
I4922-03A - 1		14643	4.2095	-7470A	PPB		24 Aug 2017 13:50:00	U	SMPL
I4932-03LX5 - 1		-193	-0.0663	-7470A	PPB		24 Aug 2017 13:52:07	U	SMPL
I4932-03A - 1		13753	3.9530	-7470A	PPB		24 Aug 2017 13:54:14	U	SMPL
CCV48 - 1	CCV48	17779	5.1133	-7470A	PPB		24 Aug 2017 13:56:20	U	SMPL
CCB48 - 1	CCB48	-282	-0.0919	-7470A	PPB		24 Aug 2017 13:58:28	U	SMPL