

245.1


LB89699

INSTRUMENT ID: CV4

Linear

 μ Abs.:

30892



A= 0.0000e+000

B= 3.2329e-004

C= 2.3927e-002

Rho= 0.9999866

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.017	-0.017	-126	0.000	-126				
0.2	0.200	0.230	0.030	638	0.0 %	638				
2.5	2.500	2.481	-0.019	7599	0.0 %	7599				
5.0	5.000	5.013	0.013	15432	0.0 %	15432				
7.5	7.500	7.482	-0.018	23069	0.0 %	23069				
10.0	10.000	10.011	0.011	30892	0.0 %	30892				

LB89699 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 24 Aug 2017 16:08:52

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-126	-	0.0000	245.1	PPB	24 Aug 2017 16:32:58	S	Std
0.2 - 1	50.2	638	-	0.2000	245.1	PPB	24 Aug 2017 16:35:07	S	Std
2.5 - 1	52.5	7599	-	2.5000	245.1	PPB	24 Aug 2017 16:37:15	S	Std
5.0 - 1	55	15432	-	5.0000	245.1	PPB	24 Aug 2017 16:39:23	S	Std
7.5 - 1	57.5	23069	-	7.5000	245.1	PPB	24 Aug 2017 16:41:31	S	Std
10.0 - 1	510	30892	-	10.0000	245.1	PPB	24 Aug 2017 16:43:39	S	Std
ICV28 - 1	ICV28	11908	3.8737	-245.1		PPB	24 Aug 2017 16:46:14	U	SMPL
ICB28 - 1	ICB28	-273	-0.0643	-245.1		PPB	24 Aug 2017 16:48:20	U	SMPL
CCV49 - 1	CCV49	15516	5.0401	-245.1		PPB	24 Aug 2017 16:50:26	U	SMPL
CCB49 - 1	CCB49	-278	-0.0659	-245.1		PPB	24 Aug 2017 16:52:33	U	SMPL
CRA - 1	CRA	601	0.2182	-245.1		PPB	24 Aug 2017 16:54:40	U	SMPL
HighStd - 1	HighStd	30590	9.9134	-245.1		PPB	24 Aug 2017 16:56:47	U	SMPL
ChkStd - 1	ChkStd	21449	6.9582	-245.1		PPB	24 Aug 2017 16:58:53	U	SMPL
PB101769BL - 1	PBW	-297	-0.0721	-245.1		PPB	24 Aug 2017 17:01:03	U	SMPL
PB101769BS - 1	LCSW	13682	4.4472	-245.1		PPB	24 Aug 2017 17:03:11	U	SMPL
I4906-02 - 1	METALS	-251	-0.0572	-245.1		PPB	24 Aug 2017 17:05:18	U	SMPL
I4923-01 - 1	SB04-GW	-349	-0.0889	-245.1		PPB	24 Aug 2017 17:07:27	U	SMPL
I4923-02 - 1	SB01-GW	-216	-0.0459	-245.1		PPB	24 Aug 2017 17:09:35	U	SMPL
I4923-02DUP - 1	SB01-GWDUP	-307	-0.0753	-245.1		PPB	24 Aug 2017 17:11:43	U	SMPL
I4923-02MS - 1	SB01-GWMS	14979	4.8665	-245.1		PPB	24 Aug 2017 17:13:50	U	SMPL
CCV50 - 1	CCV50	15242	4.9515	-245.1		PPB	24 Aug 2017 17:15:56	U	SMPL
CCB50 - 1	CCB50	-289	-0.0695	-245.1		PPB	24 Aug 2017 17:18:03	U	SMPL
I4923-02MSD - 1	SB01-GWMSD	15383	4.9971	-245.1		PPB	24 Aug 2017 17:20:10	U	SMPL
I4923-02LX5 - 1		-262	-0.0608	-245.1		PPB	24 Aug 2017 17:22:17	U	SMPL
I4923-02A - 1		12448	4.0483	-245.1		PPB	24 Aug 2017 17:24:24	U	SMPL
CCV51 - 1	CCV51	15489	5.0314	-245.1		PPB	24 Aug 2017 17:26:31	U	SMPL
CCB51 - 1	CCB51	-304	-0.0744	-245.1		PPB	24 Aug 2017 17:28:38	U	SMPL