

7470A


LB 89705

INSTRUMENT ID: CV1

Linear

 μ Abs.:

35261



A= 0.0000e+000

B= 2.8344e-004

C= -2.7751e-002

Rho= 0.9999498

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.031	-0.031	-10	0.000	-10				
0.2	0.200	0.180	-0.020	733	0.0 %	733				
2.5	2.500	2.525	0.025	9005	0.0 %	9005				
5.0	5.000	5.071	0.071	17988	0.0 %	17988				
7.5	7.500	7.488	-0.012	26517	0.0 %	26517				
10.0	10.000	9.967	-0.033	35261	0.0 %	35261				

LB89705 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 25 Aug 2017 10:08:46

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-10	-	0.0000	7470A	PPB	25 Aug 2017 10:27:10	S	Std
0.2 - 1	50.2	733	-	0.2000	7470A	PPB	25 Aug 2017 10:29:19	S	Std
2.5 - 1	52.5	9005	-	2.5000	7470A	PPB	25 Aug 2017 10:31:27	S	Std
5.0 - 1	55	17988	-	5.0000	7470A	PPB	25 Aug 2017 10:33:35	S	Std
7.5 - 1	57.5	26517	-	7.5000	7470A	PPB	25 Aug 2017 10:35:43	S	Std
10.0 - 1	50	35261	-	10.0000	7470A	PPB	25 Aug 2017 10:40:18	S	Std
ICV29 - 1	ICV29	13764	3.8736	-7470A	PPB	25 Aug 2017 10:42:51	U	SMPL	
ICB29 - 1	ICB29	-70	-0.0476	-7470A	PPB	25 Aug 2017 10:47:21	U	SMPL	
CCV52 - 1	CCV52	17612	4.9643	-7470A	PPB	25 Aug 2017 10:49:27	U	SMPL	
CCB52 - 1	CCB52	-199	-0.0842	-7470A	PPB	25 Aug 2017 10:51:34	U	SMPL	
CRA - 1	CRA	729	0.1789	-7470A	PPB	25 Aug 2017 10:53:41	U	SMPL	
HighStd - 1	HighStd	35591	10.0603	-7470A	PPB	25 Aug 2017 10:55:47	U	SMPL	
ChkStd - 1	ChkStd	25446	7.1848	-7470A	PPB	25 Aug 2017 10:57:54	U	SMPL	
PB101805BL - 1	PBW	-196	-0.0833	-7470A	PPB	25 Aug 2017 11:00:03	U	SMPL	
PB101805BS - 1	LCWS	12893	3.6267	-7470A	PPB	25 Aug 2017 11:02:11	U	SMPL	
PB101805TB - 1	PB101766TB	-211	-0.0876	-7470A	PPB	25 Aug 2017 11:04:19	U	SMPL	
I4924-03 - 1	72-11975-COMP	-157	-0.0723	-7470A	PPB	25 Aug 2017 11:06:27	U	SMPL	
I4924-07 - 1	72-11975-DEBRIS	-96	-0.0550	-7470A	PPB	25 Aug 2017 11:08:35	U	SMPL	
I4934-11 - 1	SB5-SOIL-0-7.0	-129	-0.0643	-7470A	PPB	25 Aug 2017 11:10:44	U	SMPL	
I4934-12 - 1	SB4-SOIL-0-9.0	-82	-0.0510	-7470A	PPB	25 Aug 2017 11:12:50	U	SMPL	
CCV53 - 1	CCV53	17783	5.0127	-7470A	PPB	25 Aug 2017 11:14:57	U	SMPL	
CCB53 - 1	CCB53	-194	-0.0827	-7470A	PPB	25 Aug 2017 11:17:04	U	SMPL	
I4934-13 - 1	SB3-SOIL-0-6.0	-135	-0.0660	-7470A	PPB	25 Aug 2017 11:19:12	U	SMPL	
I4934-14 - 1	SB2-SOIL-0-6.0	-91	-0.0535	-7470A	PPB	25 Aug 2017 11:21:19	U	SMPL	
I4934-15 - 1	SB1-SOIL-0-2.0	-73	-0.0484	-7470A	PPB	25 Aug 2017 11:23:25	U	SMPL	
I4943-03 - 1	OR-01-082317	-153	-0.0711	-7470A	PPB	25 Aug 2017 11:25:32	U	SMPL	
I4944-02 - 1	4-1-DRUMS-COMP	-126	-0.0635	-7470A	PPB	25 Aug 2017 11:27:40	U	SMPL	
I4945-10 - 1	SP-1-4-COMP	-121	-0.0620	-7470A	PPB	25 Aug 2017 11:29:47	U	SMPL	
I4946-03 - 1	VNJ-211	-101	-0.0564	-7470A	PPB	25 Aug 2017 11:31:56	U	SMPL	
I4947-05 - 1	ELIZABETH-RO-COMP-A	-82	-0.0510	-7470A	PPB	25 Aug 2017 11:34:04	U	SMPL	
I4947-10 - 1	ELIZABETH-RO-COMP-B	-107	-0.0581	-7470A	PPB	25 Aug 2017 11:36:13	U	SMPL	
I4947-10DUP - 1	ELIZABETH-RO-COMP-BDUP	-77	-0.0496	-7470A	PPB	25 Aug 2017 11:38:20	U	SMPL	
CCV54 - 1	CCV54	18002	5.0748	-7470A	PPB	25 Aug 2017 11:40:26	U	SMPL	
CCB54 - 1	CCB54	-211	-0.0876	-7470A	PPB	25 Aug 2017 11:42:33	U	SMPL	
I4947-10MS - 1	ELIZABETH-RO-COMP-BMS	13223	3.7202	-7470A	PPB	25 Aug 2017 11:44:40	U	SMPL	
I4947-10MSD - 1	ELIZABETH-RO-COMP-BMSD	13272	3.7341	-7470A	PPB	25 Aug 2017 11:46:47	U	SMPL	
I4947-10LX5 - 1		-203	-0.0853	-7470A	PPB	25 Aug 2017 11:48:54	U	SMPL	
I4947-10A - 1		14652	4.1253	-7470A	PPB	25 Aug 2017 11:51:02	U	SMPL	
CCV55 - 1	CCV55	18026	5.0816	-7470A	PPB	25 Aug 2017 11:53:09	U	SMPL	
CCB55 - 1	CCB55	-188	-0.0810	-7470A	PPB	25 Aug 2017 11:55:16	U	SMPL	