



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB89706

Temp.: 21.4 °C

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	08/25/2017	10:30
2	CAL2	1	Water			7.01	08/25/2017	10:31
3	CAL3	1	Water			10.02	08/25/2017	10:32
4	ICV	1	Water			7.00	08/25/2017	10:33
5	CCV1	1	Water			2.01	08/25/2017	10:34
6	I4970-01	1	Water			7.47	08/25/2017	10:35
7	I4970-01DUP	1	Water			7.48	08/25/2017	10:36
8	CCV2	1	Water			12.01	08/25/2017	10:37

LB89706

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH 14970

WorkList ID : 102283

Date : 8/25/2017 10:20:43 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/24/2017	Water	14970-01-K	pH	Cool 4 deg C	PSEG01	O32	HAMILTON-FRAC-TANK	08/24/2017	9045C

Date/Time 08/25/17 10:20 AM
Received by: JP
Relinquished by: CL

Date/Time 08/25/17 1:25 PM
Received by: CL
Relinquished by: JP