



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB89707

Temp.: 21.4 °C

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	08/25/2017	10:30
2	CAL2	1	Water			7.01	08/25/2017	10:31
3	CAL3	1	Water			10.02	08/25/2017	10:32
4	ICV	1	Water			7.00	08/25/2017	10:33
5	CCV1	1	Water			2.01	08/25/2017	10:34
6	I4970-02	1	Water			7.47	08/25/2017	10:35
7	I4970-02DUP	1	Water			7.48	08/25/2017	10:36
8	CCV2	1	Water			12.01	08/25/2017	10:37

2889767

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY I4970 WorkList ID : 102284 Date : 8/25/2017 10:21:23 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/24/2017	Water	I4970-02-K	Corrosivity	Cool 4 deg C	PSEG01	O32	HAMILTON-FRAC-TANK	08/24/2017	9040C

Date/Time 08/25/17 10:20 AM
 Received by: JO
 Relinquished by: JP

Date/Time 08/25/17 11:25 AM
 Received by: CP
 Relinquished by: JP