



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB89735

Reagent/Standard	Lot/Loq #
TURBIDITY ICV 10 NTU	WP57521
TURBIDITY CCV 10 NTU	WP57522
Turbidity std, 40NTU	WP57515
turbidity std 5 NTU	WP57518
turbidity std 20 NTU	WP57516
Turbidity std 1 NTU	WP57519
turbidity std 0 NTU	WP57520
turbidity std 10 NTU	WP57517
Turbidity std, 80NTU	WP57523

Intercept: 0.0374

Slope: 0.9903

Regression: 1

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.099	08/26/2017	09:01
2	CAL2	1	1	1.049	08/26/2017	09:04
3	CAL3	5	1	4.953	08/26/2017	09:10
4	CAL4	10	1	9.951	08/26/2017	09:15
5	CAL5	20	1	19.600	08/26/2017	09:20
6	CAL6	40	1	39.900	08/26/2017	09:24
7	CAL7	80	1	79.200	08/26/2017	09:26



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB89735

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.200	9.252	08/26/2017	09:27
2	ICB		1	0.095	0.058	08/26/2017	09:30
3	CCV1	10	1	9.890	9.949	08/26/2017	09:33
4	CCB1		1	0.096	0.059	08/26/2017	09:37
5	LB89735BL		1	0.098	0.061	08/26/2017	09:40
6	I4998-01		1	18.010	18.149	08/26/2017	09:42
7	I4998-01DUP		1	18.010	18.149	08/26/2017	09:44
8	I4998-02		1	11.010	11.080	08/26/2017	09:46
9	I4998-03		1	8.320	8.364	08/26/2017	09:48
10	I4998-04		1	20.210	20.370	08/26/2017	09:50
11	I4998-05		1	9.170	9.222	08/26/2017	09:53
12	I4998-06		1	4.460	4.466	08/26/2017	09:55
13	I4998-07		1	3.730	3.729	08/26/2017	09:57
14	I4998-08		1	4.570	4.577	08/26/2017	09:59
15	CCV2	10	1	9.900	9.959	08/26/2017	10:03
16	CCB2		1	0.101	0.064	08/26/2017	10:06
17	I4998-09		1	5.250	5.264	08/26/2017	10:07
18	I4998-10		1	50.000	50.452	08/26/2017	10:10
19	I4998-11		1	9.210	9.262	08/26/2017	10:13
20	I4998-12		1	6.410	6.435	08/26/2017	10:16
21	I4998-13		1	6.190	6.213	08/26/2017	10:18
22	I4998-14		1	6.950	6.980	08/26/2017	10:21
23	I4998-15		1	34.100	34.396	08/26/2017	10:24
24	I4998-16		1	38.900	39.243	08/26/2017	10:27
25	I4998-17		1	35.500	35.810	08/26/2017	10:30
26	I4998-18		1	10.800	10.868	08/26/2017	10:33
27	CCV3	10	1	9.200	9.252	08/26/2017	10:35
28	CCB3		1	0.099	0.062	08/26/2017	10:38
29	I4998-19		1	41.800	42.172	08/26/2017	10:41
30	I4998-20		1	44.000	44.393	08/26/2017	10:45
31	CCV4	10	1	9.500	9.555	08/26/2017	10:47
32	CCB4		1	0.103	0.066	08/26/2017	10:50

LB89735

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TUR082617

WorkList ID : 102318

Date : 8/26/2017 8:41:01 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I4998-01	Turbidity	Cool 4 deg C	HDR103	O52	20170825-105236	08/25/2017	SM2130 B
	Water	I4998-02	Turbidity	Cool 4 deg C	HDR103	O52	20170825-105253	08/25/2017	SM2130 B
	Water	I4998-03	Turbidity	Cool 4 deg C	HDR103	O52	20170825-105311	08/25/2017	SM2130 B
	Water	I4998-04	Turbidity	Cool 4 deg C	HDR103	O52	20170825-105429	08/25/2017	SM2130 B
	Water	I4998-05	Turbidity	Cool 4 deg C	HDR103	O52	20170825-105444	08/25/2017	SM2130 B
	Water	I4998-06	Turbidity	Cool 4 deg C	HDR103	O52	20170825-105514	08/25/2017	SM2130 B
	Water	I4998-07	Turbidity	Cool 4 deg C	HDR103	O52	20170825-105619	08/25/2017	SM2130 B
	Water	I4998-08	Turbidity	Cool 4 deg C	HDR103	O52	20170825-105641	08/25/2017	SM2130 B
	Water	I4998-09	Turbidity	Cool 4 deg C	HDR103	O52	20170825-105657	08/25/2017	SM2130 B
	Water	I4998-10	Turbidity	Cool 4 deg C	HDR103	O52	20170825-105331	08/25/2017	SM2130 B
	Water	I4998-11	Turbidity	Cool 4 deg C	HDR103	O52	20170825-112726	08/25/2017	SM2130 B
	Water	I4998-12	Turbidity	Cool 4 deg C	HDR103	O52	20170825-112749	08/25/2017	SM2130 B
	Water	I4998-13	Turbidity	Cool 4 deg C	HDR103	O52	20170825-11286	08/25/2017	SM2130 B
	Water	I4998-14	Turbidity	Cool 4 deg C	HDR103	O52	20170825-112821	08/25/2017	SM2130 B
	Water	I4998-15	Turbidity	Cool 4 deg C	HDR103	O52	20170825-13045	08/25/2017	SM2130 B
	Water	I4998-16	Turbidity	Cool 4 deg C	HDR103	O52	20170825-1316	08/25/2017	SM2130 B
	Water	I4998-17	Turbidity	Cool 4 deg C	HDR103	O52	20170825-13126	08/25/2017	SM2130 B
	Water	I4998-18	Turbidity	Cool 4 deg C	HDR103	O52	20170825-13227	08/25/2017	SM2130 B
	Water	I4998-19	Turbidity	Cool 4 deg C	HDR103	O52	20170825-13249	08/25/2017	SM2130 B
	Water	I4998-20	Turbidity	Cool 4 deg C	HDR103	O52	20170825-13311	08/25/2017	SM2130 B

Date/Time 8/26/17 8:45

Received by: AK

Relinquished by: JP

Date/Time

Received by: JP

Relinquished by: AK

8/26/17 13:30