



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB89736

Reagent/Standard	Lot/Loq #
TURBIDITY ICV 10 NTU	WP57521
TURBIDITY CCV 10 NTU	WP57522
Turbidity std, 40NTU	WP57515
turbidity std 5 NTU	WP57518
turbidity std 20 NTU	WP57516
Turbidity std 1 NTU	WP57519
turbidity std 0 NTU	WP57520
turbidity std 10 NTU	WP57517
Turbidity std, 80NTU	WP57523

Intercept: 0.0374

Slope: 0.9903

Regression: 1

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.099	08/26/2017	09:01
2	CAL2	1	1	1.049	08/26/2017	09:04
3	CAL3	5	1	4.953	08/26/2017	09:10
4	CAL4	10	1	9.951	08/26/2017	09:15
5	CAL5	20	1	19.600	08/26/2017	09:20
6	CAL6	40	1	39.900	08/26/2017	09:24
7	CAL7	80	1	79.200	08/26/2017	09:26



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB89736

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.200	9.252	08/26/2017	09:27
2	ICB		1	0.095	0.058	08/26/2017	09:30
3	CCV1	10	1	9.600	9.656	08/26/2017	12:06
4	CCB1		1	0.099	0.062	08/26/2017	10:08
5	LB89736BL		1	0.101	0.064	08/26/2017	12:10
6	I4999-01		1	16.400	16.523	08/26/2017	12:13
7	I4999-01DUP		1	16.100	16.220	08/26/2017	12:15
8	I4999-02		1	22.700	22.885	08/26/2017	12:17
9	I4999-03		1	56.100	56.612	08/26/2017	12:19
10	I4999-04		1	43.900	44.292	08/26/2017	12:21
11	I4999-05		1	39.100	39.445	08/26/2017	12:24
12	I4999-06		1	60.600	61.156	08/26/2017	12:28
13	I4999-07		1	57.700	58.227	08/26/2017	12:30
14	I4999-08		1	32.200	32.478	08/26/2017	12:33
15	CCV2	10	1	9.850	9.909	08/26/2017	12:36
16	CCB2		1	0.097	0.060	08/26/2017	12:39
17	I4999-09		1	9.110	9.161	08/26/2017	12:42
18	I4999-10		1	25.500	25.712	08/26/2017	12:45
19	I4999-11		1	14.000	14.099	08/26/2017	12:47
20	I4999-12		1	17.900	18.038	08/26/2017	12:50
21	I4999-13		1	14.000	14.099	08/26/2017	12:53
22	I4999-14		1	13.700	13.796	08/26/2017	12:56
23	I4999-15		1	17.500	17.634	08/26/2017	12:59
24	I4999-16		1	30.900	31.165	08/26/2017	13:02
25	I4999-17		1	22.100	22.279	08/26/2017	13:06
26	I4999-18		1	19.600	19.754	08/26/2017	13:09
27	CCV3	10	1	9.540	9.596	08/26/2017	13:12
28	CCB3		1	0.099	0.062	08/26/2017	13:16
29	I4999-19		1	24.700	24.904	08/26/2017	13:19
30	I4999-20		1	40.700	41.061	08/26/2017	13:21
31	CCV4	10	1	9.420	9.475	08/26/2017	13:24
32	CCB4		1	0.101	0.064	08/26/2017	13:27

LB 89736

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TUR082617-1 WorkList ID : 102319 Date : 8/26/2017 8:41:014AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I4999-01	Turbidity	Cool 4 deg C	HDR103	O53	20170825-131623	08/25/2017	SM2130 B
	Water	I4999-02	Turbidity	Cool 4 deg C	HDR103	O53	20170825-131646	08/25/2017	SM2130 B
	Water	I4999-03	Turbidity	Cool 4 deg C	HDR103	O53	20170825-13176	08/25/2017	SM2130 B
	Water	I4999-04	Turbidity	Cool 4 deg C	HDR103	O53	20170825-131814	08/25/2017	SM2130 B
	Water	I4999-05	Turbidity	Cool 4 deg C	HDR103	O53	20170825-131832	08/25/2017	SM2130 B
	Water	I4999-06	Turbidity	Cool 4 deg C	HDR103	O53	20170825-131848	08/25/2017	SM2130 B
	Water	I4999-07	Turbidity	Cool 4 deg C	HDR103	O53	20170825-131926	08/25/2017	SM2130 B
	Water	I4999-08	Turbidity	Cool 4 deg C	HDR103	O53	20170825-131946	08/25/2017	SM2130 B
	Water	I4999-09	Turbidity	Cool 4 deg C	HDR103	O53	20170825-133427	08/25/2017	SM2130 B
	Water	I4999-10	Turbidity	Cool 4 deg C	HDR103	O53	20170825-133446	08/25/2017	SM2130 B
	Water	I4999-11	Turbidity	Cool 4 deg C	HDR103	O53	20170825-13354	08/25/2017	SM2130 B
	Water	I4999-12	Turbidity	Cool 4 deg C	HDR103	O53	20170825-13360	08/25/2017	SM2130 B
	Water	I4999-13	Turbidity	Cool 4 deg C	HDR103	O53	20170825-133617	08/25/2017	SM2130 B
	Water	I4999-14	Turbidity	Cool 4 deg C	HDR103	O53	20170825-133634	08/25/2017	SM2130 B
	Water	I4999-15	Turbidity	Cool 4 deg C	HDR103	O53	20170825-134531	08/25/2017	SM2130 B
	Water	I4999-16	Turbidity	Cool 4 deg C	HDR103	O53	20170825-134547	08/25/2017	SM2130 B
	Water	I4999-17	Turbidity	Cool 4 deg C	HDR103	O53	20170825-13464	08/25/2017	SM2130 B
	Water	I4999-18	Turbidity	Cool 4 deg C	HDR103	O53	20170825-134753	08/25/2017	SM2130 B
	Water	I4999-19	Turbidity	Cool 4 deg C	HDR103	O53	20170825-13488	08/25/2017	SM2130 B
	Water	I4999-20	Turbidity	Cool 4 deg C	HDR103	O53	20170825-134822	08/25/2017	SM2130 B

Date/Time 8/26/17 8:45
 Received by: [Signature]
 Relinquished by: [Signature]

Date/Time 8/26/17 13:30
 Received by: [Signature]
 Relinquished by: [Signature]