



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB89737

Reagent/Standard	Lot/Loq #
TURBIDITY ICV 10 NTU	WP57521
TURBIDITY CCV 10 NTU	WP57522
Turbidity std, 40NTU	WP57515
turbidity std 5 NTU	WP57518
turbidity std 20 NTU	WP57516
Turbidity std 1 NTU	WP57519
turbidity std 0 NTU	WP57520
turbidity std 10 NTU	WP57517
Turbidity std, 80NTU	WP57523

Intercept: 0.0374

Slope: 0.9903

Regression: 1

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.099	08/26/2017	09:01
2	CAL2	1	1	1.049	08/26/2017	09:04
3	CAL3	5	1	4.953	08/26/2017	09:10
4	CAL4	10	1	9.951	08/26/2017	09:15
5	CAL5	20	1	19.600	08/26/2017	09:20
6	CAL6	40	1	39.900	08/26/2017	09:24
7	CAL7	80	1	79.200	08/26/2017	09:26



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB89737

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.200	9.252	08/26/2017	09:27
2	ICB		1	0.095	0.058	08/26/2017	09:30
3	CCV1	10	1	9.200	9.252	08/26/2017	10:47
4	CCB1		1	0.103	0.066	08/26/2017	10:50
5	LB89737BL		1	0.100	0.063	08/26/2017	10:52
6	I5000-01		1	3.770	3.769	08/26/2017	10:55
7	I5000-01DUP		1	3.790	3.789	08/26/2017	11:00
8	I5000-02		1	4.730	4.739	08/26/2017	11:03
9	I5000-03		1	4.670	4.678	08/26/2017	11:06
10	I5000-04		1	6.580	6.607	08/26/2017	11:09
11	I5000-05		1	16.500	16.624	08/26/2017	11:12
12	I5000-06		1	3.090	3.083	08/26/2017	11:15
13	I5000-07		1	30.700	30.963	08/26/2017	11:17
14	I5000-08		1	23.500	23.692	08/26/2017	11:20
15	CCV2	10	1	9.900	9.959	08/26/2017	11:23
16	CCB2		1	0.099	0.062	08/26/2017	11:26
17	I5000-09		1	23.800	23.995	08/26/2017	11:34
18	I5000-10		1	17.500	17.634	08/26/2017	11:36
19	I5000-11		1	15.100	15.210	08/26/2017	11:39
20	I5000-12		1	7.070	7.101	08/26/2017	11:42
21	I5000-13		1	18.400	18.542	08/26/2017	11:45
22	I5000-14		1	25.900	26.116	08/26/2017	11:48
23	I5000-15		1	22.000	22.178	08/26/2017	11:51
24	I5000-16		1	10.500	10.565	08/26/2017	11:54
25	I5000-17		1	20.900	21.067	08/26/2017	11:57
26	CCV3	10	1	9.600	9.656	08/26/2017	12:00
27	CCB3		1	0.098	0.061	08/26/2017	12:02

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TUR082617-2

WorkList ID : 102320

Date : 8/26/2017 8:41:15 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I5000-01	Turbidity	Cool 4 deg C	HDR103	O53	20170825-13593	08/25/2017	SM2130 B
	Water	I5000-02	Turbidity	Cool 4 deg C	HDR103	O53	20170825-135917	08/25/2017	SM2130 B
	Water	I5000-03	Turbidity	Cool 4 deg C	HDR103	O53	20170825-135929	08/25/2017	SM2130 B
	Water	I5000-04	Turbidity	Cool 4 deg C	HDR103	O53	20170825-135956	08/25/2017	SM2130 B
	Water	I5000-05	Turbidity	Cool 4 deg C	HDR103	O53	20170825-1409	08/25/2017	SM2130 B
	Water	I5000-06	Turbidity	Cool 4 deg C	HDR103	O53	20170825-14021	08/25/2017	SM2130 B
	Water	I5000-07	Turbidity	Cool 4 deg C	HDR103	O53	20170825-141238	08/25/2017	SM2130 B
	Water	I5000-08	Turbidity	Cool 4 deg C	HDR103	O53	20170825-141251	08/25/2017	SM2130 B
	Water	I5000-09	Turbidity	Cool 4 deg C	HDR103	O53	20170825-14137	08/25/2017	SM2130 B
	Water	I5000-10	Turbidity	Cool 4 deg C	HDR103	O53	20170825-142534	08/25/2017	SM2130 B
	Water	I5000-11	Turbidity	Cool 4 deg C	HDR103	O53	20170825-142549	08/25/2017	SM2130 B
	Water	I5000-12	Turbidity	Cool 4 deg C	HDR103	O53	20170825-14263	08/25/2017	SM2130 B
	Water	I5000-13	Turbidity	Cool 4 deg C	HDR103	O53	20170825-15927	08/25/2017	SM2130 B
	Water	I5000-14	Turbidity	Cool 4 deg C	HDR103	O53	20170825-15944	08/25/2017	SM2130 B
	Water	I5000-15	Turbidity	Cool 4 deg C	HDR103	O53	20170825-15101	08/25/2017	SM2130 B
	Water	I5000-16	Turbidity	Cool 4 deg C	HDR103	O53	20170825-151057	08/25/2017	SM2130 B
	Water	I5000-17	Turbidity	Cool 4 deg C	HDR103	O53	20170825-151111	08/25/2017	SM2130 B

Date/Time 8/26/17 8:45
 Received by: AK
 Relinquished by: JP

Date/Time 8/26/17 13:30
 Received by: JP
 Relinquished by: AK