


7470A

LB 89741
INSTRUMENT ID: CV1

Linear

 μ Abs.:

36186



A= 0.0000e+000

B= 2.7587e-004

C= -1.2704e-003

Rho= 0.9999154

Accept = Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.014	-0.014	-47	0.000	-47				
0.2	0.200	0.204	0.004	744	0.0 %	744				
2.5	2.500	2.550	0.050	9250	0.0 %	9250				
5.0	5.000	4.917	-0.083	17828	0.0 %	17828				
7.5	7.500	7.562	0.062	27415	0.0 %	27415				
10.0	10.000	9.981	-0.019	36186	0.0 %	36186				

LB89741 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 28 Aug 2017 10:22:12

Sample ID	Extended ID	μ Abs	Conc	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	80	-47	-	0.00007470A	PPB	28 Aug 2017 10:37:19	S	Std
0.2 - 1	80.2	744	-	0.20007470A	PPB	28 Aug 2017 10:39:27	S	Std
2.5 - 1	82.5	9250	-	2.50007470A	PPB	28 Aug 2017 10:41:36	S	Std
5.0 - 1	85	17828	-	5.00007470A	PPB	28 Aug 2017 10:43:44	S	Std
7.5 - 1	87.5	27415	-	7.50007470A	PPB	28 Aug 2017 10:45:52	S	Std
10.0 - 1	810	36186	-	10.00007470A	PPB	28 Aug 2017 10:47:59	S	Std
ICV31 - 1	ICV31	14031	3.8694	-7470A	PPB	28 Aug 2017 10:50:31	U	SMPL
ICB31 - 1	ICB31	-146	-0.0415	-7470A	PPB	28 Aug 2017 10:52:37	U	SMPL
CCV64 - 1	CCV64	18186	5.0156	-7470A	PPB	28 Aug 2017 10:54:45	U	SMPL
CCB64 - 1	CCB64	-156	-0.0443	-7470A	PPB	28 Aug 2017 10:56:51	U	SMPL
CRA - 1	CRA	669	0.1833	-7470A	PPB	28 Aug 2017 10:58:58	U	SMPL
HighStd - 1	HighStd	35814	9.8786	-7470A	PPB	28 Aug 2017 11:01:04	U	SMPL
ChkStd - 1	ChkStd	24857	6.8560	-7470A	PPB	28 Aug 2017 11:03:11	U	SMPL
PB101837BL - 1	PBW	87	0.0227	-7470A	PPB	28 Aug 2017 11:05:20	U	SMPL
PB101837BS - 1	LCSW	14565	4.0167	-7470A	PPB	28 Aug 2017 11:07:29	U	SMPL
PB101837TB - 1	PB101801TB	119	0.0316	-7470A	PPB	28 Aug 2017 11:09:37	U	SMPL
I4948-03 - 1	MANHOLE-A-COMP	125	0.0332	-7470A	PPB	28 Aug 2017 11:11:45	U	SMPL
I4948-06 - 1	TP-2-COMP	149	0.0398	-7470A	PPB	28 Aug 2017 11:13:53	U	SMPL
I4954-03 - 1	347-COMP	168	0.0451	-7470A	PPB	28 Aug 2017 11:16:02	U	SMPL
I4959-03 - 1	RT2321-RT2315-20-02-RB36575	225	0.0608	-7470A	PPB	28 Aug 2017 11:18:09	U	SMPL
CCV65 - 1	CCV65	17630	4.8623	-7470A	PPB	28 Aug 2017 11:20:16	U	SMPL
CCB65 - 1	CCB65	-118	-0.0338	-7470A	PPB	28 Aug 2017 11:22:22	U	SMPL
I4959-06 - 1	2013-RT3425-RT880-RT45915	228	0.0616	-7470A	PPB	28 Aug 2017 11:24:31	U	SMPL
I4961-15 - 1	SB-11-COMP	283	0.0768	-7470A	PPB	28 Aug 2017 11:26:38	U	SMPL
I4961-16 - 1	SB-13-COMP	372	0.1014	-7470A	PPB	28 Aug 2017 11:28:45	U	SMPL
I4961-17 - 1	SB-12-COMP	351	0.0956	-7470A	PPB	28 Aug 2017 11:30:52	U	SMPL
I4961-18 - 1	SB-01-COMP	313	0.0851	-7470A	PPB	28 Aug 2017 11:32:59	U	SMPL
I4961-19 - 1	SB-02-COMP	194	0.0522	-7470A	PPB	28 Aug 2017 11:35:07	U	SMPL
I4961-20 - 1	SB-03-COMP	188	0.0506	-7470A	PPB	28 Aug 2017 11:37:15	U	SMPL
I4961-21 - 1	SB-04-COMP	186	0.0500	-7470A	PPB	28 Aug 2017 11:39:23	U	SMPL
I4964-03 - 1	OK-03-82417	218	0.0589	-7470A	PPB	28 Aug 2017 11:41:32	U	SMPL
I4968-02 - 1	156-275	220	0.0594	-7470A	PPB	28 Aug 2017 11:43:39	U	SMPL
CCV66 - 1	CCV66	17648	4.8672	-7470A	PPB	28 Aug 2017 11:45:46	U	SMPL
CCB66 - 1	CCB66	-172	-0.0487	-7470A	PPB	28 Aug 2017 11:47:53	U	SMPL
I4969-03 - 1	BU-01-082417-A	156	0.0418	-7470A	PPB	28 Aug 2017 11:50:00	U	SMPL
I4969-06 - 1	BU-01-082417-B	344	0.0936	-7470A	PPB	28 Aug 2017 11:52:06	U	SMPL
I4970-02 - 1	HAMILTON-FRAC-TANK	372	0.1014	-7470A	PPB	28 Aug 2017 11:54:13	U	SMPL
I4970-02DUP - 1	HAMILTON-FRAC-TANKDUP	396	0.1080	-7470A	PPB	28 Aug 2017 11:56:20	U	SMPL
I4970-02MS - 1	HAMILTON-FRAC-TANKMS	13960	3.8498	-7470A	PPB	28 Aug 2017 11:58:28	U	SMPL
I4970-02MSD - 1	HAMILTON-FRAC-TANKMSD	13690	3.7753	-7470A	PPB	28 Aug 2017 12:00:35	U	SMPL
PB101838BL - 1	PBW	-181	-0.0512	-7470A	PPB	28 Aug 2017 12:02:44	U	SMPL
PB101838BS - 1	LCSW	13130	3.6209	-7470A	PPB	28 Aug 2017 12:04:53	U	SMPL
I4970-01 - 1	HAMILTON-FRAC-TANK	-334	-0.0934	-7470A	PPB	28 Aug 2017 12:21:05	U	SMPL
I4970-01DUP - 1	HAMILTON-FRAC-TANKDUP	-332	-0.0929	-7470A	PPB	28 Aug 2017 12:23:12	U	SMPL
CCV67 - 1	CCV67	17826	4.9163	-7470A	PPB	28 Aug 2017 12:25:20	U	SMPL
CCB67 - 1	CCB67	-127	-0.0363	-7470A	PPB	28 Aug 2017 12:27:27	U	SMPL
I4970-01MS - 1	HAMILTON-FRAC-TANKMS	13552	3.7373	-7470A	PPB	28 Aug 2017 12:29:34	U	SMPL
I4970-01MSD - 1	HAMILTON-FRAC-TANKMSD	13612	3.7538	-7470A	PPB	28 Aug 2017 12:31:41	U	SMPL
I4970-02LX5 - 1		-54	-0.0162	-7470A	PPB	28 Aug 2017 12:33:49	U	SMPL
I4970-02A - 1		14549	4.0123	-7470A	PPB	28 Aug 2017 12:35:56	U	SMPL
I4970-01LX5 - 1		-191	-0.0540	-7470A	PPB	28 Aug 2017 12:38:03	U	SMPL
I4970-01A - 1		14382	3.9662	-7470A	PPB	28 Aug 2017 12:40:11	U	SMPL
CCV68 - 1	CCV68	17596	4.8529	-7470A	PPB	28 Aug 2017 12:42:19	U	SMPL
CCB68 - 1	CCB68	-152	-0.0432	-7470A	PPB	28 Aug 2017 12:44:28	U	SMPL