



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB89745

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	08/28/2017	09:15
2	I4923-01		1	>100.0	>212.0	08/28/2017	09:40
3	I4923-01DUP		1	>100.0	>212.0	08/28/2017	10:02
4	I4923-02		1	>100.0	>212.0	08/28/2017	10:35
5	I4970-02		1	>100.0	>212.0	08/28/2017	10:58

Result = (Celsius * 1.8) + 32

LB89745

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP82617

WorkList ID : 102324

Date : 8/26/2017 1:44:07 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I4923-01	Flash Point	Cool 4 deg C	LIRO01	N41	SB04-GW	08/22/2017	1010A
	Water	I4923-02	Flash Point	Cool 4 deg C	LIRO01	N41	SB01-GW	08/22/2017	1010A
08/29/2017	Water	I4970-02	Flash Point	Cool 4 deg C	PSEG01	O32	HAMILTON-FRAC-TANK	08/24/2017	1010A

Date/Time 08/28/17 08:30 AM
Received by: KS
Relinquished by: CP

Date/Time 08/28/17 11:05 AM
Received by: CP
Relinquished by: KS