



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB89748

Temp.: 22.6 °C

BalanceID: WC SC-4

SLOPE.: 98.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	08/28/2017	10:15
2	CAL2	1	Water			7.01	08/28/2017	10:16
3	CAL3	1	Water			10.03	08/28/2017	10:17
4	ICV	1	Water			7.00	08/28/2017	10:18
5	CCV1	1	Water			2.02	08/28/2017	10:20
6	I4979-06	1	Solid	20.02	20	6.46	08/28/2017	10:21
7	I4979-07	1	Solid	20.01	20	6.04	08/28/2017	10:22
8	I4979-08	1	Solid	20.03	20	6.27	08/28/2017	10:23
9	I4979-09	1	Solid	20.04	20	6.50	08/28/2017	10:25
10	I4979-10	1	Solid	20.01	20	6.40	08/28/2017	10:26
11	I4980-02	1	Solid	20.02	20	8.54	08/28/2017	10:27
12	I4987-11	1	Solid	20.03	20	7.65	08/28/2017	10:29
13	I4987-12	1	Solid	20.01	20	7.62	08/28/2017	10:30
14	I4987-13	1	Solid	20.02	20	8.11	08/28/2017	10:32
15	I4987-14	1	Solid	20.01	20	8.26	08/28/2017	10:34
16	CCV2	1	Water			12.01	08/28/2017	10:35
17	I4987-15	1	Solid	20.03	20	8.52	08/28/2017	10:36
18	I4997-03	1	Solid	20.02	20	8.43	08/28/2017	10:37
19	I4997-03DUP	1	Solid	20.03	20	8.44	08/28/2017	10:39
20	CCV3	1	Water			12.01	08/28/2017	10:40

WORKLIST(Hardcopy Internal Chain)

2089742 30 2089748

WorkList Name : CORROSIONITY I4979

WorkList ID : 102346

Date : 8/28/2017 9:56:09 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I4979-06	Corrosivity	Cool 4 deg C	TECT01	O41	TP-1	08/24/2017	9045C
	Solid	I4979-07	Corrosivity	Cool 4 deg C	TECT01	O41	TP-2	08/24/2017	9045C
	Solid	I4979-08	Corrosivity	Cool 4 deg C	TECT01	O41	TP-3	08/24/2017	9045C
	Solid	I4979-09	Corrosivity	Cool 4 deg C	TECT01	O41	TP-4	08/24/2017	9045C
	Solid	I4979-10	Corrosivity	Cool 4 deg C	TECT01	O41	TP-5	08/24/2017	9045C
	Solid	I4980-02	Corrosivity	Cool 4 deg C	AREC01	O22	H26-WC1	08/23/2017	9045C
	Solid	I4987-11	Corrosivity	Cool 4 deg C	LIR001	O41	SB-05-COMP	08/24/2017	9045C
	Solid	I4987-12	Corrosivity	Cool 4 deg C	LIR001	O41	SB-06-COMP	08/24/2017	9045C
	Solid	I4987-13	Corrosivity	Cool 4 deg C	LIR001	O41	SB-07-COMP	08/25/2017	9045C
	Solid	I4987-14	Corrosivity	Cool 4 deg C	LIR001	O41	SB-08-COMP	08/25/2017	9045C
	Solid	I4987-15	Corrosivity	Cool 4 deg C	LIR001	O41	SB-09-COMP	08/25/2017	9045C
	Solid	I4997-03	Corrosivity	Cool 4 deg C	TECT01	O11	TEC-B1-COMP	08/25/2017	9045C

Date/Time 08/28/17 10:00 AM
 Received by: JO
 Relinquished by: JO

Date/Time 08/28/17 10:20 AM
 Received by: JO
 Relinquished by: JO