


7471A

LB89751
INSTRUMENT ID: CV1

Linear

 μ Abs.:

43097



A= 0.0000e+000

B= 2.3206e-004

C= -7.4427e-002

Rho= 0.9998332

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.033	-0.033	177	0.000	177				
0.2	0.200	0.199	-0.001	1180	0.0 %	1180				
2.5	2.500	2.551	0.051	11315	0.0 %	11315				
5.0	5.000	4.937	-0.063	21594	0.0 %	21594				
7.5	7.500	7.619	0.119	33152	0.0 %	33152	0			
10.0	10.000	9.927	-0.073	43097	0.0 %	43097				

LB89751 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 28 Aug 2017 14:09:14

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	177	-	0.0000	7471A	PPB	28 Aug 2017 14:28:14	S	Std
0.2 - 1	50.2	1180	-	0.2000	7471A	PPB	28 Aug 2017 14:30:31	S	Std
2.5 - 1	52.5	11315	-	2.5000	7471A	PPB	28 Aug 2017 14:32:49	S	Std
5.0 - 1	55	21594	-	5.0000	7471A	PPB	28 Aug 2017 14:35:06	S	Std
7.5 - 1	57.5	33152	-	7.5000	7471A	PPB	28 Aug 2017 14:37:23	S	Std
10.0 - 1	50	43097	-	10.0000	7471A	PPB	28 Aug 2017 14:39:40	S	Std
ICV32 - 1	ICV32	16769	3.8171	-	7471A	PPB	28 Aug 2017 14:42:23	U	SMPL
ICB32 - 1	ICB32	188	-0.0308	-	7471A	PPB	28 Aug 2017 14:44:38	U	SMPL
CCV69 - 1	CCV69	20903	4.7764	-	7471A	PPB	28 Aug 2017 14:46:54	U	SMPL
CCB69 - 1	CCB69	239	-0.0190	-	7471A	PPB	28 Aug 2017 14:49:10	U	SMPL
CRA - 1	CRA	1194	0.2027	-	7471A	PPB	28 Aug 2017 14:51:26	U	SMPL
HighStd - 1	HighStd	43493	10.0187	-	7471A	PPB	28 Aug 2017 14:53:41	U	SMPL
ChkStd - 1	ChkStd	30121	6.9156	-	7471A	PPB	28 Aug 2017 14:55:58	U	SMPL
PB101843BL - 1	PBS	-24	-0.0800	-	7471A	PPB	28 Aug 2017 14:58:16	U	SMPL
PB101843BS - 1	LCSS	14539	3.2996	-	7471A	PPB	28 Aug 2017 15:00:33	U	SMPL
I4954-01 - 1	347-COMP	2560	0.5197	-	7471A	PPB	28 Aug 2017 15:02:50	U	SMPL
I4954-01DUP - 1	347-COMPDUP	2895	0.5974	-	7471A	PPB	28 Aug 2017 15:05:08	U	SMPL
I4954-01MS - 1	347-COMPMS	19642	4.4838	-	7471A	PPB	28 Aug 2017 15:07:25	U	SMPL
I4954-01MSD - 1	347-COMPMSD	17713	4.0361	-	7471A	PPB	28 Aug 2017 15:09:43	U	SMPL
I4955-01 - 1	FK-SF-03-002-A	1774	0.3373	-	7471A	PPB	28 Aug 2017 15:11:59	U	SMPL
CCV70 - 1	CCV70	20874	4.7697	-	7471A	PPB	28 Aug 2017 15:14:15	U	SMPL
CCB70 - 1	CCB70	204	-0.0271	-	7471A	PPB	28 Aug 2017 15:16:31	U	SMPL
I4955-03 - 1	FK-SF-03-002-B	1468	0.2662	-	7471A	PPB	28 Aug 2017 15:18:47	U	SMPL
I4956-01 - 1	FK-GF-02-010-A	930	0.1414	-	7471A	PPB	28 Aug 2017 15:21:04	U	SMPL
I4956-03 - 1	FK-GF-02-010-B	1011	0.1602	-	7471A	PPB	28 Aug 2017 15:23:20	U	SMPL
I4959-01 - 1	RT2321-RT2315-20-02-RB36575	88426	20.4461	-	7471A	PPB	28 Aug 2017 15:25:36	U	SMPL
I4959-04 - 1	2013-RT3425-RT880-RT45915	142479	32.9898	-	7471A	PPB	28 Aug 2017 15:27:53	U	SMPL
I4964-01 - 1	OK-03-082417	4	-0.0735	-	7471A	PPB	28 Aug 2017 15:30:12	U	SMPL
I4965-01 - 1	TOPSOIL-OLDCASTLE-CO	6639	1.4662	-	7471A	PPB	28 Aug 2017 15:32:35	U	SMPL
I4968-01 - 1	156-275	20600	4.7061	-	7471A	PPB	28 Aug 2017 15:34:52	U	SMPL
I4969-01 - 1	BU-01-082417-A	1398	0.2500	-	7471A	PPB	28 Aug 2017 15:37:10	U	SMPL
I4969-04 - 1	BU-01-082417-B	990	0.1553	-	7471A	PPB	28 Aug 2017 15:39:26	U	SMPL
CCV71 - 1	CCV71	20731	4.7365	-	7471A	PPB	28 Aug 2017 15:41:43	U	SMPL
CCB71 - 1	CCB71	192	-0.0299	-	7471A	PPB	28 Aug 2017 15:43:59	U	SMPL
I4979-01 - 1	TP-1	15270	3.4692	-	7471A	PPB	28 Aug 2017 15:46:15	U	SMPL
I4979-02 - 1	TP-2	936	0.1428	-	7471A	PPB	28 Aug 2017 15:48:32	U	SMPL
I4979-03 - 1	TP-3	3286	0.6881	-	7471A	PPB	28 Aug 2017 15:50:48	U	SMPL
I4979-04 - 1	TP-4	11008	2.4801	-	7471A	PPB	28 Aug 2017 15:53:05	U	SMPL
I4979-05 - 1	TP-5	4511	0.9724	-	7471A	PPB	28 Aug 2017 15:55:22	U	SMPL
PB101844BL - 1	PBS	171	-0.0347	-	7471A	PPB	28 Aug 2017 15:57:39	U	SMPL
PB101844BS - 1	LCSS	14744	3.3471	-	7471A	PPB	28 Aug 2017 15:59:56	U	SMPL
I4980-01 - 1	H26-WC1	1639	0.3059	-	7471A	PPB	28 Aug 2017 16:02:14	U	SMPL
I4980-01DUP - 1	H26-WC1DUP	1812	0.3461	-	7471A	PPB	28 Aug 2017 16:04:32	U	SMPL
I4980-01MS - 1	H26-WC1MS	17665	4.0250	-	7471A	PPB	28 Aug 2017 16:06:48	U	SMPL
CCV72 - 1	CCV72	20463	4.6743	-	7471A	PPB	28 Aug 2017 16:09:04	U	SMPL
CCB72 - 1	CCB72	239	-0.0190	-	7471A	PPB	28 Aug 2017 16:11:21	U	SMPL
I4980-01MSD - 1	H26-WC1MSD	17686	4.0299	-	7471A	PPB	28 Aug 2017 16:13:37	U	SMPL
I4981-01 - 1	S14-WC-01	38684	8.9027	-	7471A	PPB	28 Aug 2017 16:15:54	U	SMPL
I4985-01 - 1	SI-1	163	-0.0366	-	7471A	PPB	28 Aug 2017 16:18:11	U	SMPL
I4986-01 - 1	SU-01-82517	1049	0.1690	-	7471A	PPB	28 Aug 2017 16:20:28	U	SMPL
I4988-01 - 1	FF1-COMP	894	0.1330	-	7471A	PPB	28 Aug 2017 16:22:45	U	SMPL
I4988-04 - 1	FF2-COMP	879	0.1296	-	7471A	PPB	28 Aug 2017 16:25:02	U	SMPL
I4988-07 - 1	FF3-COMP	290	-0.0071	-	7471A	PPB	28 Aug 2017 16:27:19	U	SMPL
I4988-10 - 1	FF4-COMP	1016	0.1614	-	7471A	PPB	28 Aug 2017 16:29:36	U	SMPL
I4988-13 - 1	FF5-COMP	671	0.0813	-	7471A	PPB	28 Aug 2017 16:31:56	U	SMPL
I4989-01 - 1	BERGEN-GEN-S1-COMP	7292	1.6178	-	7471A	PPB	28 Aug 2017 16:34:12	U	SMPL
CCV73 - 1	CCV73	20324	4.6420	-	7471A	PPB	28 Aug 2017 16:36:28	U	SMPL
CCB73 - 1	CCB73	200	-0.0280	-	7471A	PPB	28 Aug 2017 16:38:45	U	SMPL
I4989-04 - 1	BERGEN-GEN-S2-COMP	3404	0.7155	-	7471A	PPB	28 Aug 2017 16:41:01	U	SMPL
I4989-07 - 1	BERGEN-GEN-S3-COMP	4206	0.9016	-	7471A	PPB	28 Aug 2017 16:43:17	U	SMPL
I4989-10 - 1	BERGEN-GEN-S4-COMP	17258	3.9305	-	7471A	PPB	28 Aug 2017 16:45:34	U	SMPL
I4992-01 - 1	SECTION-1-STONES	322	0.0003	-	7471A	PPB	28 Aug 2017 16:47:50	U	SMPL
I4992-03 - 1	SECTION-2-STONES	2235	0.4442	-	7471A	PPB	28 Aug 2017 16:50:07	U	SMPL
I4993-01 - 1	HR-05-082517	5501	1.2022	-	7471A	PPB	28 Aug 2017 16:52:24	U	SMPL
I4994-01 - 1	EO-01-082517-A	2362	0.4737	-	7471A	PPB	28 Aug 2017 16:54:41	U	SMPL
I4994-04 - 1	EO-01-082517-B	2176	0.4305	-	7471A	PPB	28 Aug 2017 16:56:59	U	SMPL
I4994-07 - 1	EO-01-082517-C	2875	0.5928	-	7471A	PPB	28 Aug 2017 16:59:17	U	SMPL
CCV74 - 1	CCV74	20643	4.7161	-	7471A	PPB	28 Aug 2017 17:01:34	U	SMPL
CCB74 - 1	CCB74	199	-0.0282	-	7471A	PPB	28 Aug 2017 17:03:50	U	SMPL
I4954-01LX5 - 1		753	0.1003	-	7471A	PPB	28 Aug 2017 17:06:06	U	SMPL
I4954-01A - 1		20199	4.6130	-	7471A	PPB	28 Aug 2017 17:08:23	U	SMPL
I4980-01LX5 - 1		473	0.0353	-	7471A	PPB	28 Aug 2017 17:10:40	U	SMPL
I4980-01A - 1		17984	4.0990	-	7471A	PPB	28 Aug 2017 17:12:57	U	SMPL
I4959-01DLX5 - 1		17617	4.0138	-	7471A	PPB	28 Aug 2017 17:15:13	U	SMPL
I4959-04DLX5 - 1		28687	6.5828	-	7471A	PPB	28 Aug 2017 17:17:30	U	SMPL
CCV75 - 1	CCV75	20720	4.7339	-	7471A	PPB	28 Aug 2017 17:19:47	U	SMPL
CCB75 - 1	CCB75	195	-0.0292	-	7471A	PPB	28 Aug 2017 17:22:05	U	SMPL