



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By :

Run Number: LB89761

Temp.: 22.4 °C

BalanceID: WC SC-4

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	08/28/2017	16:10
2	CAL2	1	Water			7.01	08/28/2017	16:11
3	CAL3	1	Water			10.04	08/28/2017	16:12
4	ICV	1	Water			7.00	08/28/2017	16:13
5	CCV1	1	Water			2.01	08/28/2017	16:15
6	I5006-05	1	Solid	20.02	20	9.33	08/28/2017	16:17
7	I5006-06	1	Solid	20.04	20	7.93	08/28/2017	16:18
8	I5007-03	1	Solid	20.01	20	7.44	08/28/2017	16:20
9	I5008-03	1	Solid	20.03	20	8.06	08/28/2017	16:21
10	I5008-06	1	Solid	20.02	20	8.22	08/28/2017	16:22
11	I5008-06DUP	1	Solid	20.03	20	8.23	08/28/2017	16:25
12	CCV2	1	Water			12.01	08/28/2017	16:26

WORKLIST(Hardcopy Internal Chain)

10389761

WorkList Name : CORROSIONITY I5006

WorkList ID : 102360

Date : 8/28/2017 3:21:08 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I5006-05	Corrosivity	Cool 4 deg C	LIRO01	O41	SB-10-COMP	08/25/2017	9045C
	Solid	I5006-06	Corrosivity	Cool 4 deg C	LIRO01	O41	SB-14-COMP	08/25/2017	9045C
	Solid	I5007-03	Corrosivity	Cool 4 deg C	PSEG05	O52	TR-05-082817	08/28/2017	9045C
	Solid	I5008-03	Corrosivity	Cool 4 deg C	PSEG05	O52	NB-8-082817-A	08/28/2017	9045C
	Solid	I5008-06	Corrosivity	Cool 4 deg C	PSEG05	O52	NB-8-082817-B	08/28/2017	9045C

Date/Time 08/28/17 3:30 PM
 Received by: JD
 Relinquished by: JD

Date/Time 08/28/17 6:00 PM
 Received by: JD
 Relinquished by: JD