


7471A

LB 89777
INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.3345e-004

C= 6.1976e-002

Rho= 0.9998446

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.014	0.014	-204	0.000	-204				
0.2	0.200	0.250	0.050	806	0.0 %	806				
2.5	2.500	2.417	-0.083	10089	0.0 %	10089				
5.0	5.000	4.948	-0.052	20931	0.0 %	20931				
7.5	7.500	7.608	0.108	32325	0.0 %	32325	0			
10.0	10.000	9.962	-0.038	42408	0.0 %	42408				

LB89777 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 29 Aug 2017 15:44:22

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0	-204	-	0.0000	7471A	PPB	29 Aug 2017 15:54:49	S	Std
0.2 - 1	S0.2	806	-	0.2000	7471A	PPB	29 Aug 2017 15:57:06	S	Std
2.5 - 1	S2.5	10089	-	2.5000	7471A	PPB	29 Aug 2017 15:59:24	S	Std
5.0 - 1	S5	20931	-	5.0000	7471A	PPB	29 Aug 2017 16:01:41	S	Std
7.5 - 1	S7.5	32325	-	7.5000	7471A	PPB	29 Aug 2017 16:03:58	S	Std
10.0 - 1	S10	42408	-	10.0000	7471A	PPB	29 Aug 2017 16:06:15	S	Std
ICV34 - 1	ICV34	16014	3.8004	-	7471A	PPB	29 Aug 2017 16:09:59	U	SMPL
ICB34 - 1	ICB34	-227	0.0090	-	7471A	PPB	29 Aug 2017 16:12:15	U	SMPL
CCV81 - 1	CCV81	21020	4.9690	-	7471A	PPB	29 Aug 2017 16:14:31	U	SMPL
CCB81 - 1	CCB81	-219	0.0109	-	7471A	PPB	29 Aug 2017 16:16:46	U	SMPL
CRA - 1	CRA	605	0.2032	-	7471A	PPB	29 Aug 2017 16:21:37	U	SMPL
HighStd - 1	HighStd	43194	10.1454	-	7471A	PPB	29 Aug 2017 16:23:53	U	SMPL
ChkStd - 1	ChkStd	28125	6.6276	-	7471A	PPB	29 Aug 2017 16:26:09	U	SMPL
PB101895BL - 1	PBS	-418	-0.0356	-	7471A	PPB	29 Aug 2017 16:28:27	U	SMPL
PB101895BS - 1	LCSS	14720	3.4983	-	7471A	PPB	29 Aug 2017 16:30:45	U	SMPL
I4997-01 - 1	TEC-B1-COMP	10085	2.4163	-	7471A	PPB	29 Aug 2017 16:33:02	U	SMPL
I5007-01 - 1	TR-05-082817	924	0.2777	-	7471A	PPB	29 Aug 2017 16:35:18	U	SMPL
I5007-01DUP - 1	TR-05-082817DUP	1158	0.3323	-	7471A	PPB	29 Aug 2017 16:37:38	U	SMPL
I5007-01MS - 1	TR-05-082817MS	17055	4.0434	-	7471A	PPB	29 Aug 2017 16:39:56	U	SMPL
I5007-01MSD - 1	TR-05-082817MSD	17267	4.0929	-	7471A	PPB	29 Aug 2017 16:42:12	U	SMPL
CCV82 - 1	CCV82	21199	5.0108	-	7471A	PPB	29 Aug 2017 16:44:28	U	SMPL
CCB82 - 1	CCB82	-186	0.0186	-	7471A	PPB	29 Aug 2017 16:46:44	U	SMPL
I5008-01 - 1	NB-08-082817-A	3698	0.9253	-	7471A	PPB	29 Aug 2017 16:49:00	U	SMPL
I5008-04 - 1	NB-08-082817-B	1035	0.3036	-	7471A	PPB	29 Aug 2017 16:51:16	U	SMPL
I5007-01LX5 - 1		-9	0.0599	-	7471A	PPB	29 Aug 2017 16:53:32	U	SMPL
I5007-01A - 1		18004	4.2649	-	7471A	PPB	29 Aug 2017 16:55:49	U	SMPL
CCV83 - 1	CCV83	20874	4.9349	-	7471A	PPB	29 Aug 2017 16:58:06	U	SMPL
CCB83 - 1	CCB83	-402	-0.0319	-	7471A	PPB	29 Aug 2017 17:00:23	U	SMPL