



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB89782

Temp.: 22.3 °C

BalanceID: WC SC-4

SLOPE.: 98.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	08/29/2017	16:00
2	CAL2	1	Water			7.00	08/29/2017	16:01
3	CAL3	1	Water			10.03	08/29/2017	16:02
4	ICV	1	Water			7.00	08/29/2017	16:03
5	CCV1	1	Water			2.01	08/29/2017	16:05
6	I5019-07	1	Solid	20.02	20	6.76	08/29/2017	16:06
7	I5023-03	1	Solid	20.03	20	7.45	08/29/2017	16:07
8	I5023-06	1	Solid	20.01	20	7.28	08/29/2017	16:08
9	I5023-09	1	Solid	20.03	20	8.30	08/29/2017	16:10
10	I5023-12	1	Solid	20.05	20	7.63	08/29/2017	16:11
11	I5024-03	1	Solid	20.02	20	7.93	08/29/2017	16:13
12	I5024-06	1	Solid	20.03	20	7.40	08/29/2017	16:15
13	I5024-09	1	Solid	20.02	20	8.52	08/29/2017	16:17
14	I5024-12	1	Solid	20.01	20	8.57	08/29/2017	16:19
15	I5024-15	1	Solid	20.02	20	8.69	08/29/2017	16:21
16	CCV2	1	Water			12.02	08/29/2017	16:23
17	I5027-03	1	Solid	20.01	20	7.61	08/29/2017	16:25
18	I5027-03DUP	1	Solid	20.03	20	7.62	08/29/2017	16:26
19	CCV3	1	Water			12.02	08/29/2017	16:27

WORKLIST(Hardcopy Internal Chain)

1089782

WorkList Name : CORROSIONITY I5027

WorkList ID : 102397

Date : 8/29/2017 3:36:23 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I5019-07	Corrosivity	Cool 4 deg C	PSEG01	O52	STAGE2-TEST-PIT-A-B-C	08/29/2017	9045C
	Solid	I5023-03	Corrosivity	Cool 4 deg C	PSEG05	O52	PL-01-082917-A	08/29/2017	9045C
	Solid	I5023-06	Corrosivity	Cool 4 deg C	PSEG05	O52	PL-01-082917-B	08/29/2017	9045C
	Solid	I5023-09	Corrosivity	Cool 4 deg C	PSEG05	O52	PL-01-082917-C	08/29/2017	9045C
	Solid	I5023-12	Corrosivity	Cool 4 deg C	PSEG05	O52	PL-01-082917-D	08/29/2017	9045C
09/01/2017	Solid	I5024-03	Corrosivity	Cool 4 deg C	PSEG05	O52	CL-01-082917-A	08/29/2017	9045C
09/01/2017	Solid	I5024-06	Corrosivity	Cool 4 deg C	PSEG05	O52	CL-01-082917-B	08/29/2017	9045C
09/01/2017	Solid	I5024-09	Corrosivity	Cool 4 deg C	PSEG05	O52	CL-01-082917-C	08/29/2017	9045C
09/01/2017	Solid	I5024-12	Corrosivity	Cool 4 deg C	PSEG05	O52	CL-01-082917-D	08/29/2017	9045C
09/01/2017	Solid	I5024-15	Corrosivity	Cool 4 deg C	PSEG05	O52	CL-01-082917-E	08/29/2017	9045C
	Solid	I5027-03	Corrosivity	Cool 4 deg C	PSEG05	O51	SU-02-082917	08/29/2017	9045C

Date/Time 08/29/17 3:36 PM

Received by: JF

Relinquished by: JF

Date/Time 08/29/17 5:30 PM

Received by: JF

Relinquished by: JF