


7470A

LB89785
INSTRUMENT ID: CVJ

Linear

 μ Abs.:

35109



A= 0.0000e+000

B= 2.8525e-004

C= 4.1681e-002

Rho= 0.9999391

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.028	0.028	-48	0.000	-48				
0.2	0.200	0.234	0.034	675	0.0 %	675				
2.5	2.500	2.454	-0.046	8456	0.0 %	8456				
5.0	5.000	4.969	-0.031	17274	0.0 %	17274				
7.5	7.500	7.459	-0.041	26002	0.0 %	26002				
10.0	10.000	10.056	0.056	35109	0.0 %	35109				

LB89785 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 30 Aug 2017 09:37:10

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc.	Method	Units	Date	Type	Type
0.0 - 1	50	-48	-	0.0000	7470A	PPB	30 Aug 2017 10:19:26	S	Std
0.2 - 1	50.2	675	-	0.2000	7470A	PPB	30 Aug 2017 10:21:34	S	Std
2.5 - 1	52.5	8456	-	2.5000	7470A	PPB	30 Aug 2017 10:23:43	S	Std
5.0 - 1	55	17274	-	5.0000	7470A	PPB	30 Aug 2017 10:25:51	S	Std
7.5 - 1	57.5	26002	-	7.5000	7470A	PPB	30 Aug 2017 10:27:59	S	Std
10.0 - 1	511	35109	-	10.0000	7470A	PPB	30 Aug 2017 10:32:36	S	Std
ICV35 - 1	ICV35	13142	3.7904	-7470A	PPB	30 Aug 2017 10:35:08	U	SMPL	
ICB35 - 1	ICB35	-223	-0.0219	-7470A	PPB	30 Aug 2017 10:37:14	U	SMPL	
CCV84 - 1	CCV84	17451	5.0195	-7470A	PPB	30 Aug 2017 10:39:20	U	SMPL	
CCB84 - 1	CCB84	-229	-0.0236	-7470A	PPB	30 Aug 2017 10:41:27	U	SMPL	
CRA - 1	CRA	675	0.2342	-7470A	PPB	30 Aug 2017 10:43:34	U	SMPL	
HighStd - 1	HighStd	35395	10.1380	-7470A	PPB	30 Aug 2017 10:45:41	U	SMPL	
ChkStd - 1	ChkStd	23874	6.8516	-7470A	PPB	30 Aug 2017 10:47:48	U	SMPL	
PB101929BL - 1	PBW	-305	-0.0453	-7470A	PPB	30 Aug 2017 10:49:56	U	SMPL	
PB101929BS - 1	LCSW	12654	3.6512	-7470A	PPB	30 Aug 2017 10:52:05	U	SMPL	
PB101929TB - 1	PB101866TB	-206	-0.0171	-7470A	PPB	30 Aug 2017 10:54:13	U	SMPL	
I4979-06 - 1	TP-1	-58	0.0251	-7470A	PPB	30 Aug 2017 10:56:21	U	SMPL	
I4979-06DUP - 1	TP-1DUP	-60	0.0246	-7470A	PPB	30 Aug 2017 10:58:29	U	SMPL	
I4979-06MS - 1	TP-1MS	12979	3.7439	-7470A	PPB	30 Aug 2017 11:00:38	U	SMPL	
I4979-06MSD - 1	TP-1MSD	13566	3.9113	-7470A	PPB	30 Aug 2017 11:02:45	U	SMPL	
CCV85 - 1	CCV85	17592	5.0597	-7470A	PPB	30 Aug 2017 11:04:51	U	SMPL	
CCB85 - 1	CCB85	-235	-0.0254	-7470A	PPB	30 Aug 2017 11:06:58	U	SMPL	
I4979-07 - 1	TP-2	-156	-0.0028	-7470A	PPB	30 Aug 2017 11:09:05	U	SMPL	
I4979-08 - 1	TP-3	-102	0.0126	-7470A	PPB	30 Aug 2017 11:11:12	U	SMPL	
I4979-09 - 1	TP-4	-96	0.0143	-7470A	PPB	30 Aug 2017 11:13:19	U	SMPL	
I4979-10 - 1	TP-5	-24	0.0348	-7470A	PPB	30 Aug 2017 11:15:26	U	SMPL	
I4987-11 - 1	SB-05-COMP	-127	0.0055	-7470A	PPB	30 Aug 2017 11:17:33	U	SMPL	
I4987-12 - 1	SB-06-COMP	-102	0.0126	-7470A	PPB	30 Aug 2017 11:19:40	U	SMPL	
I4987-13 - 1	SB-07-COMP	-145	0.0003	-7470A	PPB	30 Aug 2017 11:21:48	U	SMPL	
I4987-14 - 1	SB-08-COMP	-129	0.0049	-7470A	PPB	30 Aug 2017 11:23:56	U	SMPL	
I4987-15 - 1	SB-09-COMP	-122	0.0069	-7470A	PPB	30 Aug 2017 11:26:04	U	SMPL	
I5003-06 - 1	NYSEG-PH4-ESMI-7B	-77	0.0197	-7470A	PPB	30 Aug 2017 11:28:11	U	SMPL	
CCV86 - 1	CCV86	17791	5.1165	-7470A	PPB	30 Aug 2017 11:30:18	U	SMPL	
CCB86 - 1	CCB86	-209	-0.0179	-7470A	PPB	30 Aug 2017 11:32:25	U	SMPL	
PB101930BL - 1	PBW	-119	0.0077	-7470A	PPB	30 Aug 2017 11:34:33	U	SMPL	
PB101930BS - 1	LCSW	13168	3.7978	-7470A	PPB	30 Aug 2017 11:36:40	U	SMPL	
PB101930TB - 1	PB101890TB	-189	-0.0122	-7470A	PPB	30 Aug 2017 11:38:47	U	SMPL	
I5006-05 - 1	SB-10-COMP	-128	0.0052	-7470A	PPB	30 Aug 2017 11:40:54	U	SMPL	
I5006-06 - 1	SB-14-COMP	-136	0.0029	-7470A	PPB	30 Aug 2017 11:43:01	U	SMPL	
I5007-03 - 1	TR-05-082817	-101	0.0129	-7470A	PPB	30 Aug 2017 11:45:09	U	SMPL	
I5008-03 - 1	NB-08-082817-A	-92	0.0154	-7470A	PPB	30 Aug 2017 11:47:17	U	SMPL	
I5008-06 - 1	NB-08-082817-B	-112	0.0097	-7470A	PPB	30 Aug 2017 11:49:25	U	SMPL	
I5008-06DUP - 1	NB-08-082817-BDUP	-99	0.0134	-7470A	PPB	30 Aug 2017 11:51:34	U	SMPL	
I5008-06MS - 1	NB-08-082817-BMS	13518	3.8976	-7470A	PPB	30 Aug 2017 11:53:41	U	SMPL	
CCV87 - 1	CCV87	17273	4.9687	-7470A	PPB	30 Aug 2017 11:55:48	U	SMPL	
CCB87 - 1	CCB87	-125	0.0060	-7470A	PPB	30 Aug 2017 12:02:33	U	SMPL	
I5008-06MSD - 1	NB-08-082817-BMSD	13673	3.9419	-7470A	PPB	30 Aug 2017 12:04:40	U	SMPL	
I4979-06LX5 - 1		156	0.0862	-7470A	PPB	30 Aug 2017 12:06:47	U	SMPL	
I4979-06A - 1		14556	4.1937	-7470A	PPB	30 Aug 2017 12:08:55	U	SMPL	
I5008-06LX5 - 1		156	0.0862	-7470A	PPB	30 Aug 2017 12:11:01	U	SMPL	
I5008-06A - 1		14239	4.1033	-7470A	PPB	30 Aug 2017 12:13:08	U	SMPL	
CCV88 - 1	CCV88	17380	4.9993	-7470A	PPB	30 Aug 2017 12:15:16	U	SMPL	
CCB88 - 1	CCB88	-231	-0.0242	-7470A	PPB	30 Aug 2017 12:17:24	U	SMPL	